

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN102119\
 Data File : BN008273.D
 Acq On : 21 Oct 2019 09:14
 Operator : JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.405

Quant Time: Oct 21 15:30:04 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 19 22:25:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	1429	0.40	ng/ul	0.00
2) Naphthalene-d8	10.36	136	6617	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.22	164	4206	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.97	188	8744	0.40	ng/ul	0.00
16) Chrysene-d12	21.18	240	8750	0.40	ng/ul	0.00
20) Perylene-d12	23.35	264	8931	0.40	ng/ul	-0.01

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.96	152	4547	0.41	ng/ul	0.00
14) Fluoranthene-d10	19.01	212	12318	0.41	ng/ul	0.00

Target Compounds				Ovalue		
3) Naphthalene	10.40	128	8108	0.430	ng/ul	99
5) 2-Methylnaphthalene	12.03	142	5409	0.419	ng/ul	100
7) Acenaphthylene	13.94	152	7844	0.392	ng/ul	99
8) Acenaphthene	14.29	153	6462	0.415	ng/ul	99
9) Fluorene	15.28	166	7152	0.404	ng/ul	97
11) Pentachlorophenol	16.67	266	542	0.256	ng/ul	96
12) Phenanthrene	17.02	178	11609	0.453	ng/ul	100
13) Anthracene	17.11	178	10127	0.450	ng/ul	100
15) Fluoranthene	19.04	202	14370	0.425	ng/ul	97
17) Pyrene	19.40	202	15023	0.431	ng/ul	97
18) Benzo(a)anthracene	21.17	228	13485	0.424	ng/ul	99
19) Chrysene	21.22	228	16380	0.420	ng/ul	99
21) Benzo(b)fluoranthene	22.71	252	14480	0.498	ng/ul	98
22) Benzo(k)fluoranthene	22.75	252	15658	0.475	ng/ul	99
23) Benzo(a)pyrene	23.26	252	14075	0.460	ng/ul	96
24) Indeno(1,2,3-cd)pyrene	25.51	276	14703	0.408	ng/ul#	94
25) Dibenzo(a,h)anthracene	25.52	278	11824	0.415	ng/ul	99
26) Benzo(g,h,i)perylene	26.17	276	12736	0.385	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN102119\
Data File : BN008273.D
Acq On : 21 Oct 2019 09:14
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.405

Quant Time: Oct 21 15:30:04 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN101019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 19 22:25:11 2019
Response via : Initial Calibration

