


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102119\
 Data File : BN008272.D
 Acq On : 21 Oct 2019 08:39
 Operator : JU
 Sample : DFTPP08
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP08

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN101019.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Oct 21 23:58:15 2019



AutoFind: Scans 2455, 2456, 2457; Background Corrected with Scan 2448

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.1	111800	PASS
68	69	0.00	2	1.8	2630	PASS
69	198	0.00	100	48.7	143104	PASS
70	69	0.00	2	0.6	930	PASS
127	198	10	80	47.9	140560	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	293632	PASS
199	198	5	9	6.8	20056	PASS
275	198	10	60	26.8	78792	PASS
365	198	1	100	3.4	9896	PASS
441	198	0.01	100	10.3	30338	PASS
442	198	50	100	67.2	197458	PASS
443	442	15	24	18.3	36168	PASS

m/z	Abundance
40.00	1193.0
41.00	412.0
43.00	367.0
44.00	3124.0
73.00	974.0
77.00	429.0
207.00	1932.0
208.00	502.0
208.90	383.0
281.00	397.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

m/z	Abundance
37.00	581.0
38.00	1632.0
39.00	9696.0
40.00	1511.0
41.00	727.0
43.00	508.0
44.00	3192.0
45.00	563.0
49.00	1160.0
50.00	31056.0
51.00	119504.0
52.00	5611.0
55.00	885.0
56.00	4075.0
57.00	9086.0
61.00	1888.0
62.00	2022.0
63.00	5728.0
63.90	831.0
65.00	2909.0
67.00	524.0
68.10	2690.0
69.00	148800.0
70.00	1013.0
73.00	2183.0
74.00	13968.0
75.00	23240.0
76.10	8477.0
77.00	158336.0
78.00	11268.0
79.00	10419.0
80.00	8504.0
81.00	11817.0
82.00	2969.0
83.00	2861.0
84.00	497.0
84.90	2269.0
86.00	3387.0
87.10	1569.0
88.00	476.0
91.00	2621.0
92.00	3042.0
93.00	18328.0
93.90	1189.0
94.90	335.0
95.00	336.0
96.00	1094.0
97.00	555.0
98.00	13215.0
99.00	11629.0
100.00	1138.0
100.90	5446.0
102.00	394.0
103.00	1983.0
104.00	4054.0
105.00	3448.0
105.90	1255.0
107.00	43800.0
108.00	7169.0
109.00	1407.0
110.00	75912.0
111.00	12277.0
111.90	1755.0
112.90	548.0
114.90	352.0
116.00	2547.0
117.00	36536.0
118.00	2579.0
118.90	379.0
120.00	528.0
122.00	2859.0
123.00	4476.0
123.90	1630.0
125.00	1710.0
127.00	140800.0
128.00	11215.0
129.00	51720.0
130.00	4465.0
130.90	1148.0
132.10	601.0
133.00	490.0
134.00	1579.0
135.00	4273.0
136.00	1656.0
137.00	2688.0
138.10	478.0
138.80	369.0
140.00	742.0
141.00	6897.0
142.00	2462.0
143.00	1748.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

144.00	410.0
145.00	547.0
146.00	1138.0
147.00	3982.0
148.00	8673.0
149.00	1957.0
150.00	590.0
151.10	1069.0
151.90	589.0
153.00	2026.0
154.00	1686.0
155.00	3886.0
156.00	5780.0
157.00	1089.0
158.00	1184.0
159.00	936.0
159.90	2297.0
161.00	3213.0
162.00	943.0
162.90	411.0
163.80	321.0
165.00	2815.0
166.00	2376.0
167.00	14110.0
168.00	9168.0
169.00	1279.0
170.00	543.0
170.90	700.0
171.90	1245.0
172.90	1291.0
174.00	3146.0
175.00	4992.0
175.90	1527.0
176.90	2747.0
178.00	1200.0
179.00	10996.0
180.00	8329.0
181.00	3289.0
182.00	628.0
183.00	328.0
184.00	786.0
185.00	4889.0
186.00	40104.0
187.00	12042.0
188.00	1214.0
189.00	2387.0
189.90	432.0
191.00	1659.0
192.00	3316.0
193.10	3988.0
193.90	897.0
195.00	490.0
196.00	7196.0
197.90	262208.0
198.90	19144.0
200.00	1641.0
201.70	1282.0
203.00	2266.0
204.00	10503.0
205.00	17096.0
206.00	74816.0
207.00	12415.0
208.00	2770.0
209.00	1187.0
210.20	1433.0
211.00	3162.0
212.10	432.0
214.90	1005.0
216.10	1602.0
217.00	20048.0
217.90	2560.0
218.90	319.0
221.00	11395.0
223.00	4612.0
224.00	42256.0
225.00	11188.0
226.00	1286.0
227.00	17360.0
228.00	2666.0
229.00	3230.0
229.90	524.0
231.00	1205.0
234.00	1358.0
235.00	1189.0
235.90	808.0
236.90	2117.0
237.90	404.0
239.00	513.0
239.90	731.0
241.00	886.0
241.90	2197.0
243.00	2401.0
244.00	28696.0
245.10	3872.0
246.00	6391.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

247.00	1382.0
247.90	383.0
248.90	903.0
251.00	386.0
253.00	697.0
255.00	158272.0
256.00	23688.0
257.00	2158.0
258.00	9710.0
259.00	1532.0
260.00	330.0
260.90	333.0
264.00	363.0
265.00	3360.0
266.00	398.0
270.90	500.0
271.90	372.0
273.00	4514.0
274.00	12427.0
275.00	64432.0
276.00	9418.0
277.00	6404.0
277.90	1091.0
281.00	622.0
283.10	788.0
284.00	400.0
285.00	1120.0
289.00	354.0
293.00	1249.0
295.10	557.0
296.00	19056.0
297.00	2715.0
301.00	344.0
301.90	419.0
303.00	2169.0
304.00	612.0
308.00	341.0
314.00	847.0
315.00	1875.0
315.90	1319.0
321.00	667.0
321.90	412.0
323.00	5850.0
324.00	1117.0
327.00	1139.0
327.90	428.0
332.90	547.0
334.00	3767.0
335.00	778.0
341.00	511.0
345.90	1088.0
352.00	1596.0
353.00	1208.0
354.00	1757.0
355.20	372.0
365.00	7091.0
365.90	1291.0
370.90	453.0
371.90	2565.0
373.00	752.0
382.90	840.0
401.90	927.0
403.00	1563.0
403.90	650.0
420.90	1026.0
422.00	1178.0
423.00	8750.0
424.00	2959.0
425.10	418.0
441.00	20624.0
442.00	127800.0
443.00	26344.0
444.00	2359.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

m/z	Abundance
37.00	688.0
38.00	2083.0
39.00	10450.0
40.00	1550.0
41.00	732.0
43.00	415.0
44.00	3255.0
45.00	449.0
49.10	1288.0
50.00	33616.0
51.00	131392.0
52.00	7318.0
53.10	390.0
55.00	1057.0
56.00	4279.0
57.00	9714.0
58.00	502.0
61.00	1499.0
62.00	2483.0
63.00	6569.0
64.00	878.0
65.00	3285.0
66.00	395.0
66.90	338.0
68.00	3219.0
69.00	172480.0
70.00	1042.0
73.00	2018.0
74.00	16848.0
75.00	26752.0
76.00	9459.0
77.00	185792.0
78.00	13350.0
79.00	12185.0
80.00	8572.0
81.00	13305.0
82.00	3425.0
83.00	3053.0
83.90	423.0
85.00	2343.0
86.00	3742.0
87.00	1973.0
88.00	781.0
91.00	2880.0
92.00	3571.0
93.00	20480.0
94.00	1230.0
94.90	403.0
96.00	1018.0
97.00	454.0
98.00	15674.0
99.00	12765.0
100.00	1355.0
101.00	6822.0
102.10	327.0
102.90	2306.0
104.00	4260.0
105.00	4342.0
106.10	1504.0
107.00	49872.0
108.00	7806.0
109.10	1800.0
110.00	92072.0
111.00	13976.0
112.00	1865.0
112.90	731.0
115.00	351.0
116.00	3094.0
117.00	41416.0
118.00	3571.0
119.00	497.0
120.00	809.0
120.90	351.0
121.90	3338.0
123.00	5389.0
124.00	2269.0
124.90	2051.0
127.00	166976.0
128.00	12525.0
129.00	64552.0
130.00	5731.0
130.90	951.0
131.80	587.0
133.00	593.0
133.90	2189.0
135.00	5518.0
136.00	2420.0
137.00	2734.0
138.00	664.0
139.00	357.0
140.00	666.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

141.00	8584.0
142.00	2858.0
142.90	1715.0
144.00	564.0
145.10	488.0
146.00	1714.0
147.00	4813.0
148.00	10353.0
149.00	2581.0
150.10	686.0
151.10	1254.0
151.80	614.0
152.90	2583.0
154.00	2074.0
155.00	4672.0
156.00	7302.0
156.90	1729.0
157.90	1699.0
159.00	1499.0
160.00	3335.0
161.00	3626.0
162.00	1160.0
163.00	352.0
163.90	549.0
165.00	3363.0
166.00	2638.0
167.00	19040.0
168.00	10502.0
169.00	1765.0
170.00	584.0
170.90	573.0
171.90	1492.0
173.00	2207.0
174.00	3880.0
175.00	6864.0
175.90	1621.0
177.00	3487.0
178.00	1133.0
179.00	13317.0
180.00	10093.0
181.00	4382.0
182.00	710.0
183.00	538.0
184.00	1082.0
185.00	7440.0
186.00	50408.0
187.00	15437.0
188.00	1929.0
189.00	3944.0
190.00	639.0
191.00	1622.0
192.00	4348.0
193.00	5293.0
194.00	1073.0
195.10	749.0
196.00	10243.0
198.00	361472.0
199.00	24152.0
199.90	1867.0
201.50	1725.0
202.90	2648.0
204.00	13967.0
205.00	23312.0
206.00	101792.0
207.00	15335.0
207.90	3697.0
209.00	1427.0
210.20	1542.0
210.90	3409.0
211.90	900.0
215.00	1110.0
216.00	2164.0
217.00	27040.0
218.00	3634.0
221.00	14716.0
221.90	3658.0
223.00	6193.0
224.00	55008.0
225.00	14513.0
226.00	1773.0
227.00	23896.0
228.00	3517.0
229.00	5165.0
230.00	614.0
231.00	2107.0
232.00	467.0
233.00	367.0
234.00	1647.0
235.00	1673.0
236.00	1248.0
237.00	2465.0
237.90	447.0
239.00	970.0
240.00	920.0
240.90	1232.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

242.00	3191.0
243.00	3075.0
244.00	41448.0
245.00	6022.0
246.00	8298.0
247.00	1733.0
248.10	374.0
249.10	1306.0
249.90	366.0
250.90	534.0
252.90	993.0
255.00	221632.0
256.00	32168.0
257.00	3034.0
258.00	13290.0
259.00	2373.0
260.90	493.0
264.00	322.0
265.00	5169.0
265.90	730.0
270.00	364.0
271.00	600.0
272.00	573.0
273.00	6523.0
274.00	16888.0
275.00	97664.0
276.00	12473.0
277.00	8255.0
277.90	1904.0
278.90	487.0
281.00	669.0
281.90	344.0
283.10	925.0
284.00	789.0
285.00	1384.0
289.00	402.0
290.10	320.0
292.00	455.0
293.00	2040.0
294.00	695.0
296.00	28216.0
297.00	3753.0
297.90	371.0
301.90	498.0
303.00	2991.0
304.00	892.0
308.00	408.0
309.90	367.0
314.00	1401.0
315.00	2877.0
316.00	1695.0
316.90	331.0
321.00	1120.0
321.90	424.0
323.00	8482.0
324.00	1456.0
327.00	1894.0
328.00	815.0
332.00	462.0
332.90	888.0
334.00	6138.0
334.90	1243.0
340.90	1113.0
341.90	403.0
345.90	1733.0
346.90	326.0
352.00	2389.0
353.00	1543.0
354.00	2343.0
355.00	486.0
364.90	12233.0
366.00	2007.0
369.90	335.0
370.90	705.0
372.00	4808.0
373.00	956.0
383.00	1154.0
384.00	408.0
390.00	695.0
390.90	345.0
392.10	322.0
400.90	324.0
401.90	1579.0
402.90	2477.0
403.90	1162.0
421.00	2313.0
422.00	2016.0
423.00	14755.0
424.00	4388.0
425.00	456.0
441.00	37064.0
442.00	236352.0
443.00	40504.0
444.00	4129.0

Instrument :

BNA_N

ClientSampleId :

DFTPP08

m/z	Abundance
37.00	473.0
38.00	1144.0
39.00	7075.0
40.00	1408.0
41.00	651.0
43.10	447.0
44.00	2886.0
45.00	452.0
49.00	570.0
50.00	21192.0
51.00	84504.0
52.00	4362.0
55.00	853.0
56.00	2535.0
57.00	6053.0
58.00	398.0
61.00	1306.0
62.00	1487.0
63.00	3760.0
64.00	635.0
65.00	2301.0
66.90	336.0
68.00	1982.0
69.00	108032.0
70.00	737.0
73.00	1988.0
74.00	11440.0
75.00	17240.0
76.10	5630.0
77.00	117912.0
78.00	8112.0
79.00	8608.0
80.00	6502.0
81.00	8640.0
81.90	2178.0
83.00	1943.0
83.90	399.0
85.00	1790.0
86.00	2103.0
86.90	1007.0
88.00	401.0
91.00	2109.0
92.00	2345.0
93.00	13939.0
94.00	1029.0
94.90	511.0
95.90	573.0
97.10	419.0
98.00	10458.0
99.00	8130.0
100.00	602.0
101.00	4404.0
103.00	1574.0
104.00	3045.0
105.00	3369.0
106.10	995.0
107.00	34048.0
108.00	5077.0
109.00	1286.0
110.00	62040.0
111.00	9391.0
111.90	936.0
113.00	525.0
116.00	2041.0
117.00	27560.0
118.00	2337.0
119.00	387.0
120.00	620.0
121.90	2473.0
123.00	3555.0
123.90	1717.0
124.90	1406.0
127.00	113904.0
128.00	8317.0
129.00	43200.0
130.00	4419.0
131.00	896.0
132.00	392.0
134.00	1446.0
135.00	3833.0
136.00	1546.0
137.00	2227.0
138.10	318.0
139.90	596.0
141.00	6013.0
142.00	1938.0
142.90	1340.0
143.80	366.0
144.90	354.0
145.90	1239.0
147.00	3502.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

148.00	7851.0
149.00	1376.0
150.00	509.0
151.00	834.0
151.90	600.0
153.00	1912.0
154.00	1329.0
155.00	3521.0
156.00	5482.0
157.10	957.0
158.00	1209.0
158.90	691.0
160.00	1895.0
161.00	2668.0
162.00	921.0
162.90	351.0
163.90	372.0
165.00	2606.0
166.00	2114.0
167.00	12252.0
168.00	7041.0
168.90	1063.0
170.00	386.0
170.90	715.0
172.00	1185.0
173.00	1718.0
174.00	2734.0
175.00	5481.0
176.10	1143.0
177.00	2748.0
178.00	681.0
179.00	9540.0
180.00	6983.0
181.00	3255.0
182.00	471.0
184.00	818.0
185.00	5245.0
186.00	36336.0
187.00	10118.0
188.10	1124.0
188.90	2160.0
190.10	504.0
190.90	1407.0
192.00	3477.0
193.00	3770.0
194.00	821.0
195.10	690.0
196.00	8076.0
198.00	257216.0
198.90	16872.0
200.00	1594.0
201.50	1270.0
203.00	2144.0
204.00	10065.0
205.00	17936.0
206.00	73720.0
207.00	10848.0
208.00	3280.0
209.00	1261.0
210.30	1344.0
211.10	2865.0
215.00	985.0
216.00	1806.0
217.00	19200.0
218.00	2804.0
221.00	10934.0
221.80	2682.0
223.00	4075.0
224.00	40512.0
225.00	10754.0
226.10	1142.0
227.00	17912.0
228.00	2363.0
229.00	3512.0
230.00	437.0
231.00	1192.0
231.90	302.0
233.00	448.0
234.00	1249.0
235.00	1091.0
235.90	879.0
237.00	2116.0
238.90	615.0
239.90	785.0
240.90	916.0
242.00	2382.0
243.00	2422.0
244.00	32048.0
245.00	3977.0
246.00	6755.0
247.00	1375.0
248.00	342.0
248.90	1373.0
251.90	338.0
253.00	771.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08

255.00	172480.0
256.00	23792.0
257.00	2067.0
258.00	11313.0
259.00	1736.0
261.00	449.0
265.00	4254.0
266.00	588.0
270.90	487.0
272.10	327.0
273.00	4915.0
274.00	12895.0
275.00	74280.0
276.00	9575.0
277.00	6774.0
278.00	1369.0
281.00	514.0
283.00	966.0
284.00	548.0
285.00	1481.0
292.10	420.0
293.00	1362.0
294.10	539.0
296.00	21384.0
297.00	2915.0
300.90	347.0
302.10	363.0
303.00	2578.0
304.10	703.0
307.90	360.0
309.80	325.0
313.90	1036.0
314.90	2681.0
316.00	1453.0
321.00	749.0
323.00	7138.0
324.00	1130.0
326.90	1469.0
328.00	688.0
331.80	334.0
333.00	678.0
334.00	5035.0
334.90	1343.0
340.90	971.0
345.90	1794.0
346.90	431.0
351.90	2227.0
353.00	1788.0
354.00	2523.0
355.00	615.0
364.90	10366.0
366.00	1564.0
371.00	547.0
371.90	3674.0
373.00	912.0
383.00	1180.0
390.00	581.0
390.90	454.0
392.10	384.0
401.90	1569.0
403.00	1961.0
403.90	682.0
404.80	314.0
421.00	1777.0
422.00	2128.0
423.00	13468.0
424.00	3375.0
425.00	419.0
441.00	33328.0
442.00	228224.0
443.00	41656.0
444.00	3577.0

Instrument :
BNA_N
ClientSampleId :
DFTPP08