




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102121\
Data File : BN017018.D
Acq On : 20 Oct 2021 13:46
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 20 15:21:13 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 13 17:23:48 2021
Response via : Initial Calibration



TIC: BN017018.D\data.ms

(70) Pentachlorophenol (C)




16.604min (-0.018) 24835.61 ng

response 200166

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.13
264.00	61.60	65.70
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102121\
Data File : BN017018.D
Acq On : 20 Oct 2021 13:46
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 20 15:21:13 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 13 17:23:48 2021
Response via : Initial Calibration



TIC: BN017018.D\data.ms

(77) Benzidine

19.227min (-0.018) 0.00 ng

response 1322800

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.67
183.00	13.20	11.74
0.00	0.00	0.00