

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102323\
 Data File : BN028353.D
 Acq On : 23 Oct 2023 20:13
 Operator : MA/JU
 Sample : PB156555BS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB156555BS

Quant Time: Oct 24 03:20:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 15:14:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.886	152	3630	0.400	ng	0.00
7) Naphthalene-d8	10.703	136	11365	0.400	ng	0.00
13) Acenaphthene-d10	14.523	164	6414	0.400	ng	-0.01
19) Phenanthrene-d10	17.294	188	15063	0.400	ng	0.00
29) Chrysene-d12	21.489	240	12087	0.400	ng	0.00
35) Perylene-d12	23.940	264	10572	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.437	112	2306	0.205	ng	0.00
5) Phenol-d6	7.069	99	2970	0.214	ng	0.01
8) Nitrobenzene-d5	9.102	82	2216	0.220	ng	0.00
11) 2-Methylnaphthalene-d10	12.294	152	7278	0.426	ng	0.00
14) 2,4,6-Tribromophenol	16.053	330	557	0.178	ng	-0.01
15) 2-Fluorobiphenyl	13.155	172	5886	0.261	ng	-0.01
27) Fluoranthene-d10	19.323	212	9363	0.232	ng	0.00
31) Terphenyl-d14	19.918	244	7558	0.295	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.299	88	1824	0.416	ng	# 87
3) n-Nitrosodimethylamine	3.661	42	2648	0.553	ng	# 92
6) bis(2-Chloroethyl)ether	7.329	93	5774	0.540	ng	91
9) Naphthalene	10.746	128	15325	0.520	ng	99
10) Hexachlorobutadiene	10.960	225	2564	0.507	ng	# 100
12) 2-Methylnaphthalene	12.366	142	9620	0.459	ng	98
16) Acenaphthylene	14.256	152	13710	0.476	ng	99
17) Acenaphthene	14.587	154	9386	0.517	ng	98
18) Fluorene	15.581	166	12764	0.534	ng	100
20) 4,6-Dinitro-2-methylph...	15.767	198	537	0.654	ng	96
21) 4-Bromophenyl-phenylether	16.475	248	3693	0.476	ng	# 88
22) Hexachlorobenzene	16.562	284	3900	0.488	ng	96
23) Atrazine	16.760	200	3369	0.463	ng	97
25) Phenanthrene	17.331	178	20849	0.506	ng	100
26) Anthracene	17.430	178	18008	0.457	ng	100
28) Fluoranthene	19.356	202	23483	0.453	ng	98
30) Pyrene	19.723	202	24022	0.581	ng	100
32) Benzo(a)anthracene	21.471	228	21260	0.495	ng	98
33) Chrysene	21.524	228	21082	0.513	ng	99
34) Bis(2-ethylhexyl)phtha...	21.336	149	9589	0.292	ng	# 97
36) Indeno(1,2,3-cd)pyrene	26.510	276	19985	0.469	ng	99
37) Benzo(b)fluoranthene	23.180	252	19844	0.601	ng	# 95
38) Benzo(k)fluoranthene	23.230	252	19202	0.571	ng	# 95
39) Benzo(a)pyrene	23.832	252	15801	0.481	ng	97
40) Dibenzo(a,h)anthracene	26.522	278	16457	0.476	ng	96
41) Benzo(g,h,i)perylene	27.308	276	17128	0.479	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102323\
Data File : BN028353.D
Acq On : 23 Oct 2023 20:13
Operator : MA/JU
Sample : PB156555BS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB156555BS

Quant Time: Oct 24 03:20:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 15:14:08 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.886 min Scan# 69
 Delta R.T. 0.008 min
 Lab File: BN028353.D
 Acq: 23 Oct 2023 20:13




Instrument :
 BNA_N
 ClientSampleId :
 PB156555BS


Tgt Ion:152 Resp: 3630
 Ion Ratio Lower Upper
 152 100
 150 149.7 121.6 182.4
 115 65.7 54.3 81.5



#2
 1,4-Dioxane
 Concen: 0.416 ng
 RT: 3.299 min Scan# 58
 Delta R.T. 0.000 min
 Lab File: BN028353.D
 Acq: 23 Oct 2023 20:13

Tgt Ion: 88 Resp: 1824
 Ion Ratio Lower Upper
 88 100
 43 46.7 26.6 39.8#
 58 76.1 55.4 83.0









#3
n-Nitrosodimethylamine
Concen: 0.553 ng
RT: 3.661 min Scan# 108
Delta R.T. -0.007 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion: 42 Resp: 2648
Ion Ratio Lower Upper
42 100
74 130.3 112.6 168.8
44 6.7 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.205 ng
RT: 5.437 min Scan# 354
Delta R.T. 0.007 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion: 112 Resp: 2306
Ion Ratio Lower Upper
112 100
64 59.7 46.9 70.3
63 32.7 25.0 37.4









#5
Phenol-d6
Concen: 0.214 ng
RT: 7.069 min Scan# 580
Delta R.T. 0.015 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion: 99 Resp: 2970
Ion Ratio Lower Upper
99 100
42 21.4 15.9 23.9
71 37.3 28.6 42.8



#6
bis(2-Chloroethyl)ether
Concen: 0.540 ng
RT: 7.329 min Scan# 616
Delta R.T. 0.007 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion: 93 Resp: 5774
Ion Ratio Lower Upper
93 100
63 68.3 62.6 93.8
95 31.5 27.2 40.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.703 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion:	136	Resp:	11365
Ion Ratio	Lower	Upper	
136	100		
137	11.9	9.8	14.8
54	7.5	6.2	9.4
68	6.3	5.7	8.5



#8
Nitrobenzene-d5
Concen: 0.220 ng
RT: 9.102 min Scan# 826
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion:	82	Resp:	2216
Ion Ratio	Lower	Upper	
82	100		
128	47.4	36.8	55.2
54	48.6	37.4	56.2









#9
Naphthalene
Concen: 0.520 ng
RT: 10.746 min Scan# 981
Delta R.T. -0.011 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion:128 Resp: 15325
Ion Ratio Lower Upper
128 100
129 12.0 9.8 14.6
127 13.9 11.4 17.0



#10
Hexachlorobutadiene
Concen: 0.507 ng
RT: 10.960 min Scan# 1000
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion:225 Resp: 2564
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.3 51.6 77.4









#11
2-Methylnaphthalene-d10
Concen: 0.426 ng
RT: 12.294 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion:152 Resp: 7278
Ion Ratio Lower Upper
152 100
151 11.7 16.1 24.1#



#12
2-Methylnaphthalene
Concen: 0.459 ng
RT: 12.366 min Scan# 1279
Delta R.T. -0.004 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion:142 Resp: 9620
Ion Ratio Lower Upper
142 100
141 90.3 71.1 106.7
115 39.6 29.9 44.9





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.523 min Scan# 11

Delta R.T. -0.011 min

Lab File: BN028353.D

Acq: 23 Oct 2023 20:13

Instrument :

BNA_N

ClientSampleId :

PB156555BS





Tgt Ion:164 Resp: 6414

Ion Ratio Lower Upper

164 100

162 105.2 81.1 121.7

160 49.6 36.6 54.8



#14

2,4,6-Tribromophenol

Concen: 0.178 ng

RT: 16.053 min Scan# 1726

Delta R.T. -0.012 min

Lab File: BN028353.D

Acq: 23 Oct 2023 20:13




Tgt Ion:330 Resp: 557


Ion Ratio Lower Upper

330 100

332 95.7 74.1 111.1

141 45.2 32.1 48.1









#15
2-Fluorobiphenyl
Concen: 0.261 ng
RT: 13.155 min Scan# 1462
Delta R.T. -0.011 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion:172 Resp: 5886
Ion Ratio Lower Upper
172 100
171 36.9 28.3 42.5
170 25.1 18.7 28.1



#16
Acenaphthylene
Concen: 0.476 ng
RT: 14.256 min Scan# 1565
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion:152 Resp: 13710
Ion Ratio Lower Upper
152 100
151 20.0 15.6 23.4
153 13.2 10.6 16.0









#17
Acenaphthene
Concen: 0.517 ng
RT: 14.587 min Scan# 1596
Delta R.T. -0.011 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion:154 Resp: 9386
Ion Ratio Lower Upper
154 100
153 114.6 90.7 136.1
152 56.9 43.6 65.4



#18
Fluorene
Concen: 0.534 ng
RT: 15.581 min Scan# 1688
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion:166 Resp: 12764
Ion Ratio Lower Upper
166 100
165 99.2 79.3 118.9
167 13.6 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.294 min Scan# 1827
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion:188 Resp: 15063
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.3 7.1 10.7



#20
4,6-Dinitro-2-methylphenol
Concen: 0.654 ng
RT: 15.767 min Scan# 1703
Delta R.T. -0.037 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion:198 Resp: 537
Ion Ratio Lower Upper
198 100
51 127.8 96.1 144.1
105 93.0 75.4 113.0









#21
4-Bromophenyl-phenylether
Concen: 0.476 ng
RT: 16.475 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion:248 Resp: 3693
Ion Ratio Lower Upper
248 100
250 99.0 80.5 120.7
141 72.1 38.9 58.3#



#22
Hexachlorobenzene
Concen: 0.488 ng
RT: 16.562 min Scan# 1767
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion:284 Resp: 3900
Ion Ratio Lower Upper
284 100
142 43.9 32.1 48.1
249 32.1 25.5 38.3









#23
Atrazine
Concen: 0.463 ng
RT: 16.760 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion:200 Resp: 3369
Ion Ratio Lower Upper
200 100
173 26.4 21.6 32.4
215 46.3 39.0 58.6



#25
Phenanthrene
Concen: 0.506 ng
RT: 17.331 min Scan# 1829
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion:178 Resp: 20849
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.3 12.5 18.7









#26
 Anthracene
 Concen: 0.457 ng
 RT: 17.430 min Scan# 1838
 Delta R.T. 0.000 min
 Lab File: BN028353.D
 Acq: 23 Oct 2023 20:13




Instrument :
 BNA_N
 ClientSampleId :
 PB156555BS


Tgt Ion:178 Resp: 18008
 Ion Ratio Lower Upper
 178 100
 176 18.2 14.6 22.0
 179 15.0 12.2 18.4



#27
 Fluoranthene-d10
 Concen: 0.232 ng
 RT: 19.323 min Scan# 2099
 Delta R.T. -0.004 min
 Lab File: BN028353.D
 Acq: 23 Oct 2023 20:13

Tgt Ion:212 Resp: 9363
 Ion Ratio Lower Upper
 212 100
 106 14.1 11.5 17.3
 104 8.3 6.7 10.1









#28
Fluoranthene
Concen: 0.453 ng
RT: 19.356 min Scan# 2106
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion:202 Resp: 23483
Ion Ratio Lower Upper
202 100
101 11.2 10.1 15.1
203 17.0 13.9 20.9



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.489 min Scan# 2515
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion:240 Resp: 12087
Ion Ratio Lower Upper
240 100
120 14.2 13.4 20.0
236 27.7 22.7 34.1









#30
Pyrene
Concen: 0.581 ng
RT: 19.723 min Scan# 2187
Delta R.T. -0.004 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion:202 Resp: 24022
Ion Ratio Lower Upper
202 100
200 20.9 16.7 25.1
203 17.7 14.5 21.7



#31
Terphenyl-d14
Concen: 0.295 ng
RT: 19.918 min Scan# 2227
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion:244 Resp: 7558
Ion Ratio Lower Upper
244 100
212 9.1 6.6 10.0
122 13.6 11.7 17.5









#32
Benzo(a)anthracene
Concen: 0.495 ng
RT: 21.471 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion:228 Resp: 21260
Ion Ratio Lower Upper
228 100
226 27.7 22.9 34.3
229 19.7 16.6 24.8



#33
Chrysene
Concen: 0.513 ng
RT: 21.524 min Scan# 2519
Delta R.T. 0.000 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion:228 Resp: 21082
Ion Ratio Lower Upper
228 100
226 30.8 25.2 37.8
229 19.9 16.6 25.0









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.292 ng
 RT: 21.336 min Scan# 24
 Delta R.T. -0.009 min
 Lab File: BN028353.D
 Acq: 23 Oct 2023 20:13




Instrument :
 BNA_N
 ClientSampleId :
 PB156555BS


Tgt Ion:149 Resp: 9589
 Ion Ratio Lower Upper
 149 100
 167 27.0 22.6 34.0
 279 4.1 4.2 6.2#



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.940 min Scan# 3141
 Delta R.T. 0.003 min
 Lab File: BN028353.D
 Acq: 23 Oct 2023 20:13

Tgt Ion:264 Resp: 10572
 Ion Ratio Lower Upper
 264 100
 260 25.1 20.7 31.1
 265 72.6 38.3 57.5#









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.469 ng
RT: 26.510 min Scan# 4029
Delta R.T. 0.006 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13




Instrument :
BNA_N
ClientSampleId :
PB156555BS


Tgt Ion:276 Resp: 19985
Ion Ratio Lower Upper
276 100
138 27.1 21.3 31.9
277 25.2 20.0 30.0



#37
Benzo(b)fluoranthene
Concen: 0.601 ng
RT: 23.180 min Scan# 2881
Delta R.T. 0.003 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13


Tgt Ion:252 Resp: 19844
Ion Ratio Lower Upper
252 100
253 26.4 21.8 32.8
125 15.0 15.6 23.4#








#38
Benzo(k)fluoranthene
Concen: 0.571 ng
RT: 23.230 min Scan# 2898
Delta R.T. 0.003 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Instrument :
BNA_N
ClientSampleId :
PB156555BS







Tgt Ion:252 Resp: 19202
Ion Ratio Lower Upper
252 100
253 26.4 22.0 33.0
125 14.9 15.3 22.9#



#39
Benzo(a)pyrene
Concen: 0.481 ng
RT: 23.832 min Scan# 3104
Delta R.T. 0.003 min
Lab File: BN028353.D
Acq: 23 Oct 2023 20:13

Tgt Ion:252 Resp: 15801
Ion Ratio Lower Upper
252 100
253 29.5 22.7 34.1
125 18.8 17.0 25.6









#40
 Dibenzo(a,h)anthracene
 Concen: 0.476 ng
 RT: 26.522 min Scan# 40
 Delta R.T. 0.003 min
 Lab File: BN028353.D
 Acq: 23 Oct 2023 20:13

Instrument :
 BNA_N
 ClientSampleId :
 PB156555BS

Tgt Ion:278 Resp: 16457
 Ion Ratio Lower Upper
 278 100
 139 20.3 18.8 28.2
 279 28.7 23.5 35.3



#41
 Benzo(g,h,i)perylene
 Concen: 0.479 ng
 RT: 27.308 min Scan# 4293
 Delta R.T. 0.003 min
 Lab File: BN028353.D
 Acq: 23 Oct 2023 20:13

Tgt Ion:276 Resp: 17128
 Ion Ratio Lower Upper
 276 100
 277 25.6 21.7 32.5
 138 25.7 21.7 32.5

