

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102323\
 Data File : BN028346.D
 Acq On : 23 Oct 2023 15:17
 Operator : MA/JU
 Sample : PB156364BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB156364BS

Quant Time: Oct 23 16:05:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 23 15:14:08 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.886	152	3348	0.400	ng	0.00
7) Naphthalene-d8	10.703	136	10469	0.400	ng	# 0.00
13) Acenaphthene-d10	14.534	164	6105	0.400	ng	0.00
19) Phenanthrene-d10	17.294	188	13898	0.400	ng	# 0.00
29) Chrysene-d12	21.489	240	10642	0.400	ng	0.00
35) Perylene-d12	23.943	264	9101	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.437	112	3179	0.306	ng	0.00
5) Phenol-d6	7.077	99	3919	0.306	ng	0.02
8) Nitrobenzene-d5	9.102	82	3010	0.324	ng	0.00
11) 2-Methylnaphthalene-d10	12.302	152	8221	0.522	ng	0.00
14) 2,4,6-Tribromophenol	16.065	330	742	0.249	ng	0.00
15) 2-Fluorobiphenyl	13.165	172	8507	0.396	ng	0.00
27) Fluoranthene-d10	19.333	212	12827	0.345	ng	0.00
31) Terphenyl-d14	19.923	244	10444	0.462	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.299	88	1231	0.304	ng	# 75
3) n-Nitrosodimethylamine	3.661	42	1766	0.400	ng	# 91
6) bis(2-Chloroethyl)ether	7.337	93	3554	0.360	ng	93
9) Naphthalene	10.757	128	10536	0.388	ng	99
10) Hexachlorobutadiene	10.960	225	1781	0.382	ng	# 100
12) 2-Methylnaphthalene	12.374	142	6707	0.348	ng	98
16) Acenaphthylene	14.266	152	9325	0.340	ng	99
17) Acenaphthene	14.598	154	6568	0.380	ng	97
18) Fluorene	15.593	166	8979	0.395	ng	100
20) 4,6-Dinitro-2-methylph...	15.792	198	276	0.461	ng	# 1
21) 4-Bromophenyl-phenylether	16.487	248	2483	0.347	ng	# 95
22) Hexachlorobenzene	16.562	284	2623	0.356	ng	97
23) Atrazine	16.773	200	2017	0.301	ng	99
25) Phenanthrene	17.331	178	14393	0.378	ng	100
26) Anthracene	17.430	178	11808	0.325	ng	100
28) Fluoranthene	19.360	202	15566	0.326	ng	98
30) Pyrene	19.727	202	15601	0.428	ng	100
32) Benzo(a)anthracene	21.480	228	12759	0.337	ng	98
33) Chrysene	21.533	228	13904	0.384	ng	98
34) Bis(2-ethylhexyl)phtha...	21.345	149	6032	0.209	ng	# 97
36) Indeno(1,2,3-cd)pyrene	26.522	276	12922	0.352	ng	100
37) Benzo(b)fluoranthene	23.183	252	12318	0.433	ng	96
38) Benzo(k)fluoranthene	23.236	252	12083	0.418	ng	96
39) Benzo(a)pyrene	23.835	252	10162	0.359	ng	94
40) Dibenzo(a,h)anthracene	26.533	278	10806	0.363	ng	96
41) Benzo(g,h,i)perylene	27.323	276	11296	0.367	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102323\
Data File : BN028346.D
Acq On : 23 Oct 2023 15:17
Operator : MA/JU
Sample : PB156364BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB156364BS


Quant Time: Oct 23 16:05:34 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN101223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 23 15:14:08 2023
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.886 min Scan# 694
 Delta R.T. 0.008 min
 Lab File: BN028346.D
 Acq: 23 Oct 2023 15:17

Instrument :
 BNA_N
 ClientSampleId :
 PB156364BS







Tgt Ion:152 Resp: 3348
 Ion Ratio Lower Upper
 152 100
 150 155.2 121.6 182.4
 115 67.1 54.3 81.5



#2
 1,4-Dioxane
 Concen: 0.304 ng
 RT: 3.299 min Scan# 58
 Delta R.T. 0.000 min
 Lab File: BN028346.D
 Acq: 23 Oct 2023 15:17

Tgt Ion: 88 Resp: 1231
 Ion Ratio Lower Upper
 88 100
 43 62.1 26.6 39.8
 58 79.9 55.4 83.0









#3
n-Nitrosodimethylamine
Concen: 0.400 ng
RT: 3.661 min Scan# 108
Delta R.T. -0.007 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17




Instrument :
BNA_N
ClientSampleId :
PB156364BS


Tgt Ion: 42 Resp: 1766
Ion Ratio Lower Upper
42 100
74 129.8 112.6 168.8
44 5.9 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.306 ng
RT: 5.437 min Scan# 354
Delta R.T. 0.007 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17

Tgt Ion: 112 Resp: 3179
Ion Ratio Lower Upper
112 100
64 60.4 46.9 70.3
63 33.6 25.0 37.4









#5
Phenol-d6
Concen: 0.306 ng
RT: 7.077 min Scan# 51
Delta R.T. 0.022 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17




Instrument :
BNA_N
ClientSampleId :
PB156364BS


Tgt Ion: 99 Resp: 3919
Ion Ratio Lower Upper
99 100
42 20.5 15.9 23.9
71 37.5 28.6 42.8



#6
bis(2-Chloroethyl)ether
Concen: 0.360 ng
RT: 7.337 min Scan# 617
Delta R.T. 0.015 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17

Tgt Ion: 93 Resp: 3554
Ion Ratio Lower Upper
93 100
63 70.1 62.6 93.8
95 32.3 27.2 40.8





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.703 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17

Instrument :

BNA_N

ClientSampleId :

PB156364BS

Tgt Ion:136 Resp: 10469





Ion Ratio Lower Upper

136 100

137 12.1 9.8 14.8

54 9.9 6.2 9.4#

68 7.6 5.7 8.5



#8

Nitrobenzene-d5

Concen: 0.324 ng

RT: 9.102 min Scan# 826

Delta R.T. 0.000 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17




Tgt Ion: 82 Resp: 3010


Ion Ratio Lower Upper

82 100

128 39.4 36.8 55.2

54 52.2 37.4 56.2









#9
Naphthalene
Concen: 0.388 ng
RT: 10.757 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17




Instrument :
BNA_N
ClientSampleId :
PB156364BS


Tgt Ion:	128	Resp:	10536
Ion Ratio	Lower	Upper	
128	100		
129	12.3	9.8	14.6
127	13.9	11.4	17.0



#10
Hexachlorobutadiene
Concen: 0.382 ng
RT: 10.960 min Scan# 1000
Delta R.T. 0.000 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17

Tgt Ion:	225	Resp:	1781
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.1	51.6	77.4





#11

2-Methylnaphthalene-d10

Concen: 0.522 ng

RT: 12.302 min Scan# 11

Delta R.T. 0.008 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17

Instrument :

BNA_N

ClientSampleId :





PB156364BS

Tgt Ion:152 Resp: 8221

Ion Ratio Lower Upper

152 100

151 15.2 16.1 24.1#



#12

2-Methylnaphthalene

Concen: 0.348 ng

RT: 12.374 min Scan# 1281

Delta R.T. 0.004 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17




Tgt Ion:142 Resp: 6707


Ion Ratio Lower Upper

142 100

141 89.3 71.1 106.7

115 41.0 29.9 44.9





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.534 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17

Instrument :

BNA_N

ClientSampleId :

PB156364BS





Tgt Ion:164 Resp: 6105

Ion Ratio Lower Upper

164 100

162 103.7 81.1 121.7

160 48.8 36.6 54.8



#14

2,4,6-Tribromophenol

Concen: 0.249 ng

RT: 16.065 min Scan# 1727

Delta R.T. 0.000 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17




Tgt Ion:330 Resp: 742


Ion Ratio Lower Upper

330 100

332 97.4 74.1 111.1

141 48.4 32.1 48.1#





#15

2-Fluorobiphenyl

Concen: 0.396 ng

RT: 13.165 min Scan# 1463

Delta R.T. 0.000 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17

Instrument :

BNA_N

ClientSampleId :

PB156364BS





Tgt Ion:172 Resp: 8507

Ion Ratio Lower Upper

172 100

171 35.9 28.3 42.5

170 23.8 18.7 28.1



#16

Acenaphthylene

Concen: 0.340 ng

RT: 14.266 min Scan# 1566

Delta R.T. 0.011 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17




Tgt Ion:152 Resp: 9325


Ion Ratio Lower Upper

152 100

151 20.1 15.6 23.4

153 13.3 10.6 16.0





#17

Acenaphthene

Concen: 0.380 ng

RT: 14.598 min Scan# 1597

Delta R.T. 0.000 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17

Instrument :

BNA_N

ClientSampleId :

PB156364BS





Tgt Ion:154 Resp: 6568

Ion Ratio Lower Upper

154 100

153 115.8 90.7 136.1

152 57.0 43.6 65.4



#18

Fluorene

Concen: 0.395 ng

RT: 15.593 min Scan# 1689

Delta R.T. 0.013 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17




Tgt Ion:166 Resp: 8979


Ion Ratio Lower Upper

166 100

165 98.9 79.3 118.9

167 13.3 10.9 16.3





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.294 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17

Instrument :

BNA_N

ClientSampleId :

PB156364BS





Tgt Ion:188 Resp: 13898

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.8 7.1 10.7#



#20

4,6-Dinitro-2-methylphenol

Concen: 0.461 ng

RT: 15.792 min Scan# 1705

Delta R.T. -0.012 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17




Tgt Ion:198 Resp: 276


Ion Ratio Lower Upper

198 100

51 343.8 96.1 144.1#

105 101.0 75.4 113.0









#21
4-Bromophenyl-phenylether
Concen: 0.347 ng
RT: 16.487 min Scan# 1761
Delta R.T. 0.013 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17




Instrument :
BNA_N
ClientSampleId :
PB156364BS


Tgt Ion:248 Resp: 2483
Ion Ratio Lower Upper
248 100
250 101.3 80.5 120.7
141 59.0 38.9 58.3#



#22
Hexachlorobenzene
Concen: 0.356 ng
RT: 16.562 min Scan# 1767
Delta R.T. 0.000 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17

Tgt Ion:284 Resp: 2623
Ion Ratio Lower Upper
284 100
142 43.7 32.1 48.1
249 31.8 25.5 38.3









#23
Atrazine
Concen: 0.301 ng
RT: 16.773 min Scan# 11
Delta R.T. 0.013 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17




Instrument :
BNA_N
ClientSampleId :
PB156364BS


Tgt Ion:200 Resp: 2017
Ion Ratio Lower Upper
200 100
173 28.5 21.6 32.4
215 49.3 39.0 58.6



#25
Phenanthrene
Concen: 0.378 ng
RT: 17.331 min Scan# 1829
Delta R.T. 0.000 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17

Tgt Ion:178 Resp: 14393
Ion Ratio Lower Upper
178 100
176 19.1 15.3 22.9
179 15.5 12.5 18.7









#26
 Anthracene
 Concen: 0.325 ng
 RT: 17.430 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BN028346.D
 Acq: 23 Oct 2023 15:17




Instrument :
 BNA_N
 ClientSampleId :
 PB156364BS


Tgt Ion:178 Resp: 11808
 Ion Ratio Lower Upper
 178 100
 176 18.1 14.6 22.0
 179 15.4 12.2 18.4



#27
 Fluoranthene-d10
 Concen: 0.345 ng
 RT: 19.333 min Scan# 2101
 Delta R.T. 0.005 min
 Lab File: BN028346.D
 Acq: 23 Oct 2023 15:17

Tgt Ion:212 Resp: 12827
 Ion Ratio Lower Upper
 212 100
 106 13.7 11.5 17.3
 104 8.2 6.7 10.1





#28

Fluoranthene

Concen: 0.326 ng

RT: 19.360 min Scan# 2108

Delta R.T. 0.005 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17

Instrument :

BNA_N

ClientSampleId :

PB156364BS





Tgt Ion:202 Resp: 15566

Ion Ratio Lower Upper

202 100

101 11.2 10.1 15.1

203 17.1 13.9 20.9



#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.489 min Scan# 2515

Delta R.T. 0.000 min

Lab File: BN028346.D

Acq: 23 Oct 2023 15:17




Tgt Ion:240 Resp: 10642


Ion Ratio Lower Upper

240 100

120 16.9 13.4 20.0

236 28.4 22.7 34.1









#30
 Pyrene
 Concen: 0.428 ng
 RT: 19.727 min Scan# 2186
 Delta R.T. 0.000 min
 Lab File: BN028346.D
 Acq: 23 Oct 2023 15:17




Instrument :
 BNA_N
 ClientSampleId :
 PB156364BS


Tgt Ion:202 Resp: 15601
 Ion Ratio Lower Upper
 202 100
 200 20.7 16.7 25.1
 203 17.9 14.5 21.7



#31
 Terphenyl-d14
 Concen: 0.462 ng
 RT: 19.923 min Scan# 2228
 Delta R.T. 0.005 min
 Lab File: BN028346.D
 Acq: 23 Oct 2023 15:17

Tgt Ion:244 Resp: 10444
 Ion Ratio Lower Upper
 244 100
 212 9.0 6.6 10.0
 122 14.2 11.7 17.5









#32
Benzo(a)anthracene
Concen: 0.337 ng
RT: 21.480 min Scan# 2148
Delta R.T. 0.009 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17




Instrument :
BNA_N
ClientSampleId :
PB156364BS


Tgt Ion:228 Resp: 12759
Ion Ratio Lower Upper
228 100
226 26.9 22.9 34.3
229 20.3 16.6 24.8



#33
Chrysene
Concen: 0.384 ng
RT: 21.533 min Scan# 2520
Delta R.T. 0.009 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17

Tgt Ion:228 Resp: 13904
Ion Ratio Lower Upper
228 100
226 29.7 25.2 37.8
229 20.7 16.6 25.0









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.209 ng
 RT: 21.345 min Scan# 2499
 Delta R.T. 0.000 min
 Lab File: BN028346.D
 Acq: 23 Oct 2023 15:17




Instrument :
 BNA_N
 ClientSampleId :
 PB156364BS


Tgt Ion:149 Resp: 6032
 Ion Ratio Lower Upper
 149 100
 167 26.8 22.6 34.0
 279 3.6 4.2 6.2#



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.943 min Scan# 3142
 Delta R.T. 0.006 min
 Lab File: BN028346.D
 Acq: 23 Oct 2023 15:17

Tgt Ion:264 Resp: 9101
 Ion Ratio Lower Upper
 264 100
 260 25.3 20.7 31.1
 265 81.0 38.3 57.5#









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.352 ng
RT: 26.522 min Scan# 4029
Delta R.T. 0.018 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17




Instrument :
BNA_N
ClientSampleId :
PB156364BS


Tgt Ion:276 Resp: 12922
Ion Ratio Lower Upper
276 100
138 26.5 21.3 31.9
277 25.0 20.0 30.0



#37
Benzo(b)fluoranthene
Concen: 0.433 ng
RT: 23.183 min Scan# 2882
Delta R.T. 0.006 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17


Tgt Ion:252 Resp: 12318
Ion Ratio Lower Upper
252 100
253 28.4 21.8 32.8
125 16.7 15.6 23.4







#38
Benzo(k)fluoranthene
Concen: 0.418 ng
RT: 23.236 min Scan# 2905
Delta R.T. 0.009 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17

Instrument :
BNA_N
ClientSampleId :
PB156364BS







Tgt Ion:252 Resp: 12083
Ion Ratio Lower Upper
252 100
253 29.0 22.0 33.0
125 17.1 15.3 22.9



#39
Benzo(a)pyrene
Concen: 0.359 ng
RT: 23.835 min Scan# 3105
Delta R.T. 0.006 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17

Tgt Ion:252 Resp: 10162
Ion Ratio Lower Upper
252 100
253 33.3 22.7 34.1
125 22.0 17.0 25.6









#40
Dibenzo(a,h)anthracene
Concen: 0.363 ng
RT: 26.533 min Scan# 4034
Delta R.T. 0.015 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17

Instrument :
BNA_N
ClientSampleId :
PB156364BS

Tgt Ion:278 Resp: 10806
Ion Ratio Lower Upper
278 100
139 20.4 18.8 28.2
279 31.0 23.5 35.3



#41
Benzo(g,h,i)perylene
Concen: 0.367 ng
RT: 27.323 min Scan# 4298
Delta R.T. 0.018 min
Lab File: BN028346.D
Acq: 23 Oct 2023 15:17

Tgt Ion:276 Resp: 11296
Ion Ratio Lower Upper
276 100
277 25.2 21.7 32.5
138 26.3 21.7 32.5

