

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102418\
 Data File : BN003290.D
 Acq On : 25 Oct 2018 02:20
 Operator : MJ/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 25 02:52:43 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-BN102418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 24 14:21:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.545	0.514	5.7	111	0.00
3	Pyridine	1.246	1.260	-1.1	109	0.00
4	n-Nitrosodimethylamine	0.459	0.446	2.8	113	0.00
5 S	2-Fluorophenol	1.148	1.121	2.4	110	0.00
6	Aniline	2.039	2.052	-0.6	113	0.00
7 S	Phenol-d6	1.633	1.628	0.3	112	0.00
8	2-Chlorophenol	1.414	1.385	2.1	112	0.00
9	Benzaldehyde	0.948	0.918	3.2	110	0.00
10 C	Phenol	1.723	1.726	-0.2	112	0.00
11	bis(2-Chloroethyl)ether	1.407	1.399	0.6	113	0.00
12	1,3-Dichlorobenzene	1.572	1.532	2.5	111	0.00
13 C	1,4-Dichlorobenzene	1.619	1.581	2.3	112	0.00
14	1,2-Dichlorobenzene	1.571	1.537	2.2	111	0.00
15	Benzyl Alcohol	1.177	1.200	-2.0	118	0.00
16	2,2'-oxybis(1-Chloropropane	2.063	2.040	1.1	114	0.00
17	2-Methylphenol	1.178	1.182	-0.3	113	0.00
18	Hexachloroethane	0.567	0.544	4.1	111	0.00
19 P	n-Nitroso-di-n-propylamine	1.177	1.198	-1.8	115	0.00
20	3+4-Methylphenols	1.673	1.688	-0.9	114	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
22	Acetophenone	0.485	0.482	0.6	115	0.00
23 S	Nitrobenzene-d5	0.349	0.341	2.3	113	0.00
24	Nitrobenzene	0.356	0.354	0.6	115	0.00
25	Isophorone	0.612	0.613	-0.2	116	0.00
26 C	2-Nitrophenol	0.187	0.188	-0.5	114	0.00
27	2,4-Dimethylphenol	0.263	0.258	1.9	113	0.00
28	bis(2-Chloroethoxy)methane	0.412	0.409	0.7	115	0.00
29 C	2,4-Dichlorophenol	0.301	0.303	-0.7	114	0.00
30	1,2,4-Trichlorobenzene	0.342	0.334	2.3	114	0.00
31	Naphthalene	0.977	0.950	2.8	114	0.00
32	Benzoic acid	0.209	0.197	5.7	108	0.00
33	4-Chloroaniline	0.432	0.435	-0.7	118	0.00
34 C	Hexachlorobutadiene	0.204	0.199	2.5	114	0.00
35	Caprolactam	0.118	0.122	-3.4	115	0.00
36 C	4-Chloro-3-methylphenol	0.335	0.338	-0.9	116	0.00
37	2-Methylnaphthalene	0.698	0.682	2.3	115	0.00
38	1-Methylnaphthalene	0.682	0.666	2.3	115	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	0.662	0.644	2.7	115	0.00
41 P	Hexachlorocyclopentadiene	0.229	0.240	-4.8	122	0.00
42 S	2,4,6-Tribromophenol	0.269	0.270	-0.4	118	0.00
43 C	2,4,6-Trichlorophenol	0.418	0.412	1.4	115	0.00
44	2,4,5-Trichlorophenol	0.436	0.433	0.7	119	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102418\
 Data File : BN003290.D
 Acq On : 25 Oct 2018 02:20
 Operator : MJ/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 25 02:52:43 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-BN102418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 24 14:21:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.474	1.416	3.9	115	0.00
46	1,1'-Biphenyl	1.755	1.680	4.3	115	0.00
47	2-Chloronaphthalene	1.295	1.244	3.9	115	0.00
48	2-Nitroaniline	0.380	0.380	0.0	115	0.00
49	Acenaphthylene	1.982	1.907	3.8	115	0.00
50	Dimethylphthalate	1.617	1.555	3.8	117	0.00
51	2,6-Dinitrotoluene	0.367	0.363	1.1	117	0.00
52 C	Acenaphthene	1.201	1.153	4.0	116	0.00
53	3-Nitroaniline	0.392	0.397	-1.3	118	0.00
54 P	2,4-Dinitrophenol	0.180	0.171	5.0	111	0.00
55	Dibenzofuran	1.954	1.878	3.9	116	0.00
56 P	4-Nitrophenol	0.247	0.241	2.4	112	0.00
57	2,4-Dinitrotoluene	0.498	0.491	1.4	115	0.00
58	Fluorene	1.508	1.439	4.6	115	0.00
59	2,3,4,6-Tetrachlorophenol	0.383	0.388	-1.3	121	0.00
60	Diethylphthalate	1.641	1.588	3.2	116	0.00
61	4-Chlorophenyl-phenylether	0.838	0.806	3.8	117	0.00
62	4-Nitroaniline	0.383	0.391	-2.1	115	0.00
63	Azobenzene	1.517	1.477	2.6	117	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
65	4,6-Dinitro-2-methylphenol	0.140	0.134	4.3	115	0.00
66 c	n-Nitrosodiphenylamine	0.626	0.602	3.8	116	0.00
67	4-Bromophenyl-phenylether	0.225	0.217	3.6	115	0.00
68	Hexachlorobenzene	0.257	0.248	3.5	115	0.00
69	Atrazine	0.222	0.217	2.3	115	0.00
70 C	Pentachlorophenol	0.107	0.127	-18.7	135	0.00
71	Phenanthrene	1.051	1.005	4.4	115	0.00
72	Anthracene	1.049	1.014	3.3	115	0.00
73	Carbazole	1.087	1.051	3.3	117	0.00
74	Di-n-butylphthalate	1.267	1.231	2.8	116	0.00
75 C	Fluoranthene	1.239	1.183	4.5	116	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
77	Benzidine	0.586	0.627	-7.0	128	0.00
78	Pyrene	1.175	1.133	3.6	116	0.00
79 S	Terphenyl-d14	0.935	0.901	3.6	115	0.00
80	Butylbenzylphthalate	0.546	0.528	3.3	115	0.00
81	Benzo(a)anthracene	1.165	1.110	4.7	113	0.00
82	3,3'-Dichlorobenzidine	0.482	0.457	5.2	112	0.00
83	Chrysene	1.120	1.056	5.7	113	0.00
84	Bis(2-ethylhexyl)phthalate	0.809	0.770	4.8	114	0.00
85 c	Di-n-octyl phthalate	1.389	1.304	6.1	111	0.00
86	Indeno(1,2,3-cd)pyrene	1.283	1.134	11.6	102	0.00
87 I	Perylene-d12	1.000	1.000	0.0	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102418\
Data File : BN003290.D
Acq On : 25 Oct 2018 02:20
Operator : MJ/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC040

Quant Time: Oct 25 02:52:43 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-BN102418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 24 14:21:25 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.101	1.086	1.4	113	0.00
89	Benzo(k)fluoranthene	1.080	1.012	6.3	110	0.00
90 C	Benzo(a)pyrene	1.049	0.999	4.8	110	0.00
91	Dibenzo(a,h)anthracene	1.039	0.940	9.5	102	0.00
92	Benzo(g,h,i)perylene	0.997	0.891	10.6	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0