

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN102618\
 Data File : BN003322.D
 Acq On : 27 Oct 2018 01:25
 Operator : MJ/SJ
 Sample : J5521-05
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 AOC8-03B

Quant Time: Oct 27 05:49:32 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 26 15:17:57 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.86	152	14409	0.40	ng	0.00
7) Naphthalene-d8	10.66	136	60897	0.40	ng	-0.01
13) Acenaphthene-d10	14.49	164	35801	0.40	ng	-0.01
19) Phenanthrene-d10	17.24	188	89138	0.40	ng	0.00
27) Chrysene-d12	21.42	240	108101	0.40	ng	0.00
34) Perylene-d12	23.74	264	112390	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.43	112	14459	0.41	ng	0.00
5) Phenol-d6	7.01	99	18352	0.41	ng	0.00
8) Nitrobenzene-d5	9.02	82	11660	0.44	ng	0.00
11) 2-Methylnaphthalene-d10	12.25	152	31291	0.32	ng	0.00
14) 2,4,6-Tribromophenol	15.98	330	4142	0.29	ng	0.00
15) 2-Fluorobiphenyl	13.12	172	46644	0.30	ng	-0.01
25) Fluoranthene-d10	19.26	212	76098	0.29	ng	0.00
29) Terphenyl-d14	19.86	244	65907	0.26	ng	0.00
Target Compounds						
22) Pentachlorophenol	16.88	266	321	0.022	ng	Qvalue # 85
32) Bis(2-ethylhexyl)phthalate	21.32	149	46932	0.560	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN102618\
Data File : BN003322.D
Acq On : 27 Oct 2018 01:25
Operator : MJ/SJ
Sample : J5521-05
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
AOC8-03B

Quant Time: Oct 27 05:49:32 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 26 15:17:57 2018
Response via : Initial Calibration

