

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102618\
 Data File : BN003305.D
 Acq On : 26 Oct 2018 15:21
 Operator : MJ/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN102618

Quant Time: Oct 26 15:59:19 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 26 15:17:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.518	0.479	7.5	96	0.00
3	n-Nitrosodimethylamine	0.238	0.216	9.2	91	0.00
4 S	2-Fluorophenol	0.991	0.940	5.1	95	0.00
5 S	Phenol-d6	1.251	1.173	6.2	95	0.00
6	bis(2-Chloroethyl)ether	1.116	1.114	0.2	97	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.175	0.173	1.1	105	0.00
9	Naphthalene	0.979	0.947	3.3	96	0.00
10	Hexachlorobutadiene	0.179	0.175	2.2	97	0.00
11 SURR	2-Methylnaphthalene-d10	0.651	0.620	4.8	95	0.00
12	2-Methylnaphthalene	0.684	0.653	4.5	95	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.159	0.133	16.4	93	0.00
15 S	2-Fluorobiphenyl	1.721	1.650	4.1	96	0.00
16	Acenaphthylene	1.783	1.611	9.6	91	0.00
17	Acenaphthene	1.236	1.184	4.2	94	0.00
18	Fluorene	1.546	1.485	3.9	94	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
20	4-Bromophenyl-phenylether	0.232	0.220	5.2	94	-0.01
21	Hexachlorobenzene	0.244	0.239	2.0	95	0.00
22	Pentachlorophenol	0.066	0.062	6.1	95	0.00
23	Phenanthrene	1.097	1.045	4.7	94	0.00
24	Anthracene	0.941	0.844	10.3	90	0.00
25 SURR	Fluoranthene-d10	1.163	1.083	6.9	93	0.00
26	Fluoranthene	1.260	1.177	6.6	93	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
28	Pvrene	1.106	1.012	8.5	93	0.00
29 S	Terphenyl-d14	0.930	0.845	9.1	94	0.00
30	Benzo(a)anthracene	1.048	0.898	14.3	91	0.00
31	Chrysene	1.131	1.054	6.8	94	0.00
32	Bis(2-ethylhexyl)phthalate	0.310	0.250	19.4	87	0.00
33	Indeno(1,2,3-cd)pyrene	1.170	1.151	1.6	99	0.00
34 I	Perylene-d12	1.000	1.000	0.0	96	0.00
35	Benzo(b)fluoranthene	1.232	1.188	3.6	98	0.00
36	Benzo(k)fluoranthene	1.236	1.148	7.1	95	0.00
37 C	Benzo(a)pyrene	1.136	1.084	4.6	96	0.00
38	Dibenzo(a,h)anthracene	1.058	0.995	6.0	99	0.00
39	Benzo(a,h,i)perylene	1.058	0.987	6.7	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102618\
Data File : BN003305.D
Acq On : 26 Oct 2018 15:21
Operator : MJ/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN102618

Quant Time: Oct 26 15:59:19 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 26 15:17:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			