

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102618\
 Data File : BN003316.D
 Acq On : 26 Oct 2018 21:48
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 27 00:44:54 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 26 15:17:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.518	0.462	10.8	97	0.00
3	n-Nitrosodimethylamine	0.238	0.211	11.3	94	0.00
4 S	2-Fluorophenol	0.991	0.853	13.9	91	0.00
5 S	Phenol-d6	1.251	1.061	15.2	90	0.00
6	bis(2-Chloroethyl)ether	1.116	1.049	6.0	96	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	-0.01
8 S	Nitrobenzene-d5	0.175	0.173	1.1	109	0.00
9	Naphthalene	0.979	0.913	6.7	97	-0.01
10	Hexachlorobutadiene	0.179	0.169	5.6	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.651	0.593	8.9	95	0.00
12	2-Methylnaphthalene	0.684	0.626	8.5	95	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	0.159	0.124	22.0	93	0.00
15 S	2-Fluorobiphenyl	1.721	1.542	10.4	96	-0.01
16	Acenaphthylene	1.783	1.436	19.5	86	0.00
17	Acenaphthene	1.236	1.100	11.0	93	0.00
18	Fluorene	1.546	1.358	12.2	92	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
20	4-Bromophenyl-phenylether	0.232	0.205	11.6	94	-0.01
21	Hexachlorobenzene	0.244	0.228	6.6	97	0.00
22	Pentachlorophenol	0.066	0.059	10.6	96	0.00
23	Phenanthrene	1.097	0.986	10.1	95	0.00
24	Anthracene	0.941	0.768	18.4	87	-0.01
25 SURR	Fluoranthene-d10	1.163	0.977	16.0	89	0.00
26	Fluoranthene	1.260	1.073	14.8	90	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
28	Pvrene	1.106	0.984	11.0	89	0.00
29 S	Terphenyl-d14	0.930	0.824	11.4	91	0.00
30	Benzo(a)anthracene	1.048	0.843	19.6	84	-0.01
31	Chrysene	1.131	1.020	9.8	90	-0.01
32	Bis(2-ethylhexyl)phthalate	0.310	0.221	28.7#	76	-0.01
33	Indeno(1,2,3-cd)pyrene	1.170	0.977	16.5	83	0.00
34 I	Perylene-d12	1.000	1.000	0.0	89	0.00
35	Benzo(b)fluoranthene	1.232	1.131	8.2	87	0.00
36	Benzo(k)fluoranthene	1.236	1.148	7.1	89	0.00
37 C	Benzo(a)pyrene	1.136	1.005	11.5	83	0.00
38	Dibenzo(a,h)anthracene	1.058	0.904	14.6	83	0.00
39	Benzo(a,h,i)perylene	1.058	0.910	14.0	84	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102618\
Data File : BN003316.D
Acq On : 26 Oct 2018 21:48
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 27 00:44:54 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 26 15:17:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			