

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102618\
 Data File : BN003318.D
 Acq On : 26 Oct 2018 23:05
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 26 23:45:07 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 26 15:17:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	97	0.00
2	1,4-Dioxane	0.400	0.356	11.0	93	0.00
3	n-Nitrosodimethylamine	0.400	0.347	13.3	88	0.00
4 S	2-Fluorophenol	0.400	0.348	13.0	88	0.00
5 S	Phenol-d6	0.400	0.342	14.5	87	0.00
6	bis(2-Chloroethyl)ether	0.400	0.374	6.5	91	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	-0.01
8 S	Nitrobenzene-d5	0.400	0.405	-1.3	106	0.00
9	Naphthalene	0.400	0.372	7.0	91	-0.01
10	Hexachlorobutadiene	0.400	0.380	5.0	93	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.364	9.0	90	0.00
12	2-Methylnaphthalene	0.400	0.364	9.0	90	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.400	0.318	20.5	89	0.00
15 S	2-Fluorobiphenyl	0.400	0.363	9.3	92	-0.01
16	Acenaphthylene	0.400	0.325	18.8	82	0.00
17	Acenaphthene	0.400	0.356	11.0	87	0.00
18	Fluorene	0.400	0.356	11.0	87	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	95	0.00
20	4-Bromophenyl-phenylether	0.400	0.351	12.3	88	-0.01
21	Hexachlorobenzene	0.400	0.373	6.8	92	0.00
22	Pentachlorophenol	0.400	0.364	9.0	94	0.00
23	Phenanthrene	0.400	0.357	10.8	89	0.00
24	Anthracene	0.400	0.322	19.5	81	0.00
25 SURR	Fluoranthene-d10	0.400	0.337	15.8	85	0.00
26	Fluoranthene	0.400	0.342	14.5	86	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	92	0.00
28	Pvrene	0.400	0.348	13.0	85	0.00
29 S	Terphenyl-d14	0.400	0.348	13.0	87	0.00
30	Benzo(a)anthracene	0.400	0.320	20.0	82	0.00
31	Chrysene	0.400	0.364	9.0	89	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.283	29.3#	74	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.356	11.0	87	0.00
34 I	Perylene-d12	0.400	0.400	0.0	89	0.00
35	Benzo(b)fluoranthene	0.400	0.366	8.5	87	0.00
36	Benzo(k)fluoranthene	0.400	0.362	9.5	87	0.00
37 C	Benzo(a)pyrene	0.400	0.356	11.0	84	0.00
38	Dibenzo(a,h)anthracene	0.400	0.353	11.8	87	0.00
39	Benzo(a,h,i)perylene	0.400	0.355	11.3	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102618\
Data File : BN003318.D
Acq On : 26 Oct 2018 23:05
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 26 23:45:07 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 26 15:17:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				