

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102618\
 Data File : BN003318.D
 Acq On : 26 Oct 2018 23:05
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 26 23:45:07 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 26 15:17:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.518	0.461	11.0	93	0.00
3	n-Nitrosodimethylamine	0.238	0.207	13.0	88	0.00
4 S	2-Fluorophenol	0.991	0.861	13.1	88	0.00
5 S	Phenol-d6	1.251	1.068	14.6	87	0.00
6	bis(2-Chloroethyl)ether	1.116	1.042	6.6	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	95	-0.01
8 S	Nitrobenzene-d5	0.175	0.178	-1.7	106	0.00
9	Naphthalene	0.979	0.910	7.0	91	-0.01
10	Hexachlorobutadiene	0.179	0.170	5.0	93	0.00
11 SURR	2-Methylnaphthalene-d10	0.651	0.592	9.1	90	0.00
12	2-Methylnaphthalene	0.684	0.623	8.9	90	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.159	0.126	20.8	89	0.00
15 S	2-Fluorobiphenyl	1.721	1.562	9.2	92	-0.01
16	Acenaphthylene	1.783	1.447	18.8	82	0.00
17	Acenaphthene	1.236	1.098	11.2	87	0.00
18	Fluorene	1.546	1.377	10.9	87	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	4-Bromophenyl-phenylether	0.232	0.203	12.5	88	-0.01
21	Hexachlorobenzene	0.244	0.228	6.6	92	0.00
22	Pentachlorophenol	0.066	0.060	9.1	94	0.00
23	Phenanthrene	1.097	0.979	10.8	89	0.00
24	Anthracene	0.941	0.758	19.4	81	0.00
25 SURR	Fluoranthene-d10	1.163	0.981	15.6	85	0.00
26	Fluoranthene	1.260	1.078	14.4	86	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
28	Pvrene	1.106	0.962	13.0	85	0.00
29 S	Terphenyl-d14	0.930	0.811	12.8	87	0.00
30	Benzo(a)anthracene	1.048	0.840	19.8	82	0.00
31	Chrysene	1.131	1.028	9.1	89	0.00
32	Bis(2-ethylhexyl)phthalate	0.310	0.219	29.4#	74	-0.01
33	Indeno(1,2,3-cd)pyrene	1.170	1.043	10.9	87	0.00
34 I	Perylene-d12	1.000	1.000	0.0	89	0.00
35	Benzo(b)fluoranthene	1.232	1.128	8.4	87	0.00
36	Benzo(k)fluoranthene	1.236	1.119	9.5	87	0.00
37 C	Benzo(a)pyrene	1.136	1.011	11.0	84	0.00
38	Dibenzo(a,h)anthracene	1.058	0.933	11.8	87	0.00
39	Benzo(a,h,i)perylene	1.058	0.941	11.1	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102618\
Data File : BN003318.D
Acq On : 26 Oct 2018 23:05
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 26 23:45:07 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 26 15:17:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			