

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102618\
 Data File : BN003327.D
 Acq On : 27 Oct 2018 04:21
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 27 05:50:22 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 26 15:17:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.518	0.451	12.9	100	0.00
3	n-Nitrosodimethylamine	0.238	0.225	5.5	106	0.00
4 S	2-Fluorophenol	0.991	0.983	0.8	111	0.00
5 S	Phenol-d6	1.251	1.238	1.0	111	0.00
6	bis(2-Chloroethyl)ether	1.116	1.075	3.7	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	108	-0.01
8 S	Nitrobenzene-d5	0.175	0.211	-20.6	142	0.00
9	Naphthalene	0.979	0.907	7.4	103	-0.01
10	Hexachlorobutadiene	0.179	0.166	7.3	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.651	0.603	7.4	103	0.00
12	2-Methylnaphthalene	0.684	0.631	7.7	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	-0.01
14 S	2,4,6-Tribromophenol	0.159	0.150	5.7	123	0.00
15 S	2-Fluorobiphenyl	1.721	1.511	12.2	103	-0.01
16	Acenaphthylene	1.783	1.553	12.9	102	-0.01
17	Acenaphthene	1.236	1.110	10.2	102	0.00
18	Fluorene	1.546	1.368	11.5	101	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
20	4-Bromophenyl-phenylether	0.232	0.204	12.1	104	-0.01
21	Hexachlorobenzene	0.244	0.215	11.9	102	0.00
22	Pentachlorophenol	0.066	0.075	-13.6	137	0.00
23	Phenanthrene	1.097	0.968	11.8	104	0.00
24	Anthracene	0.941	0.846	10.1	107	-0.01
25 SURR	Fluoranthene-d10	1.163	1.125	3.3	115	0.00
26	Fluoranthene	1.260	1.196	5.1	112	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	123	0.00
28	Pvrene	1.106	0.985	10.9	117	0.00
29 S	Terphenyl-d14	0.930	0.788	15.3	114	0.00
30	Benzo(a)anthracene	1.048	1.027	2.0	134	-0.01
31	Chrysene	1.131	1.003	11.3	116	-0.01
32	Bis(2-ethylhexyl)phthalate	0.310	0.449	-44.8#	203#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.170	1.091	6.8	122	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	133	0.00
35	Benzo(b)fluoranthene	1.232	1.045	15.2	120	0.00
36	Benzo(k)fluoranthene	1.236	1.033	16.4	119	-0.01
37 C	Benzo(a)pyrene	1.136	0.981	13.6	121	0.00
38	Dibenzo(a,h)anthracene	1.058	0.897	15.2	124	-0.01
39	Benzo(a,h,i)perylene	1.058	0.888	16.1	123	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102618\
Data File : BN003327.D
Acq On : 27 Oct 2018 04:21
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 27 05:50:22 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 26 15:17:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			