

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102618\
 Data File : BN003327.D
 Acq On : 27 Oct 2018 04:21
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 27 05:50:22 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 26 15:17:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	107	0.00
2	1,4-Dioxane	0.400	0.348	13.0	100	0.00
3	n-Nitrosodimethylamine	0.400	0.378	5.5	106	0.00
4 S	2-Fluorophenol	0.400	0.397	0.8	111	0.00
5 S	Phenol-d6	0.400	0.396	1.0	111	0.00
6	bis(2-Chloroethyl)ether	0.400	0.385	3.8	104	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	108	-0.01
8 S	Nitrobenzene-d5	0.400	0.481	-20.2	142	0.00
9	Naphthalene	0.400	0.371	7.3	103	-0.01
10	Hexachlorobutadiene	0.400	0.371	7.3	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.370	7.5	103	0.00
12	2-Methylnaphthalene	0.400	0.369	7.8	103	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	110	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.378	5.5	123	0.00
15 S	2-Fluorobiphenyl	0.400	0.351	12.3	103	-0.01
16	Acenaphthylene	0.400	0.348	13.0	102	-0.01
17	Acenaphthene	0.400	0.359	10.3	102	0.00
18	Fluorene	0.400	0.354	11.5	101	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	112	0.00
20	4-Bromophenyl-phenylether	0.400	0.353	11.8	104	-0.01
21	Hexachlorobenzene	0.400	0.352	12.0	102	0.00
22	Pentachlorophenol	0.400	0.452	-13.0	137	0.00
23	Phenanthrene	0.400	0.353	11.8	104	0.00
24	Anthracene	0.400	0.359	10.3	107	-0.01
25 SURR	Fluoranthene-d10	0.400	0.387	3.3	115	0.00
26	Fluoranthene	0.400	0.380	5.0	112	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	123	0.00
28	Pvrene	0.400	0.356	11.0	117	0.00
29 S	Terphenyl-d14	0.400	0.339	15.3	114	0.00
30	Benzo(a)anthracene	0.400	0.392	2.0	134	-0.01
31	Chrysene	0.400	0.355	11.3	116	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.579	-44.7#	203	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.373	6.8	122	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	133	0.00
35	Benzo(b)fluoranthene	0.400	0.339	15.3	120	0.00
36	Benzo(k)fluoranthene	0.400	0.334	16.5	119	-0.01
37 C	Benzo(a)pyrene	0.400	0.346	13.5	121	0.00
38	Dibenzo(a,h)anthracene	0.400	0.339	15.3	124	-0.01
39	Benzo(a,h,i)perylene	0.400	0.336	16.0	123	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102618\
Data File : BN003327.D
Acq On : 27 Oct 2018 04:21
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 27 05:50:22 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 26 15:17:57 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				