

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN102918\
 Data File : BN003334.D
 Acq On : 29 Oct 2018 13:37
 Operator : MJ/SJ
 Sample : J5521-07
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 AOC8-02B

Quant Time: Oct 30 13:08:57 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 29 16:15:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.86	152	16895	0.40	ng	0.00
7) Naphthalene-d8	10.66	136	77179	0.40	ng	-0.01
13) Acenaphthene-d10	14.49	164	47541	0.40	ng	-0.01
19) Phenanthrene-d10	17.24	188	108897	0.40	ng	0.00
27) Chrysene-d12	21.43	240	107146	0.40	ng	0.01
34) Perylene-d12	23.76	264	96526	0.40	ng	0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.42	112	16900	0.40	ng	0.00
5) Phenol-d6	7.01	99	21377	0.40	ng	0.00
8) Nitrobenzene-d5	9.02	82	16866	0.50	ng	0.00
11) 2-Methylnaphthalene-d10	12.24	152	36893	0.29	ng	0.00
14) 2,4,6-Tribromophenol	15.98	330	7533	0.40	ng	0.00
15) 2-Fluorobiphenyl	13.12	172	49728	0.24	ng	-0.01
25) Fluoranthene-d10	19.26	212	68625	0.22	ng	0.00
29) Terphenyl-d14	19.87	244	52781	0.21	ng	0.00
Target Compounds						
9) Naphthalene	10.71	128	4835	0.026	ng	# 78
18) Fluorene	15.54	166	12140	0.066	ng	# 64
23) Phenanthrene	17.28	178	24951	0.084	ng	# 80
26) Fluoranthene	19.29	202	20026	0.058	ng	# 65
28) Pyrene	19.66	202	29146	0.098	ng	# 77
32) Bis(2-ethylhexyl)phthalate	21.33	149	145919	1.756	ng	# 62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN102918\
Data File : BN003334.D
Acq On : 29 Oct 2018 13:37
Operator : MJ/SJ
Sample : J5521-07
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
AOC8-02B

Quant Time: Oct 30 13:08:57 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 29 16:15:49 2018
Response via : Initial Calibration

