

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102918\
 Data File : BN003336.D
 Acq On : 29 Oct 2018 17:15
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 30 06:45:26 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 29 16:15:49 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.518	0.446	13.9	102	0.00
3	n-Nitrosodimethylamine	0.238	0.228	4.2	111	-0.02
4 S	2-Fluorophenol	0.991	0.944	4.7	110	0.00
5 S	Phenol-d6	1.251	1.219	2.6	113	0.00
6	bis(2-Chloroethyl)ether	1.116	1.096	1.8	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	115	-0.01
8 S	Nitrobenzene-d5	0.175	0.209	-19.4	151#	0.00
9	Naphthalene	0.979	0.904	7.7	110	-0.01
10	Hexachlorobutadiene	0.179	0.167	6.7	111	-0.01
11 SURR	2-Methylnaphthalene-d10	0.651	0.598	8.1	110	0.00
12	2-Methylnaphthalene	0.684	0.630	7.9	110	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	114	-0.01
14 S	2,4,6-Tribromophenol	0.159	0.153	3.8	129	0.00
15 S	2-Fluorobiphenyl	1.721	1.531	11.0	108	-0.01
16	Acenaphthylene	1.783	1.584	11.2	108	-0.01
17	Acenaphthene	1.236	1.119	9.5	107	0.00
18	Fluorene	1.546	1.344	13.1	103	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4-Bromophenyl-phenylether	0.232	0.207	10.8	104	-0.01
21	Hexachlorobenzene	0.244	0.223	8.6	104	-0.01
22	Pentachlorophenol	0.066	0.078	-18.2	141	0.00
23	Phenanthrene	1.097	0.974	11.2	103	0.00
24	Anthracene	0.941	0.815	13.4	102	-0.01
25 SURR	Fluoranthene-d10	1.163	1.009	13.2	101	0.00
26	Fluoranthene	1.260	1.091	13.4	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
28	Pvrene	1.106	0.996	9.9	102	0.00
29 S	Terphenyl-d14	0.930	0.799	14.1	100	0.00
30	Benzo(a)anthracene	1.048	0.953	9.1	108	-0.01
31	Chrysene	1.131	0.997	11.8	100	-0.01
32	Bis(2-ethylhexyl)phthalate	0.310	0.324	-4.5	127	-0.01
33	Indeno(1,2,3-cd)pyrene	1.170	0.936	20.0	90	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	104	0.00
35	Benzo(b)fluoranthene	1.248	1.111	11.0	100	-0.01
36	Benzo(k)fluoranthene	1.236	1.091	11.7	99	-0.01
37 C	Benzo(a)pyrene	1.136	1.008	11.3	98	-0.01
38	Dibenzo(a,h)anthracene	1.058	0.857	19.0	93	-0.01
39	Benzo(a,h,i)perylene	1.058	0.832	21.4	90	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102918\
Data File : BN003336.D
Acq On : 29 Oct 2018 17:15
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 30 06:45:26 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 29 16:15:49 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			