

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102918\
 Data File : BN003354.D
 Acq On : 30 Oct 2018 04:19
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 30 06:47:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 29 16:15:49 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.518	0.459	11.4	102	0.00
3	n-Nitrosodimethylamine	0.238	0.237	0.4	111	-0.02
4 S	2-Fluorophenol	0.991	0.907	8.5	102	0.00
5 S	Phenol-d6	1.251	1.141	8.8	102	0.00
6	bis(2-Chloroethyl)ether	1.116	1.086	2.7	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	109	-0.01
8 S	Nitrobenzene-d5	0.175	0.207	-18.3	142	0.00
9	Naphthalene	0.979	0.907	7.4	105	-0.01
10	Hexachlorobutadiene	0.179	0.167	6.7	105	-0.01
11 SURR	2-Methylnaphthalene-d10	0.651	0.604	7.2	105	0.00
12	2-Methylnaphthalene	0.684	0.634	7.3	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	113	-0.01
14 S	2,4,6-Tribromophenol	0.159	0.150	5.7	125	0.00
15 S	2-Fluorobiphenyl	1.721	1.520	11.7	106	-0.01
16	Acenaphthylene	1.783	1.517	14.9	102	-0.01
17	Acenaphthene	1.236	1.104	10.7	105	0.00
18	Fluorene	1.546	1.338	13.5	101	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
20	4-Bromophenyl-phenylether	0.232	0.208	10.3	105	-0.01
21	Hexachlorobenzene	0.244	0.227	7.0	107	-0.01
22	Pentachlorophenol	0.066	0.075	-13.6	136	0.00
23	Phenanthrene	1.097	0.980	10.7	104	-0.01
24	Anthracene	0.941	0.792	15.8	99	-0.01
25 SURR	Fluoranthene-d10	1.163	0.984	15.4	99	0.00
26	Fluoranthene	1.260	1.078	14.4	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	-0.01
28	Pvrene	1.106	1.010	8.7	100	0.00
29 S	Terphenyl-d14	0.930	0.829	10.9	99	0.00
30	Benzo(a)anthracene	1.048	0.877	16.3	95	-0.01
31	Chrysene	1.131	1.022	9.6	98	-0.01
32	Bis(2-ethylhexyl)phthalate	0.310	0.279	10.0	105	-0.02
33	Indeno(1,2,3-cd)pyrene	1.170	0.906	22.6	84	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	94	-0.01
35	Benzo(b)fluoranthene	1.248	1.196	4.2	97	-0.01
36	Benzo(k)fluoranthene	1.236	1.135	8.2	93	-0.01
37 C	Benzo(a)pyrene	1.136	1.056	7.0	92	-0.01
38	Dibenzo(a,h)anthracene	1.058	0.883	16.5	86	-0.02
39	Benzo(a,h,i)perylene	1.058	0.853	19.4	84	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102918\
Data File : BN003354.D
Acq On : 30 Oct 2018 04:19
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 30 06:47:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 29 16:15:49 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			