

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102918\
 Data File : BN003354.D
 Acq On : 30 Oct 2018 04:19
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 30 06:47:38 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 29 16:15:49 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	107	0.00
2	1,4-Dioxane	0.400	0.354	11.5	102	0.00
3	n-Nitrosodimethylamine	0.400	0.397	0.8	111	-0.02
4 S	2-Fluorophenol	0.400	0.366	8.5	102	0.00
5 S	Phenol-d6	0.400	0.365	8.8	102	0.00
6	bis(2-Chloroethyl)ether	0.400	0.389	2.8	105	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	109	-0.01
8 S	Nitrobenzene-d5	0.400	0.473	-18.2	142	0.00
9	Naphthalene	0.400	0.371	7.3	105	-0.01
10	Hexachlorobutadiene	0.400	0.374	6.5	105	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.371	7.3	105	0.00
12	2-Methylnaphthalene	0.400	0.371	7.3	105	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	113	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.376	6.0	125	0.00
15 S	2-Fluorobiphenyl	0.400	0.353	11.8	106	-0.01
16	Acenaphthylene	0.400	0.340	15.0	102	-0.01
17	Acenaphthene	0.400	0.357	10.8	105	0.00
18	Fluorene	0.400	0.346	13.5	101	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	111	0.00
20	4-Bromophenyl-phenylether	0.400	0.359	10.3	105	-0.01
21	Hexachlorobenzene	0.400	0.372	7.0	107	-0.01
22	Pentachlorophenol	0.400	0.454	-13.5	136	0.00
23	Phenanthrene	0.400	0.357	10.8	104	-0.01
24	Anthracene	0.400	0.337	15.8	99	-0.01
25 SURR	Fluoranthene-d10	0.400	0.338	15.5	99	0.00
26	Fluoranthene	0.400	0.342	14.5	100	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	102	-0.01
28	Pyrene	0.400	0.365	8.8	100	0.00
29 S	Terphenyl-d14	0.400	0.356	11.0	99	0.00
30	Benzo(a)anthracene	0.400	0.335	16.3	95	-0.01
31	Chrysene	0.400	0.362	9.5	98	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.360	10.0	105	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.309	22.8	84	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	94	-0.01
35	Benzo(b)fluoranthene	0.400	0.383	4.3	97	-0.01
36	Benzo(k)fluoranthene	0.400	0.367	8.3	93	-0.01
37 C	Benzo(a)pyrene	0.400	0.372	7.0	92	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.334	16.5	86	-0.02
39	Benzo(a,h,i)perylene	0.400	0.322	19.5	84	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN102918\
Data File : BN003354.D
Acq On : 30 Oct 2018 04:19
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 30 06:47:38 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 29 16:15:49 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				