




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102921\
Data File : BN017197.D
Acq On : 30 Oct 2021 02:34
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 30 04:27:34 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 30 04:27:29 2021
Response via : Initial Calibration



TIC: BN017197.D\data.ms

(70) Pentachlorophenol (C)




16.586min (0.000) 25694.95 ng

response 520049

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	59.38
264.00	61.60	62.87
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102921\
Data File : BN017197.D
Acq On : 30 Oct 2021 02:34
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Oct 30 04:27:34 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 30 04:27:29 2021
Response via : Initial Calibration



TIC: BN017197.D\data.ms

(77) Benzidine

19.210min (0.000) 0.00 ng

response 3142713

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.03
183.00	13.20	11.63
0.00	0.00	0.00