

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102921\
 Data File : BN017198.D
 Acq On : 30 Oct 2021 03:47
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 30 04:22:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 30 00:49:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	111	0.00
2	1,4-Dioxane	0.213	0.211	0.9	121	0.00
3	n-Nitrosodimethylamine	0.351	0.318	9.4	114	0.00
4 S	2-Fluorophenol	0.897	0.873	2.7	124	0.00
5 S	Phenol-d6	0.996	0.963	3.3	124	0.00
6	bis(2-Chloroethyl)ether	1.053	1.002	4.8	114	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
8 S	Nitrobenzene-d5	0.233	0.208	10.7	124	0.00
9	Naphthalene	1.044	0.955	8.5	111	0.00
10	Hexachlorobutadiene	0.192	0.178	7.3	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.606	0.860	-41.9#	113	0.00
12	2-Methylnaphthalene	0.656	0.621	5.3	112	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
14 S	2,4,6-Tribromophenol	0.144	0.122	15.3	128	0.00
15 S	2-Fluorobiphenyl	1.495	1.405	6.0	112	0.00
16	Acenaphthylene	1.601	1.479	7.6	119	0.00
17	Acenaphthene	1.104	1.027	7.0	113	0.00
18	Fluorene	1.309	1.229	6.1	115	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4,6-Dinitro-2-methylphenol	0.044	0.005	88.6#	23#	0.00
21	4-Bromophenyl-phenylether	0.220	0.203	7.7	116	0.00
22	Hexachlorobenzene	0.240	0.222	7.5	113	0.00
23	Atrazine	0.155	0.142	8.4	132	0.00
24	Pentachlorophenol	0.088	0.057	35.2#	116	0.00
25	Phenanthrene	1.120	1.050	6.3	113	0.00
26	Anthracene	0.938	0.878	6.4	121	0.00
27 SURR	Fluoranthene-d10	1.124	1.639	-45.8#	121	0.00
28	Fluoranthene	1.214	1.202	1.0	117	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	121	-0.01
30	Pyrene	1.259	1.186	5.8	122	0.00
31 S	Terphenyl-d14	0.859	0.809	5.8	121	0.00
32	Benzo(a)anthracene	1.209	1.133	6.3	129	0.00
33	Chrysene	1.271	1.203	5.4	122	0.00
34	Bis(2-ethylhexyl)phthalate	0.510	0.517	-1.4	181#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	125	-0.01
36	Indeno(1,2,3-cd)pyrene	1.453	1.357	6.6	124	0.00
37	Benzo(b)fluoranthene	1.248	1.210	3.0	128	0.00
38	Benzo(k)fluoranthene	1.241	1.161	6.4	122	0.00
39 C	Benzo(a)pyrene	1.129	1.084	4.0	131	0.00
40	Dibenzo(a,h)anthracene	1.179	1.104	6.4	125	-0.01
41	Benzo(g,h,i)perylene	1.272	1.157	9.0	121	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4