

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102921\
 Data File : BN017207.D
 Acq On : 30 Oct 2021 09:15
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 31 23:43:09 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 30 00:49:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	109	0.00
2	1,4-Dioxane	0.400	0.332	17.0	100	0.00
3	n-Nitrosodimethylamine	0.400	0.355	11.3	110	0.00
4 S	2-Fluorophenol	0.400	0.371	7.3	116	0.00
5 S	Phenol-d6	0.400	0.364	9.0	115	0.00
6	bis(2-Chloroethyl)ether	0.400	0.379	5.3	111	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	108	0.00
8 S	Nitrobenzene-d5	0.400	0.366	8.5	122	0.00
9	Naphthalene	0.400	0.368	8.0	107	0.00
10	Hexachlorobutadiene	0.400	0.385	3.8	112	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.567	-41.7#	109	0.00
12	2-Methylnaphthalene	0.400	0.378	5.5	108	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.400	0.398	0.5	117	0.00
15 S	2-Fluorobiphenyl	0.400	0.381	4.8	107	0.00
16	Acenaphthylene	0.400	0.371	7.3	113	0.00
17	Acenaphthene	0.400	0.374	6.5	107	0.00
18	Fluorene	0.400	0.376	6.0	109	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	109	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.156	61.0#	29	0.00
21	4-Bromophenyl-phenylether	0.400	0.369	7.8	109	0.00
22	Hexachlorobenzene	0.400	0.376	6.0	108	0.00
23	Atrazine	0.400	0.391	2.3	126	0.00
24	Pentachlorophenol	0.400	0.397	0.8	99	0.00
25	Phenanthrene	0.400	0.374	6.5	106	0.00
26	Anthracene	0.400	0.374	6.5	114	0.00
27 SURR	Fluoranthene-d10	0.400	0.576	-44.0#	112	0.00
28	Fluoranthene	0.400	0.391	2.3	109	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	111	-0.01
30	Pyrene	0.400	0.378	5.5	113	0.00
31 S	Terphenyl-d14	0.400	0.380	5.0	112	0.00
32	Benzo(a)anthracene	0.400	0.370	7.5	117	0.00
33	Chrysene	0.400	0.387	3.3	114	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.440	-10.0	166	0.00
35 I	Perylene-d12	0.400	0.400	0.0	114	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.366	8.5	111	0.00
37	Benzo(b)fluoranthene	0.400	0.385	3.8	116	0.00
38	Benzo(k)fluoranthene	0.400	0.381	4.8	114	0.00
39 C	Benzo(a)pyrene	0.400	0.386	3.5	120	0.00
40	Dibenzo(a,h)anthracene	0.400	0.367	8.3	112	0.00
41	Benzo(g,h,i)perylene	0.400	0.355	11.3	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4