

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102921\
 Data File : BN017207.D
 Acq On : 30 Oct 2021 09:15
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Oct 31 23:43:09 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 30 00:49:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.561	152	36358	0.400	ng	0.00
7) Naphthalene-d8	10.332	136	161338	0.400	ng	0.00
13) Acenaphthene-d10	14.194	164	82947	0.400	ng	0.00
19) Phenanthrene-d10	16.946	188	153481	0.400	ng	0.00
29) Chrysene-d12	21.154	240	154270	0.400	ng	#-0.01
35) Perylene-d12	23.369	264	156271	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.173	112	30264	0.371	ng	0.00
5) Phenol-d6	6.749	99	32946	0.364	ng	0.00
8) Nitrobenzene-d5	8.710	82	34410	0.366	ng	0.00
11) 2-Methylnaphthalene-d10	11.929	152	138783	0.567	ng	0.00
14) 2,4,6-Tribromophenol	15.704	330	9817	0.398	ng	0.00
15) 2-Fluorobiphenyl	12.824	172	118052	0.381	ng	0.00
27) Fluoranthene-d10	18.995	212	248497	0.576	ng	0.00
31) Terphenyl-d14	19.605	244	125792	0.380	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.153	88	6439	0.332	ng	# 51
3) n-Nitrosodimethylamine	3.485	42	11307	0.355	ng	99
6) bis(2-Chloroethyl)ether	7.008	93	36253	0.379	ng	100
9) Naphthalene	10.383	128	154877	0.368	ng	99
10) Hexachlorobutadiene	10.661	225	29811	0.385	ng	# 100
12) 2-Methylnaphthalene	12.000	142	99898	0.378	ng	99
16) Acenaphthylene	13.915	152	123062	0.371	ng	100
17) Acenaphthene	14.257	154	85673	0.374	ng	99
18) Fluorene	15.247	166	102071	0.376	ng	100
20) 4,6-Dinitro-2-methylph...	15.349	198	1115	0.156	ng	# 89
21) 4-Bromophenyl-phenylether	16.155	248	31187	0.369	ng	96
22) Hexachlorobenzene	16.264	284	34643	0.376	ng	100
23) Atrazine	16.435	200	22026	0.391	ng	97
24) Pentachlorophenol	16.605	266	7947	0.397	ng	99
25) Phenanthrene	16.995	178	160597	0.374	ng	100
26) Anthracene	17.080	178	134607	0.374	ng	100
28) Fluoranthene	19.025	202	182410	0.391	ng	100
30) Pyrene	19.387	202	183802	0.378	ng	100
32) Benzo(a)anthracene	21.144	228	172718	0.370	ng	99
33) Chrysene	21.197	228	189690	0.387	ng	99
34) Bis(2-ethylhexyl)phtha...	21.101	149	79996	0.440	ng	99
36) Indeno(1,2,3-cd)pyrene	25.591	276	207683	0.366	ng	98
37) Benzo(b)fluoranthene	22.705	252	187735	0.385	ng	99
38) Benzo(k)fluoranthene	22.750	252	184534	0.381	ng	# 96
39) Benzo(a)pyrene	23.270	252	170301	0.386	ng	100
40) Dibenzo(a,h)anthracene	25.615	278	168981	0.367	ng	99
41) Benzo(g,h,i)perylene	26.272	276	176409	0.355	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN102921\
Data File : BN017207.D
Acq On : 30 Oct 2021 09:15
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Oct 31 23:43:09 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN102921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 30 00:49:11 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.561 min Scan# 497
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 36358
Ion Ratio Lower Upper
152 100
150 155.9 124.6 187.0
115 59.2 47.0 70.6



#2
1,4-Dioxane
Concen: 0.332 ng
RT: 3.153 min Scan# 19
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion: 88 Resp: 6439
Ion Ratio Lower Upper
88 100
43 36.7 103.8 155.6#
58 85.6 70.6 106.0









#3
n-Nitrosodimethylamine
Concen: 0.355 ng
RT: 3.485 min Scan# 55
Delta R.T. -0.009 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 11307
Ion Ratio Lower Upper
42 100
74 126.9 100.5 150.7
44 5.7 5.2 7.8



#4
2-Fluorophenol
Concen: 0.371 ng
RT: 5.173 min Scan# 238
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion: 112 Resp: 30264
Ion Ratio Lower Upper
112 100
64 58.5 46.6 70.0
63 29.9 23.8 35.6









#5
Phenol-d6
Concen: 0.364 ng
RT: 6.749 min Scan# 409
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion: 99 Resp: 32946




Ion	Ratio	Lower	Upper
99	100		
42	20.5	16.2	24.4
71	31.2	24.6	37.0




#6
bis(2-Chloroethyl)ether
Concen: 0.379 ng
RT: 7.008 min Scan# 437
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion: 93 Resp: 36253

Ion	Ratio	Lower	Upper
93	100		
63	80.2	64.5	96.7
95	33.1	26.6	39.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.332 min Scan# 738
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	161338
Ion	Ratio	Lower	Upper
136	100		
137	10.9	8.8	13.2
54	5.7	4.8	7.2
68	4.4	3.6	5.4



#8
Nitrobenzene-d5
Concen: 0.366 ng
RT: 8.710 min Scan# 610
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:	82	Resp:	34410
Ion	Ratio	Lower	Upper
82	100		
128	46.5	36.0	54.0
54	51.5	41.0	61.6









#9
Naphthalene
Concen: 0.368 ng
RT: 10.383 min Scan# 742
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128 Resp: 154877
Ion Ratio Lower Upper
128 100
129 11.0 8.7 13.1
127 12.5 10.2 15.2



#10
Hexachlorobutadiene
Concen: 0.385 ng
RT: 10.661 min Scan# 764
Delta R.T. -0.013 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:225 Resp: 29811
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.5 51.0 76.4









#11
2-Methylnaphthalene-d10
Concen: 0.567 ng
RT: 11.929 min Scan# 925
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 138783
Ion Ratio Lower Upper
152 100
151 19.9 15.9 23.9



#12
2-Methylnaphthalene
Concen: 0.378 ng
RT: 12.000 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:142 Resp: 99898
Ion Ratio Lower Upper
142 100
141 88.2 71.0 106.6
115 32.9 26.8 40.2









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.194 min Scan# 1188
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 82947
Ion Ratio Lower Upper
164 100
162 103.9 83.6 125.4
160 47.6 38.3 57.5



#14
2,4,6-Tribromophenol
Concen: 0.398 ng
RT: 15.704 min Scan# 1307
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:330 Resp: 9817
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 37.6 30.7 46.1









#15
2-Fluorobiphenyl
Concen: 0.381 ng
RT: 12.824 min Scan# 1080
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172 Resp: 118052
Ion Ratio Lower Upper
172 100
171 33.4 27.0 40.6
170 21.4 17.5 26.3



#16
Acenaphthylene
Concen: 0.371 ng
RT: 13.915 min Scan# 1166
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:152 Resp: 123062
Ion Ratio Lower Upper
152 100
151 19.4 15.6 23.4
153 13.1 10.4 15.6









#17
Acenaphthene
Concen: 0.374 ng
RT: 14.257 min Scan# 1193
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 85673
Ion Ratio Lower Upper
154 100
153 112.4 91.1 136.7
152 54.7 44.4 66.6



#18
Fluorene
Concen: 0.376 ng
RT: 15.247 min Scan# 1271
Delta R.T. -0.013 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:166 Resp: 102071
Ion Ratio Lower Upper
166 100
165 97.8 78.4 117.6
167 13.6 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.946 min Scan# 1408
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 153481
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.4 10.3 15.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.156 ng
RT: 15.349 min Scan# 1279
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:198 Resp: 1115
Ion Ratio Lower Upper
198 100
182 18.7 11.4 17.2#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.369 ng
RT: 16.155 min Scan# 1343
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:248 Resp: 31187




Ion	Ratio	Lower	Upper
248	100		
250	97.4	77.4	116.0
141	48.5	44.2	66.4




#22
Hexachlorobenzene
Concen: 0.376 ng
RT: 16.264 min Scan# 1352
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:284 Resp: 34643

Ion	Ratio	Lower	Upper
284	100		
142	36.1	29.0	43.4
249	30.8	24.5	36.7









#23
Atrazine
Concen: 0.391 ng
RT: 16.435 min Scan# 1366
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	200	Resp:	22026
Ion Ratio	Lower	Upper	
200	100		
173	25.1	21.7	32.5
215	48.7	37.8	56.6



#24
Pentachlorophenol
Concen: 0.397 ng
RT: 16.605 min Scan# 1380
Delta R.T. 0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:	266	Resp:	7947
Ion Ratio	Lower	Upper	
266	100		
264	63.2	50.3	75.5
268	63.9	50.7	76.1









#25
Phenanthrene
Concen: 0.374 ng
RT: 16.995 min Scan# 1412
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 160597
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.4
179 15.6 12.3 18.5



#26
Anthracene
Concen: 0.374 ng
RT: 17.080 min Scan# 1419
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:178 Resp: 134607
Ion Ratio Lower Upper
178 100
176 18.2 14.8 22.2
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.576 ng
RT: 18.995 min Scan# 1693
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 248497
Ion Ratio Lower Upper
212 100
106 13.2 10.6 15.8
104 7.2 5.8 8.6



#28
Fluoranthene
Concen: 0.391 ng
RT: 19.025 min Scan# 1699
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:202 Resp: 182410
Ion Ratio Lower Upper
202 100
101 11.1 9.0 13.4
203 17.4 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.154 min Scan# 2005
Delta R.T. -0.011 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 154270
Ion Ratio Lower Upper
240 100
120 12.9 5.0 7.4#
236 26.7 20.3 30.5



#30
Pyrene
Concen: 0.378 ng
RT: 19.387 min Scan# 1772
Delta R.T. -0.005 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:202 Resp: 183802
Ion Ratio Lower Upper
202 100
200 20.5 16.5 24.7
203 17.5 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.380 ng
 RT: 19.605 min Scan# 1816
 Delta R.T. -0.005 min
 Lab File: BN017207.D
 Acq: 30 Oct 2021 09:15




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244 Resp: 125792
 Ion Ratio Lower Upper
 244 100
 212 7.3 5.8 8.6
 122 12.5 9.2 13.8



#32
 Benzo(a)anthracene
 Concen: 0.370 ng
 RT: 21.144 min Scan# 2004
 Delta R.T. -0.000 min
 Lab File: BN017207.D
 Acq: 30 Oct 2021 09:15

Tgt Ion:228 Resp: 172718
 Ion Ratio Lower Upper
 228 100
 226 26.6 21.9 32.9
 229 19.7 15.4 23.2









#33
Chrysene
Concen: 0.387 ng
RT: 21.197 min Scan# 2009
Delta R.T. -0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:228 Resp: 189690




Ion	Ratio	Lower	Upper
228	100		
226	29.0	23.6	35.4
229	19.7	15.5	23.3




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.440 ng
RT: 21.101 min Scan# 2000
Delta R.T. 0.000 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:149 Resp: 79996


Ion	Ratio	Lower	Upper
149	100		
167	27.5	21.6	32.4
279	4.1	3.2	4.8








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.369 min Scan# 2404
Delta R.T. -0.007 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:264 Resp: 156271
Ion Ratio Lower Upper
264 100
260 24.4 19.4 29.0
265 23.1 18.2 27.4



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.366 ng
RT: 25.591 min Scan# 3053
Delta R.T. -0.007 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:276 Resp: 207683
Ion Ratio Lower Upper
276 100
138 27.0 20.4 30.6
277 25.1 20.0 30.0





#37

Benzo(b)fluoranthene

Concen: 0.385 ng

RT: 22.705 min Scan# 2210

Delta R.T. -0.007 min

Lab File: BN017207.D

Acq: 30 Oct 2021 09:15





Tgt Ion:252 Resp: 187735

Ion Ratio Lower Upper

252 100

253 22.2 17.5 26.3

125 10.2 7.9 11.9



#38

Benzo(k)fluoranthene

Concen: 0.381 ng

RT: 22.750 min Scan# 2223

Delta R.T. -0.003 min

Lab File: BN017207.D

Acq: 30 Oct 2021 09:15




Tgt Ion:252 Resp: 184534


Ion Ratio Lower Upper

252 100

253 22.7 17.6 26.4


125 14.8 9.2 13.8#








#39
Benzo(a)pyrene
Concen: 0.386 ng
RT: 23.270 min Scan# 2375
Delta R.T. -0.003 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:252 Resp: 170301
Ion Ratio Lower Upper
252 100
253 22.1 17.6 26.4
125 11.5 9.1 13.7



#40
Dibenzo(a,h)anthracene
Concen: 0.367 ng
RT: 25.615 min Scan# 3060
Delta R.T. -0.007 min
Lab File: BN017207.D
Acq: 30 Oct 2021 09:15

Tgt Ion:278 Resp: 168981
Ion Ratio Lower Upper
278 100
139 17.9 13.2 19.8
279 23.7 19.0 28.6





#41
 Benzo(g,h,i)perylene
 Concen: 0.355 ng
 RT: 26.272 min Scan# 3252
 Delta R.T. -0.003 min
 Lab File: BN017207.D
 Acq: 30 Oct 2021 09:15

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:	276	Resp:	176409
Ion	Ratio	Lower	Upper
276	100		
277	23.8	19.0	28.4
138	21.4	17.4	26.0

