

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN103018\
 Data File : BN003356.D
 Acq On : 30 Oct 2018 12:03
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 30 18:14:12 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 29 16:15:49 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	131	0.00
2	1,4-Dioxane	0.400	0.367	8.3	128	0.00
3	n-Nitrosodimethylamine	0.400	0.422	-5.5	144	-0.02
4 S	2-Fluorophenol	0.400	0.375	6.3	127	0.00
5 S	Phenol-d6	0.400	0.373	6.8	127	0.00
6	bis(2-Chloroethyl)ether	0.400	0.410	-2.5	135	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	132	-0.01
8 S	Nitrobenzene-d5	0.400	0.493	-23.2	178	0.00
9	Naphthalene	0.400	0.388	3.0	131	-0.01
10	Hexachlorobutadiene	0.400	0.389	2.8	132	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	131	-0.01
12	2-Methylnaphthalene	0.400	0.386	3.5	131	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	131	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.383	4.3	148	-0.01
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	132	-0.02
16	Acenaphthylene	0.400	0.365	8.8	127	-0.01
17	Acenaphthene	0.400	0.387	3.3	131	0.00
18	Fluorene	0.400	0.377	5.8	128	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	126	-0.01
20	4-Bromophenyl-phenylether	0.400	0.389	2.8	129	-0.02
21	Hexachlorobenzene	0.400	0.408	-2.0	133	-0.01
22	Pentachlorophenol	0.400	0.460	-15.0	158	0.00
23	Phenanthrene	0.400	0.390	2.5	129	-0.01
24	Anthracene	0.400	0.363	9.3	122	-0.01
25 SURR	Fluoranthene-d10	0.400	0.361	9.8	121	0.00
26	Fluoranthene	0.400	0.368	8.0	123	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	119	0.00
28	Pvrene	0.400	0.389	2.8	123	0.00
29 S	Terphenyl-d14	0.400	0.382	4.5	124	0.00
30	Benzo(a)anthracene	0.400	0.359	10.3	119	-0.01
31	Chrysene	0.400	0.366	8.5	116	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.362	9.5	123	-0.02
33	Indeno(1,2,3-cd)pyrene	0.400	0.323	19.3	102	-0.01
34 I	Perylene-d12	0.400	0.400	0.0	106	0.00
35	Benzo(b)fluoranthene	0.400	0.396	1.0	113	-0.01
36	Benzo(k)fluoranthene	0.400	0.384	4.0	109	-0.01
37 C	Benzo(a)pyrene	0.400	0.387	3.3	108	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.359	10.3	104	-0.02
39	Benzo(a,h,i)perylene	0.400	0.347	13.3	102	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN103018\
Data File : BN003356.D
Acq On : 30 Oct 2018 12:03
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 30 18:14:12 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 29 16:15:49 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				