

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN103018\
 Data File : BN003356.D
 Acq On : 30 Oct 2018 12:03
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 30 18:14:12 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 29 16:15:49 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	131	0.00
2	1,4-Dioxane	0.518	0.475	8.3	128	0.00
3	n-Nitrosodimethylamine	0.238	0.252	-5.9	144	-0.02
4 S	2-Fluorophenol	0.991	0.928	6.4	127	0.00
5 S	Phenol-d6	1.251	1.165	6.9	127	0.00
6	bis(2-Chloroethyl)ether	1.116	1.144	-2.5	135	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	132	-0.01
8 S	Nitrobenzene-d5	0.175	0.216	-23.4	178#	0.00
9	Naphthalene	0.979	0.949	3.1	131	-0.01
10	Hexachlorobutadiene	0.179	0.174	2.8	132	-0.01
11 SURR	2-Methylnaphthalene-d10	0.651	0.628	3.5	131	-0.01
12	2-Methylnaphthalene	0.684	0.661	3.4	131	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	131	-0.01
14 S	2,4,6-Tribromophenol	0.159	0.152	4.4	148	-0.01
15 S	2-Fluorobiphenyl	1.721	1.627	5.5	132	-0.02
16	Acenaphthylene	1.783	1.626	8.8	127	-0.01
17	Acenaphthene	1.236	1.195	3.3	131	0.00
18	Fluorene	1.546	1.456	5.8	128	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	126	-0.01
20	4-Bromophenyl-phenylether	0.232	0.225	3.0	129	-0.02
21	Hexachlorobenzene	0.244	0.249	-2.0	133	-0.01
22	Pentachlorophenol	0.066	0.076	-15.2	158#	0.00
23	Phenanthrene	1.097	1.069	2.6	129	-0.01
24	Anthracene	0.941	0.855	9.1	122	-0.01
25 SURR	Fluoranthene-d10	1.163	1.050	9.7	121	0.00
26	Fluoranthene	1.260	1.161	7.9	123	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
28	Pvrene	1.106	1.074	2.9	123	0.00
29 S	Terphenyl-d14	0.930	0.889	4.4	124	0.00
30	Benzo(a)anthracene	1.048	0.942	10.1	119	-0.01
31	Chrysene	1.131	1.036	8.4	116	-0.01
32	Bis(2-ethylhexyl)phthalate	0.310	0.281	9.4	123	-0.02
33	Indeno(1,2,3-cd)pyrene	1.170	0.947	19.1	102	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	106	0.00
35	Benzo(b)fluoranthene	1.248	1.237	0.9	113	-0.01
36	Benzo(k)fluoranthene	1.236	1.188	3.9	109	-0.01
37 C	Benzo(a)pyrene	1.136	1.098	3.3	108	-0.01
38	Dibenzo(a,h)anthracene	1.058	0.949	10.3	104	-0.02
39	Benzo(a,h,i)perylene	1.058	0.918	13.2	102	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN103018\
Data File : BN003356.D
Acq On : 30 Oct 2018 12:03
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 30 18:14:12 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 29 16:15:49 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			