

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN103018\
 Data File : BN003364.D
 Acq On : 30 Oct 2018 18:36
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 31 02:26:30 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 29 16:15:49 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.518	0.426	17.8	101	0.00
3	n-Nitrosodimethylamine	0.238	0.216	9.2	109	-0.02
4 S	2-Fluorophenol	0.991	0.848	14.4	102	0.00
5 S	Phenol-d6	1.251	1.091	12.8	105	0.00
6	bis(2-Chloroethyl)ether	1.116	1.013	9.2	105	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	120	-0.01
8 S	Nitrobenzene-d5	0.175	0.198	-13.1	148	0.00
9	Naphthalene	0.979	0.823	15.9	104	-0.01
10	Hexachlorobutadiene	0.179	0.150	16.2	104	-0.01
11 SURR	2-Methylnaphthalene-d10	0.651	0.551	15.4	105	-0.01
12	2-Methylnaphthalene	0.684	0.578	15.5	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	125	-0.01
14 S	2,4,6-Tribromophenol	0.159	0.138	13.2	128	-0.01
15 S	2-Fluorobiphenyl	1.721	1.352	21.4	104	-0.02
16	Acenaphthylene	1.783	1.384	22.4	103	-0.01
17	Acenaphthene	1.236	1.008	18.4	105	0.00
18	Fluorene	1.546	1.206	22.0	101	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	121	-0.01
20	4-Bromophenyl-phenylether	0.232	0.187	19.4	103	-0.02
21	Hexachlorobenzene	0.244	0.202	17.2	104	-0.01
22	Pentachlorophenol	0.066	0.070	-6.1	140	0.00
23	Phenanthrene	1.097	0.874	20.3	101	-0.01
24	Anthracene	0.941	0.720	23.5	99	-0.01
25 SURR	Fluoranthene-d10	1.163	0.874	24.8	96	0.00
26	Fluoranthene	1.260	0.950	24.6	97	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	102	-0.01
28	Pvrene	1.106	0.986	10.8	97	0.00
29 S	Terphenyl-d14	0.930	0.802	13.8	96	-0.01
30	Benzo(a)anthracene	1.048	0.825	21.3	90	-0.01
31	Chrysene	1.131	0.912	19.4	88	-0.01
32	Bis(2-ethylhexyl)phthalate	0.310	0.303	2.3	114	-0.02
33	Indeno(1,2,3-cd)pyrene	1.170	0.763	34.8#	71	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	87	-0.01
35	Benzo(b)fluoranthene	1.248	1.083	13.2	81	-0.02
36	Benzo(k)fluoranthene	1.236	1.033	16.4	78	-0.02
37 C	Benzo(a)pyrene	1.136	0.944	16.9	76	-0.02
38	Dibenzo(a,h)anthracene	1.058	0.800	24.4	72	-0.03
39	Benzo(a,h,i)perylene	1.058	0.787	25.6#	71	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN103018\
Data File : BN003364.D
Acq On : 30 Oct 2018 18:36
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 31 02:26:30 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 29 16:15:49 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			