

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103118\
 Data File : BN003366.D
 Acq On : 31 Oct 2018 11:08
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 31 14:09:40 2018
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 29 16:15:49 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	-0.02
2	1,4-Dioxane	0.518	0.498	3.9	108	0.00
3	n-Nitrosodimethylamine	0.238	0.234	1.7	107	-0.02
4 S	2-Fluorophenol	0.991	0.915	7.7	100	-0.02
5 S	Phenol-d6	1.251	1.144	8.6	100	-0.02
6	bis(2-Chloroethyl)ether	1.116	1.091	2.2	103	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	105	-0.01
8 S	Nitrobenzene-d5	0.175	0.214	-22.3	140	-0.01
9	Naphthalene	0.979	0.952	2.8	105	-0.01
10	Hexachlorobutadiene	0.179	0.174	2.8	105	-0.01
11 SURR	2-Methylnaphthalene-d10	0.651	0.629	3.4	105	-0.01
12	2-Methylnaphthalene	0.684	0.661	3.4	105	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	104	-0.01
14 S	2,4,6-Tribromophenol	0.159	0.144	9.4	111	-0.01
15 S	2-Fluorobiphenyl	1.721	1.646	4.4	106	-0.02
16	Acenaphthylene	1.783	1.585	11.1	98	-0.01
17	Acenaphthene	1.236	1.195	3.3	104	-0.01
18	Fluorene	1.546	1.440	6.9	100	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	100	-0.01
20	4-Bromophenyl-phenylether	0.232	0.224	3.4	102	-0.02
21	Hexachlorobenzene	0.244	0.247	-1.2	105	-0.01
22	Pentachlorophenol	0.066	0.067	-1.5	110	0.00
23	Phenanthrene	1.097	1.063	3.1	102	-0.01
24	Anthracene	0.941	0.833	11.5	94	-0.01
25 SURR	Fluoranthene-d10	1.163	1.040	10.6	95	0.00
26	Fluoranthene	1.260	1.140	9.5	96	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	88	-0.01
28	Pyrene	1.106	1.118	-1.1	95	0.00
29 S	Terphenyl-d14	0.930	0.926	0.4	96	0.00
30	Benzo(a)anthracene	1.048	0.904	13.7	85	-0.01
31	Chrysene	1.131	1.079	4.6	90	-0.01
32	Bis(2-ethylhexyl)phthalate	0.310	0.275	11.3	89	-0.02
33	Indeno(1,2,3-cd)pyrene	1.170	0.915	21.8	73	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	76	-0.01
35	Benzo(b)fluoranthene	1.248	1.283	-2.8	85	-0.01
36	Benzo(k)fluoranthene	1.236	1.200	2.9	80	-0.02
37 C	Benzo(a)pyrene	1.136	1.107	2.6	79	-0.02
38	Dibenzo(a,h)anthracene	1.058	0.946	10.6	75	-0.02
39	Benzo(g,h,i)perylene	1.058	0.920	13.0	73	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN103118\
Data File : BN003366.D
Acq On : 31 Oct 2018 11:08
Operator : MJ/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 31 14:09:40 2018
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 29 16:15:49 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		