

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN110118\
 Data File : BN003379.D
 Acq On : 01 Nov 2018 15:28
 Operator : MJ/SJ
 Sample : PB114320BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK20

Quant Time: Nov 01 17:26:19 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN110118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 01 13:18:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.86	152	9084	0.40	ng/ul	0.00
2) Naphthalene-d8	10.65	136	36902	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.49	164	22598	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.33	188	2628128	0.40	ng/ul	0.10
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.41
20) Perylene-d12	23.73	264	47723	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.24	152	14978	0.25	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
9) Fluorene	15.49	166	4588	0.044	ng/ul#	12
23) Benzo(a)pyrene	23.58	252	9820	0.057	ng/ul#	45

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN110118\
Data File : BN003379.D
Acq On : 01 Nov 2018 15:28
Operator : MJ/SJ
Sample : PB114320BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK20

Quant Time: Nov 01 17:26:19 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN110118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 01 13:18:49 2018
Response via : Initial Calibration

