

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110122\
 Data File : BN022444.D
 Acq On : 01 Nov 2022 21:42
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 02 05:37:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 02 05:31:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	119	0.00
2	1,4-Dioxane	0.400	0.463	-15.8	135	0.00
3	n-Nitrosodimethylamine	0.400	0.470	-17.5	142	0.00
4 S	2-Fluorophenol	0.400	0.449	-12.2	135	0.00
5 S	Phenol-d6	0.400	0.458	-14.5	141	0.00
6	bis(2-Chloroethyl)ether	0.400	0.463	-15.8	139	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	133	0.00
8 S	Nitrobenzene-d5	0.400	0.352	12.0	124	0.00
9	Naphthalene	0.400	0.391	2.3	134	0.00
10	Hexachlorobutadiene	0.400	0.445	-11.2	150	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.401	-0.3	141	0.00
12	2-Methylnaphthalene	0.400	0.354	11.5	130	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	127	0.01
14 S	2,4,6-Tribromophenol	0.400	0.286	28.5#	102	0.00
15 S	2-Fluorobiphenyl	0.400	0.411	-2.7	134	0.00
16	Acenaphthylene	0.400	0.363	9.3	127	0.00
17	Acenaphthene	0.400	0.381	4.8	127	0.00
18	Fluorene	0.400	0.349	12.8	115	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	98	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.188	53.0#	57	0.00
21	4-Bromophenyl-phenylether	0.400	0.419	-4.7	109	0.00
22	Hexachlorobenzene	0.400	0.445	-11.2	113	0.00
23	Atrazine	0.400	0.331	17.3	88	0.00
24	Pentachlorophenol	0.400	0.231	42.3#	69	0.00
25	Phenanthrene	0.400	0.381	4.8	98	0.00
26	Anthracene	0.400	0.354	11.5	95	0.00
27 SURR	Fluoranthene-d10	0.400	0.316	21.0	81	0.00
28	Fluoranthene	0.400	0.323	19.3	82	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	76	0.00
30	Pyrene	0.400	0.409	-2.2	80	0.00
31 S	Terphenyl-d14	0.400	0.475	-18.7	94	0.00
32	Benzo(a)anthracene	0.400	0.379	5.3	75	0.00
33	Chrysene	0.400	0.396	1.0	78	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.268	33.0#	49	0.00
35 I	Perylene-d12	0.400	0.400	0.0	73	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.426	-6.5	76	0.00
37	Benzo(b)fluoranthene	0.400	0.382	4.5	73	0.00
38	Benzo(k)fluoranthene	0.400	0.381	4.8	74	0.00
39 C	Benzo(a)pyrene	0.400	0.386	3.5	73	0.00
40	Dibenzo(a,h)anthracene	0.400	0.433	-8.2	78	0.00
41	Benzo(g,h,i)perylene	0.400	0.435	-8.7	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4