

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110123\
 Data File : BN028412.D
 Acq On : 01 Nov 2023 22:22
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN110123

Quant Time: Nov 02 06:29:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 02 06:26:50 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	67	0.00
2	1,4-Dioxane	0.400	0.348	13.0	59	0.00
3	n-Nitrosodimethylamine	0.400	0.390	2.5	66	0.00
4 S	2-Fluorophenol	0.400	0.355	11.3	64	0.00
5 S	Phenol-d6	0.400	0.346	13.5	64	0.00
6	bis(2-Chloroethyl)ether	0.400	0.399	0.3	68	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	69	0.00
8 S	Nitrobenzene-d5	0.400	0.354	11.5	66	0.00
9	Naphthalene	0.400	0.407	-1.7	70	0.00
10	Hexachlorobutadiene	0.400	0.393	1.8	68	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.400	0.0	69	0.00
12	2-Methylnaphthalene	0.400	0.406	-1.5	70	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	68	0.00
14 S	2,4,6-Tribromophenol	0.400	0.352	12.0	70	0.00
15 S	2-Fluorobiphenyl	0.400	0.369	7.8	62	0.00
16	Acenaphthylene	0.400	0.362	9.5	64	0.00
17	Acenaphthene	0.400	0.405	-1.3	69	0.00
18	Fluorene	0.400	0.415	-3.7	71	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	68	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.279	30.3#	52	0.00
21	4-Bromophenyl-phenylether	0.400	0.411	-2.7	69	0.00
22	Hexachlorobenzene	0.400	0.404	-1.0	67	0.00
23	Atrazine	0.400	0.351	12.3	63	0.00
24	Pentachlorophenol	0.400	0.345	13.8	67	0.00
25	Phenanthrene	0.400	0.413	-3.2	69	0.00
26	Anthracene	0.400	0.378	5.5	65	0.00
27 SURR	Fluoranthene-d10	0.400	0.399	0.3	67	0.00
28	Fluoranthene	0.400	0.396	1.0	68	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	66	0.00
30	Pyrene	0.400	0.421	-5.2	67	0.00
31 S	Terphenyl-d14	0.400	0.404	-1.0	67	0.00
32	Benzo(a)anthracene	0.400	0.385	3.8	66	0.00
33	Chrysene	0.400	0.413	-3.2	69	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.387	3.3	63	0.00
35 I	Perylene-d12	0.400	0.400	0.0	64	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.474	-18.5	75	-0.01
37	Benzo(b)fluoranthene	0.400	0.473	-18.2	70	0.00
38	Benzo(k)fluoranthene	0.400	0.394	1.5	69	0.00
39 C	Benzo(a)pyrene	0.400	0.391	2.3	70	0.00
40	Dibenzo(a,h)anthracene	0.400	0.434	-8.5	76	-0.01
41	Benzo(g,h,i)perylene	0.400	0.437	-9.2	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN110123