


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110218\
Data File : BN003396.D
Acq On : 02 Nov 2018 07:52
Operator : MJ/SJ
Sample : DFTPP42
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP42

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN110118.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Nov 05 13:38:10 2018



AutoFind: Scans 2491, 2492, 2493; Background Corrected with Scan 2484

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	27.4	243882	PASS
68	69	0.00	2	1.8	4825	PASS
69	198	0.00	100	29.5	263274	PASS
70	69	0.00	2	0.6	1507	PASS
127	198	10	80	41.2	367061	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	891392	PASS
199	198	5	9	6.8	60197	PASS
275	198	10	60	29.6	263594	PASS
365	198	1	100	3.5	30946	PASS
441	198	0.01	100	13.0	116045	PASS
442	198	50	100	83.9	747861	PASS
443	442	15	24	19.4	144941	PASS

m/z	Abundance
35.90	340.0
37.00	1025.0
38.00	3217.0
39.00	18824.0
40.00	1749.0
40.90	710.0
42.90	391.0
44.00	4995.0
44.90	442.0
49.00	1946.0
50.00	62904.0
51.00	248320.0
52.00	12909.0
53.10	609.0
55.00	1810.0
56.00	7359.0
57.00	16584.0
58.00	764.0
61.00	3331.0
62.00	3233.0
63.00	10232.0
64.00	1580.0
65.00	4923.0
66.00	470.0
67.10	328.0
68.10	4484.0
69.00	259648.0
70.00	1574.0
73.00	2265.0
74.00	26616.0
75.00	41248.0
76.00	14666.0
77.00	303296.0
78.00	21272.0
79.00	20416.0
80.00	15639.0
81.00	21472.0
82.00	5588.0
83.00	5007.0
84.10	970.0
85.00	3993.0
86.00	6529.0
87.00	3083.0
88.00	981.0
88.90	472.0
91.00	5085.0
92.00	5710.0
93.00	33656.0
94.00	2226.0
95.00	597.0
96.00	1570.0
97.00	547.0
98.00	25696.0
99.00	22952.0
100.00	2112.0
101.00	11581.0
102.10	726.0
103.00	3754.0
104.00	7924.0
105.00	6749.0
106.00	2522.0
107.00	91840.0
108.00	13831.0
109.00	2717.0
110.00	179520.0
111.00	26840.0
112.00	3435.0
113.00	1097.0
115.10	535.0
116.00	4572.0
117.00	77312.0
118.00	6511.0
119.00	935.0
119.90	1099.0
121.00	524.0
122.00	6414.0
123.00	9233.0
123.90	5096.0
125.00	3670.0
127.00	343808.0
128.00	27080.0
129.00	129416.0
130.00	11726.0
131.00	2370.0
132.00	1244.0
133.00	805.0
134.00	3761.0
135.00	10808.0
136.00	4335.0
137.00	6448.0
138.00	1300.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42

138.90	755.0
140.00	1634.0
141.00	18032.0
142.00	5747.0
143.00	3773.0
144.00	1218.0
145.00	782.0
146.00	3424.0
147.00	9027.0
148.00	20840.0
149.00	4539.0
150.00	1901.0
151.10	2287.0
151.80	1397.0
152.90	4984.0
154.00	4482.0
155.00	9638.0
156.00	14600.0
157.00	3377.0
157.90	3536.0
159.00	2792.0
159.90	5983.0
161.00	7772.0
161.90	2102.0
163.10	508.0
164.10	891.0
165.00	6480.0
166.00	5060.0
167.00	35296.0
168.00	24432.0
169.00	3714.0
170.10	1228.0
171.00	1321.0
171.90	3151.0
173.00	4334.0
174.00	8001.0
175.00	13552.0
176.00	4693.0
177.00	6915.0
178.00	2174.0
178.90	25888.0
180.00	19776.0
181.00	9422.0
182.10	1522.0
182.90	654.0
184.00	2194.0
185.00	13110.0
186.00	100544.0
187.00	28280.0
188.00	3971.0
188.90	5985.0
190.00	1423.0
191.00	2595.0
192.00	8012.0
193.00	10120.0
194.00	2187.0
195.10	1099.0
196.00	23384.0
198.00	771904.0
199.00	53944.0
200.00	4062.0
201.60	3109.0
203.00	4784.0
204.00	26472.0
205.00	43016.0
206.00	194944.0
207.00	27904.0
208.00	6945.0
209.00	2221.0
211.00	7638.0
213.00	463.0
213.80	305.0
215.00	2091.0
216.00	3801.0
217.00	48656.0
218.00	6461.0
219.00	686.0
221.00	35664.0
221.80	8127.0
223.00	11109.0
224.00	106376.0
225.00	27184.0
226.10	3238.0
227.00	41888.0
228.00	6045.0
229.00	9703.0
230.00	1096.0
231.00	3622.0
232.00	808.0
233.00	660.0
234.00	2856.0
235.00	3233.0
236.00	1981.0
237.00	4241.0
238.00	556.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42

239.00	1791.0
240.00	1671.0
241.00	2729.0
242.00	5522.0
243.00	6401.0
244.00	88216.0
245.00	11685.0
246.00	16349.0
247.00	3180.0
248.00	598.0
249.00	3259.0
250.00	642.0
251.00	719.0
252.10	546.0
253.00	2063.0
255.00	438656.0
256.00	66192.0
257.00	5527.0
258.00	25536.0
259.00	3972.0
259.90	833.0
260.90	759.0
264.00	761.0
265.00	10814.0
265.90	1491.0
266.90	325.0
269.90	313.0
271.00	824.0
272.00	1440.0
273.00	12771.0
274.00	35184.0
275.00	207872.0
276.00	29088.0
277.00	17248.0
278.00	2992.0
278.90	666.0
281.10	633.0
282.00	583.0
283.00	1906.0
283.90	1536.0
285.00	3293.0
285.90	521.0
288.90	907.0
289.90	541.0
291.10	312.0
292.00	645.0
293.00	3755.0
294.00	1346.0
296.00	54096.0
297.00	7511.0
298.10	645.0
300.90	626.0
302.00	969.0
303.00	6062.0
304.00	1620.0
305.10	328.0
308.00	840.0
309.00	612.0
310.00	654.0
312.90	606.0
314.00	2416.0
314.90	6595.0
315.90	2670.0
317.10	687.0
320.90	1893.0
321.90	694.0
323.00	16688.0
324.00	3338.0
325.00	348.0
325.80	309.0
327.00	3439.0
328.00	1597.0
329.00	432.0
332.00	1395.0
333.00	1810.0
334.00	11872.0
335.00	3023.0
336.00	361.0
341.10	1747.0
342.10	655.0
345.90	3614.0
347.00	587.0
350.90	345.0
352.00	4807.0
353.00	3340.0
354.10	5559.0
355.10	1158.0
359.10	454.0
365.00	22552.0
366.00	3638.0
369.80	569.0
371.00	1204.0
372.00	8646.0
373.00	2232.0
383.00	2498.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42

383.90	651.0
385.10	321.0
390.00	881.0
390.90	1033.0
392.10	490.0
401.00	538.0
402.00	3210.0
403.00	4846.0
404.00	2186.0
405.00	509.0
421.10	3824.0
422.10	3944.0
423.00	29288.0
424.10	8249.0
425.10	1009.0
441.00	76648.0
442.00	490688.0
443.00	98120.0
444.10	8997.0
445.10	415.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42

m/z	Abundance
36.00	398.0
37.00	1347.0
38.00	3257.0
39.00	21168.0
40.00	1905.0
41.00	924.0
43.00	393.0
44.00	5105.0
45.00	571.0
49.00	2474.0
50.00	73832.0
51.00	287680.0
52.00	14875.0
53.00	718.0
55.00	1656.0
56.00	9241.0
57.00	19544.0
58.00	754.0
61.00	3803.0
62.00	4318.0
63.00	11823.0
64.00	1662.0
65.00	5741.0
65.90	530.0
67.20	650.0
68.10	6079.0
69.00	311360.0
70.00	1717.0
73.00	2786.0
74.00	32360.0
75.00	50576.0
76.00	17816.0
77.00	364352.0
78.00	24544.0
79.00	23296.0
80.00	17912.0
81.00	25552.0
82.00	6912.0
83.00	5384.0
84.10	874.0
85.00	4914.0
86.00	7705.0
87.00	3371.0
87.90	1458.0
89.00	623.0
91.00	6002.0
92.00	6672.0
93.00	41624.0
94.00	2816.0
95.00	776.0
96.00	2253.0
97.00	967.0
98.00	32552.0
99.00	26784.0
99.90	2492.0
101.00	15065.0
102.00	1027.0
103.00	4719.0
104.00	10690.0
105.00	8648.0
106.00	2858.0
107.00	114552.0
108.00	18640.0
109.10	3617.0
110.00	222720.0
111.00	32368.0
112.00	3925.0
112.90	1479.0
114.30	346.0
115.10	469.0
116.00	6184.0
117.00	97280.0
118.00	7177.0
119.00	1197.0
120.00	1603.0
121.00	509.0
122.00	7806.0
123.00	12533.0
124.00	5933.0
125.00	5344.0
127.00	437952.0
128.00	34056.0
129.00	164864.0
130.00	13990.0
131.00	3045.0
132.00	1507.0
132.90	1046.0
134.00	5027.0
135.00	13270.0
136.00	5502.0
137.00	7515.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42

137.90	1458.0
138.90	888.0
140.00	1971.0
141.00	22272.0
142.00	6815.0
143.00	4810.0
144.00	1285.0
145.00	1274.0
145.90	4365.0
147.00	12072.0
148.00	26368.0
149.00	5462.0
150.00	2033.0
151.20	2828.0
151.80	1850.0
153.00	7075.0
153.90	5242.0
155.00	12512.0
156.00	19592.0
157.00	4193.0
157.90	4123.0
159.00	3006.0
160.00	7334.0
161.00	10936.0
162.00	3203.0
163.00	1049.0
164.00	1401.0
165.00	8512.0
166.00	7020.0
167.00	45864.0
168.00	26904.0
169.00	4163.0
170.00	1161.0
171.00	2255.0
172.00	4346.0
173.00	5400.0
174.00	9979.0
175.00	17624.0
176.00	5779.0
177.00	8472.0
178.00	2548.0
179.00	34536.0
180.00	24392.0
181.00	11640.0
182.00	2099.0
183.00	1000.0
183.90	3277.0
185.00	18616.0
186.00	135040.0
187.00	38752.0
188.00	4794.0
189.00	8683.0
190.00	1409.0
191.00	4336.0
192.00	10421.0
193.00	12940.0
194.00	2832.0
196.00	31136.0
198.00	1076224.0
199.00	72712.0
200.00	5920.0
201.50	5033.0
202.90	6774.0
204.00	35504.0
205.00	59544.0
206.00	273984.0
207.00	37720.0
208.00	8341.0
209.00	3525.0
210.20	4840.0
211.00	11122.0
213.00	848.0
214.00	460.0
215.00	2684.0
216.00	5369.0
217.00	67832.0
218.00	9133.0
219.00	1086.0
221.00	51792.0
223.00	15963.0
224.00	150912.0
225.00	39384.0
226.10	4337.0
227.00	59672.0
228.00	8400.0
229.00	13970.0
230.00	1797.0
231.00	5789.0
231.90	936.0
232.90	1002.0
233.90	3577.0
235.00	4121.0
236.00	3385.0
237.00	6069.0
238.00	902.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42

239.10	2341.0
240.00	1974.0
241.00	3755.0
242.00	7920.0
243.00	9295.0
244.00	129624.0
245.00	17000.0
246.00	22504.0
247.00	4636.0
247.90	1156.0
249.00	4444.0
250.00	953.0
250.90	895.0
252.00	1062.0
253.00	2506.0
255.00	635264.0
256.00	93448.0
257.00	7533.0
258.00	37776.0
259.00	5796.0
260.00	1517.0
261.00	1253.0
263.00	467.0
264.00	936.0
265.00	14687.0
266.00	2080.0
267.00	347.0
269.90	728.0
271.00	1393.0
272.00	1751.0
273.00	19032.0
274.00	53496.0
275.00	316480.0
276.00	42088.0
277.00	23848.0
278.00	3958.0
279.00	1070.0
280.90	837.0
281.90	814.0
283.00	2550.0
284.10	1925.0
285.00	4365.0
286.10	748.0
289.00	978.0
289.90	746.0
291.00	567.0
292.00	1215.0
293.00	5171.0
293.90	1599.0
296.00	80352.0
297.00	11254.0
298.00	680.0
300.90	1256.0
301.90	1556.0
303.00	9289.0
304.00	2364.0
304.90	354.0
308.00	1279.0
309.00	883.0
309.90	1257.0
311.10	358.0
313.00	937.0
314.10	3924.0
315.00	9123.0
316.00	5535.0
317.00	1095.0
319.80	308.0
321.00	2574.0
322.00	1538.0
323.00	27088.0
324.10	5347.0
325.00	619.0
326.00	578.0
327.00	5136.0
328.00	2157.0
329.00	421.0
332.00	2167.0
333.00	2739.0
334.00	17864.0
335.00	4520.0
336.10	476.0
338.90	405.0
340.00	502.0
341.00	3126.0
342.00	916.0
346.00	6391.0
347.00	1045.0
350.90	433.0
352.00	8098.0
353.00	5423.0
354.00	9693.0
355.10	1962.0
359.00	660.0
365.00	36016.0
366.00	5751.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42

367.00	340.0
370.00	806.0
371.00	2142.0
372.10	14644.0
373.00	3471.0
374.00	467.0
377.00	613.0
383.00	4166.0
383.90	1309.0
390.00	1733.0
391.10	1186.0
392.00	927.0
401.00	1006.0
402.00	5724.0
403.00	8795.0
404.00	3283.0
405.10	599.0
409.90	375.0
414.90	385.0
421.00	7719.0
422.00	6714.0
423.00	49864.0
424.00	12480.0
425.00	1454.0
441.00	136384.0
442.00	848320.0
443.00	167808.0
444.10	15808.0
445.00	851.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42

m/z	Abundance
36.00	349.0
36.90	703.0
38.00	2311.0
39.00	14135.0
40.00	1691.0
41.00	697.0
43.00	337.0
44.00	4996.0
45.00	438.0
49.10	1653.0
50.00	48856.0
51.00	195648.0
52.00	9960.0
53.00	518.0
55.00	1370.0
56.00	5527.0
57.00	13193.0
57.90	511.0
61.00	2783.0
62.00	3112.0
63.00	8224.0
64.00	1177.0
65.00	4115.0
67.10	383.0
68.00	3913.0
69.00	218816.0
70.00	1232.0
73.00	2195.0
74.00	22504.0
75.00	34968.0
76.10	12228.0
77.00	253376.0
78.00	17520.0
79.00	16064.0
80.00	13129.0
81.00	18152.0
82.00	4580.0
83.00	4379.0
84.00	679.0
85.00	3847.0
86.00	5371.0
87.00	2789.0
87.90	1110.0
89.10	400.0
91.00	4303.0
92.00	5062.0
93.00	29688.0
94.00	2185.0
94.90	569.0
96.00	1671.0
97.00	489.0
98.00	23224.0
99.00	19440.0
99.90	1568.0
101.00	10646.0
102.00	704.0
103.00	3384.0
103.90	6768.0
105.00	6383.0
106.00	1948.0
107.00	80800.0
108.00	12974.0
109.00	2365.0
110.00	157376.0
111.00	24232.0
112.00	3070.0
113.00	943.0
114.00	342.0
114.80	466.0
116.00	4745.0
117.00	69448.0
118.00	4893.0
118.90	648.0
120.00	1264.0
121.00	359.0
122.00	5706.0
123.00	9393.0
124.00	4464.0
125.00	3663.0
127.00	319424.0
128.00	24016.0
129.00	122920.0
130.00	10211.0
131.00	2230.0
132.00	1377.0
133.00	621.0
134.00	3552.0
135.00	10305.0
136.00	4289.0
137.00	5766.0
137.80	1123.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42

138.90	664.0
140.00	1637.0
141.00	16632.0
142.00	5451.0
142.90	3081.0
144.00	1118.0
145.00	1087.0
146.00	2997.0
147.00	8149.0
148.00	19080.0
149.00	4179.0
150.00	1335.0
151.10	2619.0
151.90	1218.0
153.00	5537.0
153.90	4165.0
155.00	9777.0
156.00	13787.0
157.00	2887.0
157.90	3378.0
159.00	2435.0
160.00	5188.0
161.00	7977.0
161.90	2335.0
163.00	770.0
164.00	896.0
164.90	6264.0
166.00	5139.0
167.00	34544.0
168.00	20624.0
169.00	3298.0
170.00	1290.0
170.90	1896.0
171.90	2894.0
173.00	4259.0
174.00	7947.0
175.00	14735.0
176.00	3944.0
176.90	6893.0
177.90	2380.0
178.90	27728.0
180.00	20088.0
181.00	8606.0
182.00	1441.0
182.80	665.0
183.90	2008.0
185.00	12500.0
186.00	106800.0
187.00	31392.0
188.00	3477.0
189.00	6718.0
190.10	1044.0
191.10	3207.0
192.00	8718.0
193.00	9658.0
194.10	1837.0
195.00	1345.0
196.00	25336.0
198.00	826048.0
199.00	53936.0
200.00	4695.0
201.40	3875.0
203.00	5116.0
204.00	28352.0
205.00	48120.0
206.00	214080.0
207.00	29552.0
208.00	7403.0
209.00	2762.0
210.30	3971.0
211.10	8824.0
212.90	494.0
214.00	320.0
215.00	2278.0
216.00	4855.0
217.00	54704.0
218.00	7438.0
219.00	950.0
221.00	40064.0
223.00	12511.0
224.00	121720.0
225.10	30344.0
226.00	3575.0
227.00	47264.0
228.00	7186.0
229.00	9744.0
230.00	1452.0
231.00	4679.0
231.90	670.0
233.10	961.0
233.90	3248.0
235.00	3648.0
236.00	2487.0
237.00	4485.0
237.90	556.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42

239.00	1906.0
240.00	1692.0
241.00	3115.0
242.00	7035.0
243.10	7085.0
244.00	104600.0
245.00	14391.0
246.00	18776.0
247.00	3813.0
247.90	1002.0
249.00	3844.0
250.00	921.0
251.00	1145.0
252.10	979.0
253.00	1946.0
255.00	533696.0
256.00	78088.0
257.00	6485.0
258.00	29592.0
259.00	5063.0
260.00	1254.0
261.00	1149.0
264.00	986.0
265.00	12369.0
266.00	1418.0
266.90	330.0
269.90	712.0
270.90	1277.0
272.00	1391.0
273.00	16920.0
274.00	44768.0
275.00	266432.0
276.00	35040.0
277.00	20648.0
277.90	3591.0
278.90	813.0
281.00	591.0
282.00	765.0
283.00	2431.0
284.00	1558.0
285.00	3600.0
286.00	771.0
288.90	993.0
289.90	612.0
291.00	640.0
292.00	1130.0
293.00	4636.0
294.00	1145.0
295.00	1715.0
296.00	70760.0
297.00	9386.0
297.90	542.0
301.00	832.0
302.00	1255.0
303.00	7959.0
304.00	2093.0
305.00	351.0
307.90	1140.0
308.80	565.0
310.00	684.0
312.10	307.0
313.00	618.0
314.00	3615.0
315.00	8671.0
316.00	4448.0
317.00	919.0
321.00	2499.0
322.10	1365.0
323.00	24088.0
324.00	4346.0
325.10	467.0
325.90	514.0
327.00	4916.0
328.00	2799.0
328.90	416.0
332.00	1832.0
333.00	2415.0
334.00	16424.0
335.00	4257.0
336.10	568.0
339.00	423.0
340.00	439.0
341.00	2774.0
342.00	1014.0
346.00	5774.0
346.90	1186.0
348.00	331.0
351.00	372.0
352.00	8132.0
353.00	5303.0
354.00	8268.0
355.00	1585.0
358.90	587.0
365.00	34272.0
366.00	5236.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42

367.10	427.0
370.00	735.0
371.00	1935.0
372.00	14212.0
373.00	3453.0
373.90	333.0
377.00	361.0
382.90	3575.0
384.00	963.0
384.90	314.0
390.00	1638.0
391.00	1268.0
391.90	1073.0
401.00	882.0
402.00	5409.0
403.00	8915.0
404.00	2979.0
405.00	656.0
414.80	331.0
421.00	6455.0
422.00	6629.0
423.00	51000.0
424.00	10307.0
425.00	1115.0
441.10	135104.0
442.00	904576.0
443.00	168896.0
444.10	15955.0
445.00	947.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42

Scan 2484 (17.598 min): BN003396.D
DFTPP42

m/z	Abundance
40.00	1272.0
40.90	381.0
44.00	5427.0
54.90	345.0
73.00	397.0
77.00	334.0
147.00	394.0
190.90	312.0
207.00	2174.0
208.00	568.0
208.90	374.0
281.00	669.0

Instrument :
BNA_N
ClientSampleId :
DFTPP42