

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110223\
 Data File : BN028427.D
 Acq On : 02 Nov 2023 18:35
 Operator : MA/JU
 Sample : PB156693BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB156693BL

Quant Time: Nov 02 19:28:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 02 06:26:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.724	152	11326	0.400	ng	0.00
7) Naphthalene-d8	10.509	136	31576	0.400	ng	0.00
13) Acenaphthene-d10	14.363	164	15584	0.400	ng	0.00
19) Phenanthrene-d10	17.109	188	34077	0.400	ng	#-0.01
29) Chrysene-d12	21.328	240	22172	0.400	ng	0.00
35) Perylene-d12	23.637	264	19941	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.305	112	11104	0.375	ng	0.00
5) Phenol-d6	6.894	99	13357	0.356	ng	0.00
8) Nitrobenzene-d5	8.875	82	7631	0.374	ng	0.00
11) 2-Methylnaphthalene-d10	12.110	152	17264	0.382	ng	0.00
14) 2,4,6-Tribromophenol	15.855	330	1763	0.290	ng	0.00
15) 2-Fluorobiphenyl	13.006	172	6029	0.326	ng	0.00
27) Fluoranthene-d10	19.152	212	31787	0.377	ng	0.00
31) Terphenyl-d14	19.765	244	20607	0.441	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110223\
Data File : BN028427.D
Acq On : 02 Nov 2023 18:35
Operator : MA/JU
Sample : PB156693BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB156693BL

Quant Time: Nov 02 19:28:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 02 06:26:50 2023
Response via : Initial Calibration

