

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110518\
 Data File : BN003422.D
 Acq On : 05 Nov 2018 10:38
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.437

Quant Time: Nov 05 13:40:20 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN110118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 02 23:49:03 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	82	0.00
3	Naphthalene	1.122	1.140	-1.6	84	0.00
4 SURR	2-Methylnaphthalene-d10	0.640	0.649	-1.4	84	0.00
5	2-Methylnaphthalene	0.807	0.832	-3.1	85	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	80	0.00
7	Acenaphthylene	1.738	2.070	-19.1	95	0.00
8 C	Acenaphthene	1.553	1.596	-2.8	82	0.00
9	Fluorene	1.829	1.805	1.3	79	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	71	0.00
11	Pentachlorophenol	0.080	0.087	-8.7	84	0.00
12	Phenanthrene	1.312	1.312	0.0	72	0.00
13	Anthracene	1.069	1.185	-10.9	79	0.00
14 SURR	Fluoranthene-d10	1.124	1.084	3.6	69	0.00
15 C	Fluoranthene	1.474	1.393	5.5	67	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	53	0.00
17	Pyrene	1.455	1.805	-24.1#	67	0.00
18	Benzo(a)anthracene	1.254	1.487	-18.6	63	0.00
19	Chrysene	1.425	1.419	0.4	53	0.00
20 I	Perylene-d12	1.000	1.000	0.0	58	0.00
21	Benzo(b)fluoranthene	1.724	1.527	11.4	53	0.00
22	Benzo(k)fluoranthene	1.669	1.500	10.1	54	0.00
23 C	Benzo(a)pyrene	1.439	1.450	-0.8	59	0.00
24	Indeno(1,2,3-cd)pyrene	1.541	1.717	-11.4	64	0.00
25	Dibenzo(a,h)anthracene	1.256	1.362	-8.4	62	0.00
26	Benzo(g,h,i)perylene	1.321	1.476	-11.7	64	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0