


Data Path : Z:\SVOASRV\HPCHEM1\BNA\_N\DATA\BN110518\  
Data File : BN003421.D  
Acq On : 05 Nov 2018 10:03  
Operator : MJ/SJ  
Sample : DFTPP43  
Misc :  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SOM-EPA-SIM-BN110118.M  
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
Last Update : Tue Nov 06 09:59:55 2018



AutoFind: Scans 2491, 2492, 2493; Background Corrected with Scan 2484

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 25.8      | 193472  | PASS             |
| 68          | 69           | 0.00         | 2            | 1.9       | 4034    | PASS             |
| 69          | 198          | 0.00         | 100          | 28.4      | 213033  | PASS             |
| 70          | 69           | 0.00         | 2            | 0.4       | 870     | PASS             |
| 127         | 198          | 10           | 80           | 40.6      | 304640  | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 749845  | PASS             |
| 199         | 198          | 5            | 9            | 6.9       | 51509   | PASS             |
| 275         | 198          | 10           | 60           | 30.2      | 226410  | PASS             |
| 365         | 198          | 1            | 100          | 3.6       | 27232   | PASS             |
| 441         | 198          | 0.01         | 100          | 13.9      | 104056  | PASS             |
| 442         | 198          | 50           | 100          | 88.0      | 659712  | PASS             |
| 443         | 442          | 15           | 24           | 19.5      | 128773  | PASS             |

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 841.0     |
| 38.00  | 2286.0    |
| 39.00  | 15617.0   |
| 40.00  | 1738.0    |
| 41.00  | 1101.0    |
| 43.10  | 948.0     |
| 44.00  | 7487.0    |
| 44.90  | 519.0     |
| 49.00  | 1865.0    |
| 50.00  | 53144.0   |
| 51.00  | 221056.0  |
| 52.00  | 11486.0   |
| 53.00  | 523.0     |
| 55.00  | 1651.0    |
| 56.00  | 6704.0    |
| 57.00  | 15967.0   |
| 57.90  | 721.0     |
| 61.00  | 3016.0    |
| 62.00  | 3405.0    |
| 63.00  | 8381.0    |
| 64.00  | 1508.0    |
| 65.00  | 4405.0    |
| 65.90  | 387.0     |
| 67.00  | 576.0     |
| 68.10  | 4760.0    |
| 69.00  | 240640.0  |
| 70.00  | 1354.0    |
| 71.00  | 888.0     |
| 73.00  | 2210.0    |
| 74.00  | 24512.0   |
| 75.00  | 36680.0   |
| 76.00  | 13209.0   |
| 77.00  | 281280.0  |
| 78.00  | 19992.0   |
| 79.00  | 18504.0   |
| 80.00  | 13461.0   |
| 81.00  | 20544.0   |
| 82.00  | 4861.0    |
| 83.00  | 5214.0    |
| 84.10  | 782.0     |
| 85.00  | 4198.0    |
| 86.00  | 5498.0    |
| 87.00  | 2544.0    |
| 88.00  | 1084.0    |
| 89.00  | 675.0     |
| 91.00  | 4068.0    |
| 92.00  | 5194.0    |
| 93.00  | 31512.0   |
| 94.00  | 2075.0    |
| 95.00  | 558.0     |
| 96.10  | 1799.0    |
| 97.10  | 992.0     |
| 98.00  | 23736.0   |
| 99.00  | 19384.0   |
| 100.00 | 1911.0    |
| 101.00 | 11329.0   |
| 102.00 | 665.0     |
| 103.00 | 3972.0    |
| 103.90 | 7471.0    |
| 105.00 | 6646.0    |
| 106.00 | 2146.0    |
| 107.00 | 86472.0   |
| 108.00 | 13763.0   |
| 109.10 | 3091.0    |
| 110.00 | 172352.0  |
| 111.00 | 26456.0   |
| 112.00 | 3156.0    |
| 112.90 | 1101.0    |
| 115.10 | 533.0     |
| 116.00 | 4960.0    |
| 117.00 | 71232.0   |
| 118.00 | 5705.0    |
| 119.00 | 890.0     |
| 119.90 | 1424.0    |
| 121.10 | 532.0     |
| 122.00 | 5533.0    |
| 123.00 | 8785.0    |
| 124.00 | 4455.0    |
| 125.00 | 3825.0    |
| 127.00 | 328320.0  |
| 128.00 | 24296.0   |
| 129.00 | 122128.0  |
| 130.00 | 10610.0   |
| 131.00 | 1828.0    |
| 132.10 | 1435.0    |
| 132.90 | 955.0     |
| 134.00 | 3816.0    |
| 135.00 | 9361.0    |
| 135.90 | 4132.0    |
| 137.00 | 5547.0    |
| 138.00 | 1036.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

|        |          |
|--------|----------|
| 138.90 | 886.0    |
| 140.00 | 1796.0   |
| 141.00 | 16968.0  |
| 142.00 | 5950.0   |
| 143.00 | 3824.0   |
| 144.00 | 918.0    |
| 145.10 | 1256.0   |
| 145.90 | 3115.0   |
| 147.00 | 8267.0   |
| 148.00 | 17984.0  |
| 149.00 | 4275.0   |
| 150.00 | 1298.0   |
| 151.20 | 2070.0   |
| 151.80 | 1496.0   |
| 153.00 | 5890.0   |
| 154.00 | 4034.0   |
| 155.00 | 8574.0   |
| 156.00 | 13971.0  |
| 157.00 | 2932.0   |
| 157.90 | 3126.0   |
| 159.00 | 2201.0   |
| 160.00 | 5217.0   |
| 161.00 | 8193.0   |
| 162.00 | 2561.0   |
| 163.00 | 830.0    |
| 163.90 | 1045.0   |
| 164.90 | 6376.0   |
| 166.10 | 5027.0   |
| 167.00 | 31984.0  |
| 168.00 | 19416.0  |
| 169.00 | 3093.0   |
| 170.00 | 980.0    |
| 170.90 | 1740.0   |
| 171.90 | 2847.0   |
| 173.00 | 3672.0   |
| 174.00 | 7120.0   |
| 175.00 | 13448.0  |
| 176.00 | 4444.0   |
| 177.00 | 6365.0   |
| 177.90 | 2395.0   |
| 179.00 | 25176.0  |
| 180.00 | 17872.0  |
| 181.00 | 8401.0   |
| 182.00 | 1548.0   |
| 182.90 | 1026.0   |
| 184.00 | 2072.0   |
| 185.00 | 12329.0  |
| 186.00 | 96504.0  |
| 187.00 | 28704.0  |
| 188.00 | 3447.0   |
| 189.00 | 6026.0   |
| 190.00 | 835.0    |
| 191.00 | 3128.0   |
| 192.00 | 7981.0   |
| 193.00 | 8365.0   |
| 194.00 | 2135.0   |
| 195.00 | 1435.0   |
| 196.00 | 22744.0  |
| 198.00 | 754240.0 |
| 199.00 | 53624.0  |
| 200.00 | 4535.0   |
| 201.40 | 4241.0   |
| 203.00 | 4592.0   |
| 204.00 | 25048.0  |
| 205.00 | 42456.0  |
| 206.00 | 190336.0 |
| 207.00 | 27360.0  |
| 208.00 | 6569.0   |
| 209.00 | 2279.0   |
| 210.20 | 3439.0   |
| 211.00 | 7343.0   |
| 213.00 | 379.0    |
| 214.90 | 2088.0   |
| 216.00 | 3716.0   |
| 217.00 | 46696.0  |
| 218.00 | 5845.0   |
| 218.90 | 591.0    |
| 221.00 | 35744.0  |
| 223.00 | 10957.0  |
| 224.00 | 104384.0 |
| 225.00 | 27240.0  |
| 226.00 | 3177.0   |
| 227.00 | 40104.0  |
| 228.00 | 6340.0   |
| 229.00 | 9151.0   |
| 230.00 | 999.0    |
| 231.00 | 4022.0   |
| 232.00 | 645.0    |
| 233.10 | 546.0    |
| 234.00 | 2675.0   |
| 235.00 | 3303.0   |
| 236.00 | 1908.0   |
| 237.00 | 4260.0   |
| 238.10 | 598.0    |
| 239.00 | 1503.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

|        |          |
|--------|----------|
| 240.00 | 1388.0   |
| 241.00 | 2795.0   |
| 242.00 | 6214.0   |
| 243.10 | 5897.0   |
| 244.00 | 84752.0  |
| 245.00 | 12009.0  |
| 246.00 | 14595.0  |
| 247.00 | 3520.0   |
| 248.00 | 930.0    |
| 249.00 | 3086.0   |
| 250.00 | 677.0    |
| 251.00 | 741.0    |
| 251.90 | 741.0    |
| 253.00 | 1813.0   |
| 255.00 | 433600.0 |
| 256.00 | 64352.0  |
| 257.00 | 5316.0   |
| 258.00 | 23464.0  |
| 259.00 | 4020.0   |
| 260.00 | 647.0    |
| 261.00 | 681.0    |
| 264.10 | 720.0    |
| 265.00 | 10526.0  |
| 266.00 | 1451.0   |
| 267.10 | 353.0    |
| 269.80 | 375.0    |
| 271.00 | 872.0    |
| 272.10 | 1228.0   |
| 273.00 | 13740.0  |
| 274.00 | 35504.0  |
| 275.00 | 204480.0 |
| 276.00 | 28424.0  |
| 277.00 | 16576.0  |
| 278.00 | 2947.0   |
| 279.10 | 588.0    |
| 281.10 | 691.0    |
| 282.10 | 657.0    |
| 283.00 | 1790.0   |
| 284.00 | 1062.0   |
| 285.00 | 2680.0   |
| 285.80 | 453.0    |
| 287.90 | 324.0    |
| 288.90 | 899.0    |
| 290.10 | 459.0    |
| 291.00 | 454.0    |
| 292.00 | 673.0    |
| 293.00 | 3697.0   |
| 294.00 | 1035.0   |
| 295.00 | 1271.0   |
| 296.00 | 52328.0  |
| 297.00 | 7534.0   |
| 298.00 | 613.0    |
| 301.00 | 596.0    |
| 301.90 | 919.0    |
| 303.00 | 5999.0   |
| 304.00 | 1738.0   |
| 305.00 | 325.0    |
| 307.90 | 860.0    |
| 309.00 | 643.0    |
| 310.00 | 756.0    |
| 313.00 | 516.0    |
| 314.00 | 2630.0   |
| 315.00 | 5871.0   |
| 316.00 | 2888.0   |
| 317.00 | 737.0    |
| 321.00 | 1859.0   |
| 323.00 | 18160.0  |
| 324.10 | 3357.0   |
| 325.00 | 311.0    |
| 325.90 | 399.0    |
| 327.00 | 3809.0   |
| 328.00 | 1900.0   |
| 329.20 | 375.0    |
| 332.10 | 1520.0   |
| 333.00 | 1880.0   |
| 334.00 | 10811.0  |
| 335.00 | 3486.0   |
| 336.10 | 484.0    |
| 338.90 | 389.0    |
| 341.00 | 2078.0   |
| 341.90 | 619.0    |
| 346.00 | 4125.0   |
| 347.00 | 854.0    |
| 352.00 | 5306.0   |
| 353.00 | 3410.0   |
| 354.10 | 5598.0   |
| 355.00 | 1320.0   |
| 359.00 | 363.0    |
| 365.00 | 23368.0  |
| 366.00 | 3630.0   |
| 370.00 | 495.0    |
| 371.00 | 1282.0   |
| 372.00 | 8588.0   |
| 373.00 | 2436.0   |
| 383.00 | 2369.0   |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

|        |          |
|--------|----------|
| 384.00 | 899.0    |
| 390.00 | 1410.0   |
| 391.10 | 953.0    |
| 392.00 | 689.0    |
| 401.00 | 517.0    |
| 402.00 | 3335.0   |
| 403.00 | 5311.0   |
| 404.00 | 1716.0   |
| 405.10 | 352.0    |
| 421.00 | 4043.0   |
| 422.00 | 4060.0   |
| 423.00 | 29152.0  |
| 424.00 | 7430.0   |
| 425.00 | 946.0    |
| 441.00 | 81048.0  |
| 442.00 | 492928.0 |
| 443.10 | 103616.0 |
| 444.00 | 10198.0  |
| 445.00 | 585.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 1019.0    |
| 38.00  | 2556.0    |
| 39.00  | 16314.0   |
| 40.00  | 1913.0    |
| 41.10  | 1200.0    |
| 43.00  | 953.0     |
| 44.00  | 7863.0    |
| 45.00  | 369.0     |
| 49.10  | 1789.0    |
| 50.00  | 57808.0   |
| 51.00  | 225216.0  |
| 52.00  | 11583.0   |
| 52.90  | 500.0     |
| 55.00  | 2052.0    |
| 56.00  | 7515.0    |
| 57.00  | 17048.0   |
| 58.00  | 703.0     |
| 61.00  | 2884.0    |
| 62.00  | 3817.0    |
| 63.00  | 8910.0    |
| 64.00  | 1305.0    |
| 65.00  | 4807.0    |
| 65.90  | 440.0     |
| 67.10  | 552.0     |
| 68.00  | 4531.0    |
| 69.00  | 249216.0  |
| 70.00  | 1361.0    |
| 71.00  | 691.0     |
| 73.00  | 2530.0    |
| 74.00  | 25984.0   |
| 75.00  | 41080.0   |
| 76.00  | 14028.0   |
| 77.00  | 293888.0  |
| 78.00  | 20768.0   |
| 79.00  | 19432.0   |
| 80.00  | 14801.0   |
| 81.00  | 20520.0   |
| 82.00  | 5146.0    |
| 83.00  | 4938.0    |
| 84.00  | 816.0     |
| 85.00  | 4596.0    |
| 86.00  | 6104.0    |
| 87.00  | 2959.0    |
| 88.00  | 1002.0    |
| 89.00  | 532.0     |
| 91.00  | 5184.0    |
| 92.00  | 5355.0    |
| 93.00  | 33504.0   |
| 94.00  | 2538.0    |
| 95.00  | 896.0     |
| 96.00  | 1748.0    |
| 97.10  | 710.0     |
| 98.00  | 26736.0   |
| 99.00  | 21048.0   |
| 100.00 | 2074.0    |
| 101.00 | 12746.0   |
| 102.00 | 711.0     |
| 103.00 | 4401.0    |
| 104.00 | 7728.0    |
| 105.00 | 7196.0    |
| 106.00 | 2441.0    |
| 107.00 | 90512.0   |
| 108.00 | 14693.0   |
| 109.10 | 2711.0    |
| 110.00 | 183552.0  |
| 111.00 | 27624.0   |
| 112.00 | 3461.0    |
| 113.00 | 1151.0    |
| 114.80 | 469.0     |
| 116.00 | 5448.0    |
| 117.00 | 75120.0   |
| 118.00 | 5752.0    |
| 119.00 | 982.0     |
| 120.10 | 1413.0    |
| 121.10 | 555.0     |
| 122.00 | 6523.0    |
| 123.00 | 9887.0    |
| 124.00 | 4430.0    |
| 125.00 | 3943.0    |
| 127.00 | 358400.0  |
| 128.00 | 26560.0   |
| 129.00 | 135360.0  |
| 130.00 | 11385.0   |
| 131.00 | 2379.0    |
| 132.00 | 1427.0    |
| 133.00 | 771.0     |
| 134.00 | 4476.0    |
| 135.00 | 12032.0   |
| 136.00 | 4978.0    |
| 137.00 | 6158.0    |
| 137.80 | 1214.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

|        |          |
|--------|----------|
| 138.90 | 869.0    |
| 140.00 | 1790.0   |
| 141.00 | 17072.0  |
| 142.00 | 5613.0   |
| 143.00 | 3851.0   |
| 144.00 | 1142.0   |
| 144.90 | 1422.0   |
| 146.00 | 3077.0   |
| 147.00 | 9760.0   |
| 148.00 | 20872.0  |
| 149.00 | 4504.0   |
| 150.00 | 1420.0   |
| 151.10 | 2709.0   |
| 151.90 | 1751.0   |
| 153.00 | 5964.0   |
| 154.00 | 4563.0   |
| 155.00 | 9791.0   |
| 156.00 | 15685.0  |
| 157.00 | 3811.0   |
| 157.90 | 3851.0   |
| 159.00 | 2642.0   |
| 160.00 | 5938.0   |
| 161.00 | 9127.0   |
| 162.00 | 2564.0   |
| 163.10 | 822.0    |
| 164.00 | 1279.0   |
| 165.00 | 7371.0   |
| 166.00 | 5987.0   |
| 167.00 | 37344.0  |
| 168.00 | 19896.0  |
| 169.10 | 3561.0   |
| 170.10 | 1476.0   |
| 171.00 | 1717.0   |
| 172.00 | 3362.0   |
| 173.00 | 4606.0   |
| 174.00 | 8157.0   |
| 175.00 | 15541.0  |
| 176.00 | 4290.0   |
| 177.00 | 7696.0   |
| 177.90 | 2494.0   |
| 179.00 | 28992.0  |
| 180.00 | 19248.0  |
| 181.00 | 8935.0   |
| 182.00 | 1375.0   |
| 183.00 | 989.0    |
| 184.00 | 2463.0   |
| 185.00 | 14558.0  |
| 186.00 | 113112.0 |
| 187.00 | 31480.0  |
| 188.10 | 3747.0   |
| 189.00 | 7215.0   |
| 190.00 | 1315.0   |
| 191.00 | 3647.0   |
| 192.00 | 9622.0   |
| 193.00 | 10360.0  |
| 194.00 | 2437.0   |
| 194.90 | 1754.0   |
| 196.00 | 27272.0  |
| 198.00 | 891968.0 |
| 199.00 | 60624.0  |
| 200.00 | 4490.0   |
| 201.40 | 4777.0   |
| 203.00 | 5333.0   |
| 204.00 | 29120.0  |
| 205.00 | 51280.0  |
| 206.00 | 224384.0 |
| 207.00 | 31368.0  |
| 208.00 | 7962.0   |
| 209.00 | 2513.0   |
| 211.00 | 9568.0   |
| 213.10 | 794.0    |
| 214.00 | 320.0    |
| 215.00 | 2317.0   |
| 216.00 | 4882.0   |
| 217.00 | 58200.0  |
| 218.00 | 7207.0   |
| 219.00 | 875.0    |
| 221.00 | 41448.0  |
| 221.80 | 9176.0   |
| 223.00 | 13864.0  |
| 224.00 | 125456.0 |
| 225.00 | 31176.0  |
| 226.00 | 4057.0   |
| 227.00 | 50640.0  |
| 228.00 | 6895.0   |
| 229.00 | 11190.0  |
| 230.00 | 1500.0   |
| 231.00 | 4840.0   |
| 232.00 | 726.0    |
| 233.00 | 663.0    |
| 233.90 | 3172.0   |
| 235.00 | 3547.0   |
| 236.00 | 2371.0   |
| 237.00 | 4540.0   |
| 238.00 | 823.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

|        |          |
|--------|----------|
| 239.00 | 2176.0   |
| 239.90 | 1827.0   |
| 241.00 | 3146.0   |
| 242.00 | 6506.0   |
| 243.10 | 7641.0   |
| 244.00 | 106616.0 |
| 245.00 | 13949.0  |
| 246.00 | 18432.0  |
| 247.00 | 4032.0   |
| 248.00 | 939.0    |
| 249.00 | 4379.0   |
| 249.90 | 620.0    |
| 250.90 | 933.0    |
| 252.00 | 890.0    |
| 253.00 | 2197.0   |
| 255.00 | 551232.0 |
| 256.00 | 81088.0  |
| 257.00 | 6372.0   |
| 258.00 | 30240.0  |
| 259.00 | 4836.0   |
| 260.00 | 942.0    |
| 260.90 | 1043.0   |
| 263.00 | 405.0    |
| 263.90 | 811.0    |
| 265.00 | 11674.0  |
| 266.00 | 1650.0   |
| 267.00 | 309.0    |
| 270.00 | 827.0    |
| 271.00 | 1204.0   |
| 272.10 | 1349.0   |
| 273.00 | 17336.0  |
| 274.00 | 44504.0  |
| 275.00 | 278272.0 |
| 276.00 | 36896.0  |
| 277.00 | 21216.0  |
| 278.00 | 3444.0   |
| 279.00 | 1030.0   |
| 281.00 | 660.0    |
| 282.00 | 708.0    |
| 283.00 | 2394.0   |
| 284.00 | 1677.0   |
| 285.10 | 3760.0   |
| 286.00 | 894.0    |
| 287.90 | 366.0    |
| 289.00 | 933.0    |
| 289.90 | 800.0    |
| 291.10 | 318.0    |
| 292.00 | 1187.0   |
| 293.00 | 5051.0   |
| 294.00 | 1179.0   |
| 296.00 | 71312.0  |
| 297.00 | 9579.0   |
| 297.90 | 671.0    |
| 301.00 | 1095.0   |
| 302.00 | 1201.0   |
| 303.00 | 8601.0   |
| 304.00 | 2415.0   |
| 305.00 | 379.0    |
| 308.00 | 953.0    |
| 309.00 | 752.0    |
| 310.10 | 798.0    |
| 310.90 | 349.0    |
| 312.00 | 362.0    |
| 313.00 | 910.0    |
| 314.00 | 3945.0   |
| 315.00 | 7892.0   |
| 316.00 | 4923.0   |
| 316.90 | 731.0    |
| 320.10 | 307.0    |
| 321.00 | 1964.0   |
| 322.00 | 1267.0   |
| 323.00 | 24408.0  |
| 324.00 | 4060.0   |
| 324.90 | 505.0    |
| 326.00 | 472.0    |
| 326.90 | 4174.0   |
| 328.00 | 2694.0   |
| 328.90 | 411.0    |
| 332.00 | 1797.0   |
| 333.00 | 2327.0   |
| 334.00 | 14786.0  |
| 335.00 | 4503.0   |
| 336.20 | 510.0    |
| 339.10 | 484.0    |
| 339.90 | 502.0    |
| 341.00 | 3072.0   |
| 342.00 | 672.0    |
| 345.90 | 5523.0   |
| 346.90 | 754.0    |
| 351.10 | 382.0    |
| 352.00 | 7179.0   |
| 353.00 | 5324.0   |
| 354.10 | 8816.0   |
| 355.10 | 1395.0   |
| 358.90 | 715.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

|        |          |
|--------|----------|
| 365.00 | 32272.0  |
| 366.00 | 4630.0   |
| 366.90 | 505.0    |
| 370.00 | 958.0    |
| 370.90 | 2173.0   |
| 372.00 | 13228.0  |
| 373.00 | 3389.0   |
| 374.00 | 474.0    |
| 377.10 | 357.0    |
| 383.00 | 3527.0   |
| 384.00 | 1116.0   |
| 385.00 | 358.0    |
| 390.00 | 1618.0   |
| 391.00 | 1074.0   |
| 392.00 | 910.0    |
| 401.00 | 923.0    |
| 402.00 | 5639.0   |
| 403.00 | 7627.0   |
| 404.10 | 2202.0   |
| 405.00 | 599.0    |
| 414.90 | 515.0    |
| 421.00 | 6311.0   |
| 422.00 | 6233.0   |
| 423.00 | 44480.0  |
| 424.10 | 10011.0  |
| 425.10 | 1360.0   |
| 441.10 | 120864.0 |
| 442.00 | 793856.0 |
| 443.10 | 153472.0 |
| 444.00 | 14938.0  |
| 445.00 | 868.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

| m/z    | Abundance |
|--------|-----------|
| 37.00  | 546.0     |
| 38.00  | 1673.0    |
| 39.00  | 10403.0   |
| 40.00  | 1434.0    |
| 41.00  | 1157.0    |
| 43.10  | 1541.0    |
| 44.00  | 8335.0    |
| 44.90  | 556.0     |
| 49.00  | 1203.0    |
| 50.00  | 34480.0   |
| 51.00  | 134144.0  |
| 52.00  | 6787.0    |
| 53.00  | 466.0     |
| 55.00  | 1255.0    |
| 56.00  | 4621.0    |
| 57.00  | 10489.0   |
| 57.90  | 562.0     |
| 61.00  | 1810.0    |
| 62.00  | 1900.0    |
| 63.00  | 5360.0    |
| 64.00  | 763.0     |
| 65.00  | 2730.0    |
| 65.90  | 324.0     |
| 66.90  | 446.0     |
| 68.00  | 2813.0    |
| 69.00  | 151360.0  |
| 70.00  | 858.0     |
| 71.00  | 1134.0    |
| 73.00  | 1457.0    |
| 74.00  | 14959.0   |
| 75.00  | 24648.0   |
| 76.00  | 8532.0    |
| 77.00  | 177344.0  |
| 78.00  | 11801.0   |
| 79.00  | 12280.0   |
| 80.00  | 8818.0    |
| 81.00  | 12732.0   |
| 82.00  | 3242.0    |
| 83.00  | 3053.0    |
| 84.10  | 585.0     |
| 85.00  | 2852.0    |
| 86.00  | 3696.0    |
| 87.00  | 1739.0    |
| 88.00  | 690.0     |
| 89.00  | 311.0     |
| 91.00  | 3608.0    |
| 92.00  | 3505.0    |
| 93.00  | 20640.0   |
| 94.00  | 1326.0    |
| 95.10  | 553.0     |
| 96.00  | 1418.0    |
| 97.10  | 699.0     |
| 98.00  | 16271.0   |
| 99.00  | 13479.0   |
| 100.00 | 1223.0    |
| 101.00 | 7618.0    |
| 102.00 | 525.0     |
| 103.00 | 2650.0    |
| 103.90 | 4688.0    |
| 105.00 | 4963.0    |
| 106.00 | 1605.0    |
| 107.00 | 57456.0   |
| 108.00 | 9840.0    |
| 109.10 | 1893.0    |
| 110.00 | 114760.0  |
| 111.00 | 17104.0   |
| 112.00 | 1715.0    |
| 113.00 | 752.0     |
| 114.90 | 594.0     |
| 116.00 | 3323.0    |
| 117.00 | 46576.0   |
| 118.00 | 3385.0    |
| 118.90 | 645.0     |
| 120.00 | 971.0     |
| 121.20 | 448.0     |
| 121.90 | 4360.0    |
| 123.00 | 6219.0    |
| 123.90 | 2851.0    |
| 125.00 | 2832.0    |
| 127.00 | 227200.0  |
| 128.00 | 17072.0   |
| 129.00 | 85936.0   |
| 130.00 | 7321.0    |
| 131.00 | 1379.0    |
| 132.00 | 712.0     |
| 133.00 | 651.0     |
| 133.90 | 2367.0    |
| 135.00 | 7418.0    |
| 136.00 | 3040.0    |
| 137.00 | 4158.0    |
| 137.90 | 800.0     |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

|        |          |
|--------|----------|
| 139.00 | 728.0    |
| 140.00 | 1307.0   |
| 141.00 | 11746.0  |
| 142.00 | 4010.0   |
| 142.90 | 2387.0   |
| 144.00 | 714.0    |
| 145.00 | 781.0    |
| 146.00 | 2252.0   |
| 147.00 | 6290.0   |
| 148.00 | 13269.0  |
| 149.00 | 2950.0   |
| 150.00 | 975.0    |
| 151.00 | 1625.0   |
| 151.90 | 1103.0   |
| 153.00 | 4113.0   |
| 154.00 | 2893.0   |
| 155.00 | 6430.0   |
| 156.00 | 9960.0   |
| 157.00 | 1870.0   |
| 157.90 | 2431.0   |
| 159.00 | 1656.0   |
| 160.00 | 3732.0   |
| 161.00 | 5909.0   |
| 162.00 | 1575.0   |
| 163.00 | 624.0    |
| 164.00 | 728.0    |
| 165.00 | 4824.0   |
| 166.00 | 3566.0   |
| 167.00 | 24776.0  |
| 168.00 | 12901.0  |
| 169.00 | 2546.0   |
| 169.90 | 926.0    |
| 171.00 | 1038.0   |
| 171.90 | 2309.0   |
| 173.00 | 3084.0   |
| 174.00 | 5360.0   |
| 175.00 | 10299.0  |
| 176.00 | 3231.0   |
| 177.00 | 4814.0   |
| 178.00 | 1634.0   |
| 179.00 | 19984.0  |
| 180.00 | 14333.0  |
| 181.00 | 6696.0   |
| 182.00 | 1116.0   |
| 182.80 | 739.0    |
| 184.00 | 1862.0   |
| 185.00 | 9424.0   |
| 186.00 | 75184.0  |
| 187.00 | 20544.0  |
| 188.00 | 2105.0   |
| 189.00 | 4613.0   |
| 190.00 | 964.0    |
| 191.00 | 2387.0   |
| 192.00 | 5683.0   |
| 193.00 | 7158.0   |
| 194.00 | 1632.0   |
| 195.10 | 892.0    |
| 196.00 | 17280.0  |
| 198.00 | 603328.0 |
| 199.00 | 40280.0  |
| 200.00 | 3123.0   |
| 201.50 | 3168.0   |
| 203.00 | 3502.0   |
| 204.00 | 20152.0  |
| 205.00 | 34904.0  |
| 206.00 | 152960.0 |
| 207.00 | 20192.0  |
| 208.00 | 5419.0   |
| 209.00 | 1740.0   |
| 210.20 | 3171.0   |
| 210.90 | 6264.0   |
| 213.10 | 457.0    |
| 215.00 | 1910.0   |
| 216.00 | 3251.0   |
| 217.00 | 40472.0  |
| 218.00 | 5213.0   |
| 219.00 | 416.0    |
| 221.00 | 30128.0  |
| 221.90 | 5924.0   |
| 223.00 | 9341.0   |
| 224.00 | 86600.0  |
| 225.00 | 22152.0  |
| 226.00 | 2758.0   |
| 227.00 | 35560.0  |
| 228.00 | 4337.0   |
| 229.00 | 7108.0   |
| 230.00 | 908.0    |
| 231.00 | 3288.0   |
| 232.00 | 652.0    |
| 233.00 | 711.0    |
| 234.00 | 2320.0   |
| 235.00 | 2832.0   |
| 236.00 | 1728.0   |
| 237.00 | 3356.0   |
| 238.00 | 675.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

|        |          |
|--------|----------|
| 238.90 | 1540.0   |
| 240.00 | 1327.0   |
| 241.00 | 2091.0   |
| 242.00 | 5022.0   |
| 243.10 | 5986.0   |
| 244.00 | 76448.0  |
| 245.00 | 10126.0  |
| 246.00 | 13767.0  |
| 247.00 | 2964.0   |
| 247.90 | 791.0    |
| 249.00 | 3022.0   |
| 250.00 | 737.0    |
| 250.90 | 773.0    |
| 252.10 | 601.0    |
| 253.10 | 1516.0   |
| 255.00 | 396928.0 |
| 256.00 | 57184.0  |
| 257.00 | 4278.0   |
| 258.00 | 21416.0  |
| 259.00 | 3176.0   |
| 260.00 | 642.0    |
| 260.90 | 866.0    |
| 263.90 | 672.0    |
| 265.00 | 8174.0   |
| 266.00 | 1490.0   |
| 269.90 | 433.0    |
| 271.00 | 752.0    |
| 272.00 | 1186.0   |
| 273.00 | 12497.0  |
| 274.00 | 34152.0  |
| 275.00 | 196480.0 |
| 276.00 | 25296.0  |
| 277.00 | 14871.0  |
| 278.00 | 2758.0   |
| 279.00 | 668.0    |
| 281.00 | 732.0    |
| 282.00 | 541.0    |
| 283.00 | 1902.0   |
| 284.00 | 1186.0   |
| 285.00 | 2829.0   |
| 286.00 | 461.0    |
| 288.90 | 669.0    |
| 289.90 | 491.0    |
| 291.00 | 334.0    |
| 291.90 | 779.0    |
| 293.00 | 2960.0   |
| 294.00 | 1083.0   |
| 295.00 | 1264.0   |
| 296.00 | 53320.0  |
| 297.00 | 7336.0   |
| 298.00 | 444.0    |
| 301.00 | 670.0    |
| 302.10 | 891.0    |
| 303.00 | 6647.0   |
| 304.00 | 1652.0   |
| 304.90 | 342.0    |
| 308.00 | 821.0    |
| 309.10 | 550.0    |
| 310.00 | 801.0    |
| 312.90 | 381.0    |
| 314.00 | 2652.0   |
| 315.00 | 5814.0   |
| 316.00 | 3541.0   |
| 316.90 | 593.0    |
| 321.00 | 1744.0   |
| 322.00 | 1178.0   |
| 323.00 | 19216.0  |
| 324.00 | 3142.0   |
| 324.90 | 328.0    |
| 326.00 | 433.0    |
| 326.90 | 3384.0   |
| 328.00 | 1728.0   |
| 328.90 | 353.0    |
| 332.00 | 1197.0   |
| 333.00 | 1934.0   |
| 334.00 | 12264.0  |
| 335.10 | 3110.0   |
| 336.00 | 434.0    |
| 339.00 | 361.0    |
| 340.00 | 395.0    |
| 341.00 | 1956.0   |
| 342.00 | 690.0    |
| 346.00 | 3958.0   |
| 346.90 | 956.0    |
| 351.10 | 599.0    |
| 352.00 | 5815.0   |
| 353.00 | 3808.0   |
| 354.00 | 6352.0   |
| 355.00 | 1158.0   |
| 358.90 | 504.0    |
| 365.00 | 26056.0  |
| 366.00 | 3509.0   |
| 369.90 | 724.0    |
| 371.10 | 1631.0   |
| 372.00 | 10851.0  |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

|        |          |
|--------|----------|
| 373.00 | 2672.0   |
| 373.90 | 443.0    |
| 377.00 | 399.0    |
| 383.00 | 3216.0   |
| 383.90 | 811.0    |
| 390.00 | 1329.0   |
| 391.00 | 678.0    |
| 392.10 | 795.0    |
| 401.00 | 833.0    |
| 402.00 | 4459.0   |
| 403.00 | 6801.0   |
| 404.00 | 2144.0   |
| 405.00 | 449.0    |
| 421.00 | 5437.0   |
| 422.00 | 5031.0   |
| 423.00 | 37784.0  |
| 424.10 | 8596.0   |
| 425.10 | 797.0    |
| 441.10 | 110256.0 |
| 442.00 | 692352.0 |
| 443.00 | 129232.0 |
| 444.00 | 11700.0  |
| 445.10 | 913.0    |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43

| m/z    | Abundance |
|--------|-----------|
| 39.00  | 338.0     |
| 39.90  | 1270.0    |
| 41.00  | 1047.0    |
| 43.10  | 1183.0    |
| 44.00  | 8640.0    |
| 55.00  | 1041.0    |
| 56.10  | 446.0     |
| 57.00  | 1196.0    |
| 67.00  | 312.0     |
| 69.00  | 705.0     |
| 70.00  | 321.0     |
| 71.00  | 937.0     |
| 73.00  | 608.0     |
| 81.00  | 382.0     |
| 83.00  | 733.0     |
| 83.90  | 452.0     |
| 85.10  | 540.0     |
| 96.00  | 346.0     |
| 97.10  | 594.0     |
| 111.00 | 358.0     |
| 191.00 | 323.0     |
| 207.00 | 1895.0    |
| 208.00 | 471.0     |
| 281.00 | 363.0     |

Instrument :  
BNA\_N  
ClientSampleId :  
DFTPP43