

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110519\
Quantitation Report (QT Reviewed)

Data File : BN008454.D
Acq On : 05 Nov 2019 19:25
Operator : JU
Sample : K5567-04
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Nov 06 01:47:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN110519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 06 00:56:29 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.56	152	1214	0.40	ng/ul	0.00
2) Naphthalene-d8	10.32	136	5290	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.19	164	3126	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.95	188	6618m	0.40	ng/ul	0.00
16) Chrysene-d12	21.16	240	5220	0.40	ng/ul	0.00
20) Perylene-d12	23.32	264	3195m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.93	152	2628	0.32	ng/ul	0.00
14) Fluoranthene-d10	18.99	212	8777	0.45	ng/ul	0.00
Target Compounds						
5) 2-Methylnaphthalene	12.01	142	104	0.025	ng/ul	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN008454.D
Acq On : 05 Nov 2019 19:25
Operator : JU
Sample : K5567-04
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Nov 06 01:47:55 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN110519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 06 00:56:29 2019
Response via : Initial Calibration

