

Data File : BN008457.D  
Acq On : 05 Nov 2019 21:12  
Operator : JU  
Sample : K5567-07  
Misc :  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :

Quant Time: Nov 06 02:02:37 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SOM-EPA-SIM-BN110519.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Nov 06 00:56:29 2019  
Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|-------|--------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 7.56  | 152  | 1420     | 0.40  | ng/ul  | 0.00     |
| 2) Naphthalene-d8           | 10.32 | 136  | 6233     | 0.40  | ng/ul  | 0.00     |
| 6) Acenaphthene-d10         | 14.19 | 164  | 3771     | 0.40  | ng/ul  | 0.00     |
| 10) Phenanthrene-d10        | 16.95 | 188  | 7761m    | 0.40  | ng/ul  | 0.00     |
| 16) Chrysene-d12            | 21.16 | 240  | 6498     | 0.40  | ng/ul  | 0.00     |
| 20) Perylene-d12            | 23.32 | 264  | 4095m    | 0.40  | ng/ul  | 0.00     |
| System Monitoring Compounds |       |      |          |       |        |          |
| 4) 2-Methylnaphthalene-d10  | 11.93 | 152  | 2767     | 0.29  | ng/ul  | 0.00     |
| 14) Fluoranthene-d10        | 18.99 | 212  | 8536     | 0.37  | ng/ul  | 0.00     |
| Target Compounds            |       |      |          |       |        |          |
| 5) 2-Methylnaphthalene      | 12.01 | 142  | 72       | 0.021 | ng/ul  | 97       |
| 21) Benzo(b)fluoranthene    | 22.69 | 252  | 126      | 0.031 | ng/ul# | 1        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN008457.D  
Acq On : 05 Nov 2019 21:12  
Operator : JU  
Sample : K5567-07  
Misc :  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :

Quant Time: Nov 06 02:02:37 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_N\METHODS\SOM-EPA-SIM-BN110519.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Nov 06 00:56:29 2019  
Response via : Initial Calibration

