

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN110618\
 Data File : BN003442.D
 Acq On : 06 Nov 2018 21:29
 Operator : MJ/SJ
 Sample : J5826-12
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SB-S

Quant Time: Nov 07 10:47:02 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-BN102418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 07 08:25:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	112700	20.00	ng	0.00
21) Naphthalene-d8	10.65	136	465408	20.00	ng	0.00
39) Acenaphthene-d10	14.49	164	270109	20.00	ng	0.00
64) Phenanthrene-d10	17.23	188	582988	20.00	ng	0.00
76) Chrysene-d12	21.40	240	476775	20.00	ng	0.00
87) Perylene-d12	23.72	264	478903	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	744372	115.04	ng	0.01
7) Phenol-d6	7.01	99	871332	94.70	ng	0.00
23) Nitrobenzene-d5	9.02	82	631011	77.76	ng	0.00
42) 2,4,6-Tribromophenol	15.97	330	498967	137.19	ng	0.00
45) 2-Fluorobiphenyl	13.11	172	1650270	82.87	ng	0.00
79) Terphenyl-d14	19.85	244	2258487	101.30	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN110618\
Data File : BN003442.D
Acq On : 06 Nov 2018 21:29
Operator : MJ/SJ
Sample : J5826-12
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SB-S

Quant Time: Nov 07 10:47:02 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-BN102418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 07 08:25:32 2018
Response via : Initial Calibration

