

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN110719\
Quantitation Report (QT Reviewed)

Data File : BN008491.D
Acq On : 07 Nov 2019 19:02
Operator : JU
Sample : K5565-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Nov 08 07:25:21 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN110519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 08 06:44:20 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.55	152	1430	0.40	ng/ul	0.00
2) Naphthalene-d8	10.31	136	5876	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.19	164	3093	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.94	188	5796	0.40	ng/ul	0.00
16) Chrysene-d12	21.15	240	4970	0.40	ng/ul	0.00
20) Perylene-d12	23.31	264	4955m	0.40	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	11.92	152	1620	0.17	ng/ul	0.00
14) Fluoranthene-d10	18.98	212	3379	0.20	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN008491.D
Acq On : 07 Nov 2019 19:02
Operator : JU
Sample : K5565-06
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Nov 08 07:25:21 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN110519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 08 06:44:20 2019
Response via : Initial Calibration

