

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN110719\
Quantitation Report (QT Reviewed)

Data File : BN008496.D
Acq On : 07 Nov 2019 22:01
Operator : JU
Sample : K5565-11
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Nov 08 07:32:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN110519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 08 06:44:20 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.55	152	1468	0.40	ng/ul	0.00
2) Naphthalene-d8	10.31	136	6520	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.19	164	3768	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.94	188	7453m	0.40	ng/ul	0.00
16) Chrysene-d12	21.15	240	4929	0.40	ng/ul	0.00
20) Perylene-d12	23.32	264	4004m	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	11.92	152	2473	0.24	ng/ul	0.00
14) Fluoranthene-d10	18.98	212	5490	0.25	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BN008496.D
Acq On : 07 Nov 2019 22:01
Operator : JU
Sample : K5565-11
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Nov 08 07:32:59 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN110519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 08 06:44:20 2019
Response via : Initial Calibration

