

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110723\
 Data File : BN028489.D
 Acq On : 07 Nov 2023 13:12
 Operator : MA/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Nov 07 16:51:20 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 07 15:43:27 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.717	152	7738	0.400	ng	0.00
7) Naphthalene-d8	10.509	136	23811	0.400	ng	0.00
13) Acenaphthene-d10	14.364	164	15170	0.400	ng	0.00
19) Phenanthrene-d10	17.109	188	35496	0.400	ng	0.00
29) Chrysene-d12	21.329	240	30578	0.400	ng	0.00
35) Perylene-d12	23.643	264	30301	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.290	112	18853	0.815	ng	0.00
5) Phenol-d6	6.886	99	23801	0.800	ng	0.00
8) Nitrobenzene-d5	8.864	82	17042	0.801	ng	0.00
11) 2-Methylnaphthalene-d10	12.102	152	34033	0.824	ng	0.00
14) 2,4,6-Tribromophenol	15.855	330	6359	0.748	ng	0.00
15) 2-Fluorobiphenyl	12.985	172	54124	0.811	ng	0.00
27) Fluoranthene-d10	19.153	212	88400	0.815	ng	0.00
31) Terphenyl-d14	19.766	244	65869	0.810	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.196	88	7669	0.818	ng	96
3) n-Nitrosodimethylamine	3.506	42	8860	0.828	ng	# 98
6) bis(2-Chloroethyl)ether	7.146	93	20151	0.818	ng	99
9) Naphthalene	10.551	128	60808	0.818	ng	99
10) Hexachlorobutadiene	10.850	225	10817	0.819	ng	# 99
12) 2-Methylnaphthalene	12.178	142	41703	0.819	ng	99
16) Acenaphthylene	14.075	152	64246	0.805	ng	100
17) Acenaphthene	14.417	154	42885	0.806	ng	99
18) Fluorene	15.412	166	60217	0.801	ng	99
20) 4,6-Dinitro-2-methylph...	15.497	198	4674	0.714	ng	# 78
21) 4-Bromophenyl-phenylether	16.315	248	19000	0.802	ng	97
22) Hexachlorobenzene	16.414	284	20276	0.817	ng	98
23) Atrazine	16.588	200	16894	0.808	ng	98
24) Pentachlorophenol	16.761	266	8136	0.766	ng	99
25) Phenanthrene	17.146	178	95013	0.807	ng	99
26) Anthracene	17.245	178	86450	0.802	ng	100
28) Fluoranthene	19.185	202	114198	0.809	ng	100
30) Pyrene	19.547	202	117877	0.809	ng	100
32) Benzo(a)anthracene	21.311	228	102815	0.804	ng	100
33) Chrysene	21.364	228	105848	0.829	ng	99
34) Bis(2-ethylhexyl)phtha...	21.266	149	61183	0.774	ng	99
36) Indeno(1,2,3-cd)pyrene	26.023	276	98585	0.807	ng	100
37) Benzo(b)fluoranthene	22.941	252	105257	0.843	ng	98
38) Benzo(k)fluoranthene	22.988	252	102156	0.838	ng	96
39) Benzo(a)pyrene	23.538	252	87077	0.816	ng	96
40) Dibenzo(a,h)anthracene	26.040	278	76692	0.805	ng	94
41) Benzo(g,h,i)perylene	26.745	276	85014	0.805	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110723\
Data File : BN028489.D
Acq On : 07 Nov 2023 13:12
Operator : MA/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Nov 07 16:51:20 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 07 15:43:27 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.717 min Scan# 64
 Delta R.T. -0.000 min
 Lab File: BN028489.D
 Acq: 07 Nov 2023 13:12




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:152 Resp: 7738
 Ion Ratio Lower Upper
 152 100
 150 151.2 122.5 183.7
 115 58.3 47.5 71.3



#2
 1,4-Dioxane
 Concen: 0.818 ng
 RT: 3.196 min Scan# 16
 Delta R.T. -0.000 min
 Lab File: BN028489.D
 Acq: 07 Nov 2023 13:12

Tgt Ion: 88 Resp: 7669
 Ion Ratio Lower Upper
 88 100
 43 32.2 29.0 43.4
 58 77.0 60.2 90.2









#3
 n-Nitrosodimethylamine
 Concen: 0.828 ng
 RT: 3.506 min Scan# 59
 Delta R.T. -0.007 min
 Lab File: BN028489.D
 Acq: 07 Nov 2023 13:12




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion: 42 Resp: 8860
 Ion Ratio Lower Upper
 42 100
 74 123.6 100.8 151.2
 44 6.4 7.3 10.9#



#4
 2-Fluorophenol
 Concen: 0.815 ng
 RT: 5.290 min Scan# 306
 Delta R.T. -0.000 min
 Lab File: BN028489.D
 Acq: 07 Nov 2023 13:12

Tgt Ion: 112 Resp: 18853
 Ion Ratio Lower Upper
 112 100
 64 55.0 44.6 67.0
 63 28.4 23.3 34.9









#5
Phenol-d6
Concen: 0.800 ng
RT: 6.886 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 99 Resp: 23801
Ion Ratio Lower Upper
99 100
42 19.6 15.8 23.8
71 31.6 25.1 37.7



#6
bis(2-Chloroethyl)ether
Concen: 0.818 ng
RT: 7.146 min Scan# 563
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12

Tgt Ion: 93 Resp: 20151
Ion Ratio Lower Upper
93 100
63 76.7 60.9 91.3
95 32.9 26.2 39.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.509 min Scan# 911
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	23811
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.0	13.4
54	6.6	5.1	7.7
68	5.2	4.2	6.4



#8
Nitrobenzene-d5
Concen: 0.801 ng
RT: 8.864 min Scan# 776
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12

Tgt Ion:	82	Resp:	17042
Ion	Ratio	Lower	Upper
82	100		
128	44.0	36.5	54.7
54	51.6	43.8	65.8





#9

Naphthalene

Concen: 0.818 ng

RT: 10.551 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN028489.D

Acq: 07 Nov 2023 13:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:128 Resp: 60808

Ion Ratio Lower Upper

128 100

129 11.1 9.1 13.7

127 13.1 10.6 15.8



#10

Hexachlorobutadiene

Concen: 0.819 ng

RT: 10.850 min Scan# 962

Delta R.T. -0.000 min

Lab File: BN028489.D

Acq: 07 Nov 2023 13:12




Tgt Ion:225 Resp: 10817


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 64.0 51.9 77.9









#11
2-Methylnaphthalene-d10
Concen: 0.824 ng
RT: 12.102 min Scan# 1180
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 34033
Ion Ratio Lower Upper
152 100
151 20.2 16.1 24.1



#12
2-Methylnaphthalene
Concen: 0.819 ng
RT: 12.178 min Scan# 1200
Delta R.T. 0.004 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12

Tgt Ion:142 Resp: 41703
Ion Ratio Lower Upper
142 100
141 88.3 71.3 106.9
115 33.2 27.6 41.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.364 min Scan# 1495
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 15170
Ion Ratio Lower Upper
164 100
162 102.3 82.5 123.7
160 46.5 38.1 57.1



#14
2,4,6-Tribromophenol
Concen: 0.748 ng
RT: 15.855 min Scan# 1630
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12

Tgt Ion:330 Resp: 6359
Ion Ratio Lower Upper
330 100
332 96.4 77.0 115.4
141 40.4 32.6 49.0









#15
2-Fluorobiphenyl
Concen: 0.811 ng
RT: 12.985 min Scan# 1366
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172 Resp: 54124
Ion Ratio Lower Upper
172 100
171 34.5 27.9 41.9
170 23.1 18.9 28.3



#16
Acenaphthylene
Concen: 0.805 ng
RT: 14.075 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12

Tgt Ion:152 Resp: 64246
Ion Ratio Lower Upper
152 100
151 19.3 15.4 23.0
153 12.9 10.5 15.7





#17

Acenaphthene

Concen: 0.806 ng

RT: 14.417 min Scan# 1500

Delta R.T. -0.000 min

Lab File: BN028489.D

Acq: 07 Nov 2023 13:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:154 Resp: 42885

Ion Ratio Lower Upper

154 100

153 114.3 92.0 138.0

152 55.9 45.4 68.0



#18

Fluorene

Concen: 0.801 ng

RT: 15.412 min Scan# 1593

Delta R.T. -0.000 min

Lab File: BN028489.D

Acq: 07 Nov 2023 13:12




Tgt Ion:166 Resp: 60217


Ion Ratio Lower Upper

166 100

165 98.8 78.5 117.7

167 13.4 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.109 min Scan# 1731
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:188 Resp: 35496
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.7 8.3 12.5



#20
4,6-Dinitro-2-methylphenol
Concen: 0.714 ng
RT: 15.497 min Scan# 1601
Delta R.T. 0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12

Tgt Ion:198 Resp: 4674
Ion Ratio Lower Upper
198 100
51 33.5 42.1 63.1#
105 31.0 32.6 48.8#









#21
4-Bromophenyl-phenylether
Concen: 0.802 ng
RT: 16.315 min Scan# 1667
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 19000
Ion Ratio Lower Upper
248 100
250 101.9 84.6 126.8
141 34.3 28.2 42.2



#22
Hexachlorobenzene
Concen: 0.817 ng
RT: 16.414 min Scan# 1675
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12

Tgt Ion:284 Resp: 20276
Ion Ratio Lower Upper
284 100
142 38.6 31.8 47.6
249 30.7 25.4 38.0





#23

Atrazine

Concen: 0.808 ng

RT: 16.588 min Scan# 1689

Delta R.T. -0.000 min

Lab File: BN028489.D

Acq: 07 Nov 2023 13:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:200 Resp: 16894

Ion Ratio Lower Upper

200 100

173 20.5 18.7 28.1

215 48.2 38.3 57.5



#24

Pentachlorophenol

Concen: 0.766 ng

RT: 16.761 min Scan# 1703

Delta R.T. -0.000 min

Lab File: BN028489.D

Acq: 07 Nov 2023 13:12




Tgt Ion:266 Resp: 8136


Ion Ratio Lower Upper

266 100

264 62.8 50.5 75.7

268 64.0 51.9 77.9









#25
Phenanthrene
Concen: 0.807 ng
RT: 17.146 min Scan# 1734
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:178 Resp: 95013
Ion Ratio Lower Upper
178 100
176 18.9 15.1 22.7
179 15.2 12.5 18.7



#26
Anthracene
Concen: 0.802 ng
RT: 17.245 min Scan# 1742
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12

Tgt Ion:178 Resp: 86450
Ion Ratio Lower Upper
178 100
176 18.4 14.8 22.2
179 15.2 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.815 ng
RT: 19.153 min Scan# 1982
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 88400
Ion Ratio Lower Upper
212 100
106 14.4 11.5 17.3
104 8.1 6.4 9.6



#28
Fluoranthene
Concen: 0.809 ng
RT: 19.185 min Scan# 1989
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12

Tgt Ion:202 Resp: 114198
Ion Ratio Lower Upper
202 100
101 11.5 9.1 13.7
203 17.2 13.7 20.5





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.329 min Scan# 2417

Delta R.T. -0.000 min

Lab File: BN028489.D

Acq: 07 Nov 2023 13:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:240 Resp: 30578

Ion Ratio Lower Upper

240 100

120 10.1 7.6 11.4

236 27.4 21.9 32.9



#30

Pyrene

Concen: 0.809 ng

RT: 19.547 min Scan# 2067

Delta R.T. -0.000 min

Lab File: BN028489.D

Acq: 07 Nov 2023 13:12




Tgt Ion:202 Resp: 117877


Ion Ratio Lower Upper

202 100

200 20.6 16.5 24.7

203 17.9 14.3 21.5









#31
 Terphenyl-d14
 Concen: 0.810 ng
 RT: 19.766 min Scan# 2114
 Delta R.T. -0.000 min
 Lab File: BN028489.D
 Acq: 07 Nov 2023 13:12




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:244 Resp: 65869
 Ion Ratio Lower Upper
 244 100
 212 8.1 6.6 9.8
 122 14.2 11.4 17.2



#32
 Benzo(a)anthracene
 Concen: 0.804 ng
 RT: 21.311 min Scan# 2415
 Delta R.T. -0.000 min
 Lab File: BN028489.D
 Acq: 07 Nov 2023 13:12

Tgt Ion:228 Resp: 102815
 Ion Ratio Lower Upper
 228 100
 226 27.0 21.4 32.2
 229 19.5 15.7 23.5





#33

Chrysene

Concen: 0.829 ng

RT: 21.364 min Scan# 2421

Delta R.T. -0.000 min

Lab File: BN028489.D

Acq: 07 Nov 2023 13:12

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:228 Resp: 105848

Ion Ratio Lower Upper

228 100

226 29.1 23.8 35.8

229 19.6 15.7 23.5



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.774 ng

RT: 21.266 min Scan# 2410

Delta R.T. -0.000 min

Lab File: BN028489.D

Acq: 07 Nov 2023 13:12




Tgt Ion:149 Resp: 61183


Ion Ratio Lower Upper

149 100

167 27.0 21.0 31.4

279 3.3 2.6 3.8









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.643 min Scan# 2961
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:264 Resp: 30301
Ion Ratio Lower Upper
264 100
260 25.7 20.7 31.1
265 40.2 31.6 47.4



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.807 ng
RT: 26.023 min Scan# 3775
Delta R.T. 0.003 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12

Tgt Ion:276 Resp: 98585
Ion Ratio Lower Upper
276 100
138 28.2 22.4 33.6
277 24.9 19.8 29.6









#37
Benzo(b)fluoranthene
Concen: 0.843 ng
RT: 22.941 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 105257
Ion Ratio Lower Upper
252 100
253 22.6 18.6 27.8
125 12.5 11.4 17.0



#38
Benzo(k)fluoranthene
Concen: 0.838 ng
RT: 22.988 min Scan# 2737
Delta R.T. 0.003 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12


Tgt Ion:252 Resp: 102156
Ion Ratio Lower Upper
252 100
253 22.6 19.0 28.6
125 12.8 12.0 18.0








#39
Benzo(a)pyrene
Concen: 0.816 ng
RT: 23.538 min Scan# 2925
Delta R.T. -0.000 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8





Tgt Ion:252 Resp: 87077
Ion Ratio Lower Upper
252 100
253 22.5 19.3 28.9
125 14.2 13.3 19.9




#40
Dibenzo(a,h)anthracene
Concen: 0.805 ng
RT: 26.040 min Scan# 3781
Delta R.T. -0.003 min
Lab File: BN028489.D
Acq: 07 Nov 2023 13:12



Tgt Ion:278 Resp: 76692
Ion Ratio Lower Upper
278 100
139 20.4 17.0 25.6
279 24.9 23.8 35.8





#41
 Benzo(g,h,i)perylene
 Concen: 0.805 ng
 RT: 26.745 min Scan# 40
 Delta R.T. -0.003 min
 Lab File: BN028489.D
 Acq: 07 Nov 2023 13:12

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Tgt Ion:	276	Resp:	85014
Ion Ratio	Lower	Upper	
276	100		
277	24.1	19.5	29.3
138	25.4	20.1	30.1

