




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110725\
Data File : BN038146.D
Acq On : 07 Nov 2025 09:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 07 09:43:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 07 09:43:12 2025
Response via : Initial Calibration



TIC: BN038146.D\data.ms

(70) Pentachlorophenol (C)




16.851min (0.000) 278.07 ng

response 2058575

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.80
264.00	61.60	62.64
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110725\
Data File : BN038146.D
Acq On : 07 Nov 2025 09:00
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 07 09:43:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 07 09:43:12 2025
Response via : Initial Calibration



TIC: BN038146.D\data.ms

(77) Benzidine

19.451min (0.000) 0.00 ng

response 10437790

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	15.39
183.00	13.20	12.84
0.00	0.00	0.00