

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110725\
 Data File : BN038163.D
 Acq On : 07 Nov 2025 20:22
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 08 03:06:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN102825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 29 13:09:46 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.413	0.413	0.0	103	0.00
3	n-Nitrosodimethylamine	0.532	0.501	5.8	99	0.00
4 S	2-Fluorophenol	0.942	1.042	-10.6	116	0.00
5 S	Phenol-d6	1.148	1.319	-14.9	124	0.00
6	bis(2-Chloroethyl)ether	1.128	1.115	1.2	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.323	0.320	0.9	117	0.00
9	Naphthalene	1.102	1.073	2.6	113	0.00
10	Hexachlorobutadiene	0.186	0.181	2.7	111	0.00
11 SURR	2-Methylnaphthalene-d10	0.571	0.580	-1.6	120	0.00
12	2-Methylnaphthalene	0.735	0.734	0.1	118	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
14 S	2,4,6-Tribromophenol	0.165	0.160	3.0	133	0.00
15 S	2-Fluorobiphenyl	1.602	1.503	6.2	117	0.00
16	Acenaphthylene	1.811	1.699	6.2	122	0.00
17	Acenaphthene	1.267	1.186	6.4	121	0.00
18	Fluorene	1.661	1.576	5.1	122	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
20	4,6-Dinitro-2-methylphenol	0.055	0.032	41.8#	86	0.00
21	4-Bromophenyl-phenylether	0.262	0.257	1.9	124	0.00
22	Hexachlorobenzene	0.298	0.289	3.0	117	-0.01
23	Atrazine	0.192	0.181	5.7	124	0.00
24	Pentachlorophenol	0.128	0.071	44.5#	79	0.00
25	Phenanthrene	1.270	1.192	6.1	117	0.00
26	Anthracene	1.111	1.044	6.0	120	0.00
27 SURR	Fluoranthene-d10	1.009	0.915	9.3	115	0.00
28	Fluoranthene	1.419	1.252	11.8	112	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
30	Pyrene	1.806	1.993	-10.4	111	0.00
31 S	Terphenyl-d14	0.869	1.009	-16.1	119	0.00
32	Benzo(a)anthracene	1.427	1.361	4.6	103	0.00
33	Chrysene	1.549	1.505	2.8	101	0.00
34	Bis(2-ethylhexyl)phthalate	0.752	1.016	-35.1#	145	0.00
35 I	Perylene-d12	1.000	1.000	0.0	104	0.00
36	Indeno(1,2,3-cd)pyrene	1.643	1.756	-6.9	114	0.00
37	Benzo(b)fluoranthene	1.710	1.564	8.5	100	0.00
38	Benzo(k)fluoranthene	1.721	1.582	8.1	102	0.00
39 C	Benzo(a)pyrene	1.369	1.318	3.7	104	0.00
40	Dibenzo(a,h)anthracene	1.265	1.354	-7.0	118	0.01
41	Benzo(g,h,i)perylene	1.444	1.504	-4.2	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4