

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110821\
 Data File : BN017338.D
 Acq On : 08 Nov 2021 17:07
 Operator : CG/JU
 Sample : M4412-11
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BG367

Quant Time: Nov 08 17:38:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 08 11:43:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.514	152	5305	0.400	ng/ul	0.00
4) Naphthalene-d8	10.282	136	21282	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.169	164	13373	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.922	188	28020	0.400	ng/ul	0.00
17) Chrysene-d12	21.135	240	24521	0.400	ng/ul	0.00
23) Perylene-d12	23.329	264	25657	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.038	96	27127	4.599	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.888	152	10146	0.326	ng/ul	0.00
18) Fluoranthene-d10	18.961	212	30040	0.467	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110821\
Data File : BN017338.D
Acq On : 08 Nov 2021 17:07
Operator : CG/JU
Sample : M4412-11
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BG367

Quant Time: Nov 08 17:38:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 08 11:43:44 2021
Response via : Initial Calibration

