

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110821\
 Data File : BN017329.D
 Acq On : 08 Nov 2021 10:59
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4289

Quant Time: Nov 08 11:43:56 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 08 11:43:44 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.510	152	5769	0.400	ng/ul	0.00
4) Naphthalene-d8	10.283	136	21783	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.165	164	12406	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.918	188	26005	0.400	ng/ul	0.00
17) Chrysene-d12	21.135	240	23025	0.400	ng/ul	0.00
23) Perylene-d12	23.329	264	19033	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.042	96	2491	0.388	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.883	152	12342	0.388	ng/ul	0.00
18) Fluoranthene-d10	18.961	212	27872	0.462	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.076	88	3015	0.438	ng/ul#	74
5) Naphthalene	10.327	128	24490	0.406	ng/ul	99
7) 2-Methylnaphthalene	11.960	142	16385	0.393	ng/ul	100
8) 1-Methylnaphthalene	12.180	142	16750	0.392	ng/ul	100
10) Acenaphthylene	13.878	152	18507	0.377	ng/ul	100
11) Acenaphthene	14.225	153	18082	0.420	ng/ul	100
12) Fluorene	15.224	166	20288	0.403	ng/ul	99
14) Pentachlorophenol	16.576	266	1635	0.279	ng/ul	100
15) Phenanthrene	16.960	178	34670	0.420	ng/ul	100
16) Anthracene	17.053	178	28024	0.410	ng/ul	100
19) Fluoranthene	18.989	202	39083	0.474	ng/ul	100
20) Pyrene	19.356	202	38964	0.461	ng/ul	98
21) Benzo(a)anthracene	21.120	228	28535	0.378	ng/ul	99
22) Chrysene	21.170	228	36972	0.432	ng/ul	100
24) Benzo(b)fluoranthene	22.672	252	33134	0.467	ng/ul	99
25) Benzo(k)fluoranthene	22.715	252	36332	0.452	ng/ul	99
26) Benzo(a)pyrene	23.233	252	26401	0.415	ng/ul	100
27) Indeno(1,2,3-cd)pyrene	25.535	276	32168	0.356	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.558	278	24672	0.345	ng/ul	99
29) Benzo(g,h,i)perylene	26.202	276	30179	0.395	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110821\
Data File : BN017329.D
Acq On : 08 Nov 2021 10:59
Operator : CG/JU
Sample : SSTDC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4289

Quant Time: Nov 08 11:43:56 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN102721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 08 11:43:44 2021
Response via : Initial Calibration

