

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110823\
 Data File : BN028513.D
 Acq On : 09 Nov 2023 00:30
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 09 02:07:54 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN110723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 08 00:27:19 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	127	0.00
2	1,4-Dioxane	0.400	0.378	5.5	121	0.00
3	n-Nitrosodimethylamine	0.400	0.389	2.8	126	0.00
4 S	2-Fluorophenol	0.400	0.389	2.8	123	0.00
5 S	Phenol-d6	0.400	0.414	-3.5	133	0.00
6	bis(2-Chloroethyl)ether	0.400	0.404	-1.0	129	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	137	-0.01
8 S	Nitrobenzene-d5	0.400	0.384	4.0	135	0.00
9	Naphthalene	0.400	0.397	0.8	137	0.00
10	Hexachlorobutadiene	0.400	0.399	0.3	137	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.387	3.3	136	0.00
12	2-Methylnaphthalene	0.400	0.388	3.0	136	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	127	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.334	16.5	116	0.00
15 S	2-Fluorobiphenyl	0.400	0.400	0.0	129	0.00
16	Acenaphthylene	0.400	0.389	2.8	131	0.00
17	Acenaphthene	0.400	0.388	3.0	126	0.00
18	Fluorene	0.400	0.360	10.0	116	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	110	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.385	3.8	95	-0.01
21	4-Bromophenyl-phenylether	0.400	0.393	1.8	113	-0.01
22	Hexachlorobenzene	0.400	0.416	-4.0	118	0.00
23	Atrazine	0.400	0.357	10.8	104	-0.01
24	Pentachlorophenol	0.400	0.318	20.5	101	0.00
25	Phenanthrene	0.400	0.385	3.8	109	0.00
26	Anthracene	0.400	0.384	4.0	112	-0.01
27 SURR	Fluoranthene-d10	0.400	0.361	9.8	102	0.00
28	Fluoranthene	0.400	0.360	10.0	102	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	91	0.00
30	Pyrene	0.400	0.429	-7.2	100	0.00
31 S	Terphenyl-d14	0.400	0.422	-5.5	98	0.00
32	Benzo(a)anthracene	0.400	0.390	2.5	92	0.00
33	Chrysene	0.400	0.399	0.3	92	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.417	-4.2	103	0.00
35 I	Perylene-d12	0.400	0.400	0.0	73	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.389	2.8	74	-0.01
37	Benzo(b)fluoranthene	0.400	0.404	-1.0	76	0.00
38	Benzo(k)fluoranthene	0.400	0.414	-3.5	79	0.00
39 C	Benzo(a)pyrene	0.400	0.396	1.0	75	0.00
40	Dibenzo(a,h)anthracene	0.400	0.401	-0.3	77	-0.01
41	Benzo(g,h,i)perylene	0.400	0.381	4.8	72	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4