

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN110918\
 Data File : BN003484.D
 Acq On : 09 Nov 2018 22:17
 Operator : MJ/SJ
 Sample : J5860-19
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EP2-SB-109(17-18)

Manual Integrations
 APPROVED

Sohil
 11/12/2018 5:07:52 PM

Quant Time: Nov 12 10:18:51 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN102418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 07 08:25:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	109791	20.00	ng	0.00
21) Naphthalene-d8	10.65	136	426623	20.00	ng	0.00
39) Acenaphthene-d10	14.49	164	265889	20.00	ng	0.00
64) Phenanthrene-d10	17.23	188	588955	20.00	ng	0.00
76) Chrysene-d12	21.40	240	512481	20.00	ng	0.00
87) Perylene-d12	23.72	264	498206	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	921378	146.17	ng	0.00
7) Phenol-d6	7.01	99	1250156	139.48	ng	0.00
23) Nitrobenzene-d5	9.03	82	682475	91.74	ng	0.01
42) 2,4,6-Tribromophenol	15.97	330	669304	186.94	ng	0.00
45) 2-Fluorobiphenyl	13.11	172	1780563	90.83	ng	0.00
79) Terphenyl-d14	19.85	244	2616528	109.19	ng	0.00
Target Compounds						
10) Phenol	7.03	94	91272	9.649	ng	99
37) 2-Methylnaphthalene	12.32	142	1952429	131.049	ng	99
38) 1-Methylnaphthalene	12.53	142	1200783m	82.522	ng	
50) Dimethylphthalate	13.94	163	336571	15.655	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN110918\
Data File : BN003484.D
Acq On : 09 Nov 2018 22:17
Operator : MJ/SJ
Sample : J5860-19
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EP2-SB-109(17-18)

Manual Integrations
APPROVED

Sohil
11/12/2018 5:07:52 PM

Quant Time: Nov 12 10:18:51 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN102418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 07 08:25:32 2018
Response via : Initial Calibration

