

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN110918\
 Data File : BN003476.D
 Acq On : 09 Nov 2018 17:36
 Operator : MJ/SJ
 Sample : J5860-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EP2-SB-104(15-17)

Quant Time: Nov 12 03:46:43 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN102418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 07 08:25:32 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	103425	20.00	ng	0.00
21) Naphthalene-d8	10.65	136	438792	20.00	ng	0.00
39) Acenaphthene-d10	14.49	164	257925	20.00	ng	0.00
64) Phenanthrene-d10	17.23	188	557774	20.00	ng	0.00
76) Chrysene-d12	21.40	240	450418	20.00	ng	0.00
87) Perylene-d12	23.72	264	451946	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	875082	147.37	ng	0.00
7) Phenol-d6	7.01	99	1185449	140.40	ng	0.00
23) Nitrobenzene-d5	9.02	82	681886	89.12	ng	0.00
42) 2,4,6-Tribromophenol	15.97	330	583606	168.04	ng	0.00
45) 2-Fluorobiphenyl	13.10	172	1811674	95.28	ng	0.00
79) Terphenyl-d14	19.85	244	2344388	111.31	ng	0.00
Target Compounds						
10) Phenol	7.03	94	87008	9.765	ng	98
50) Dimethylphthalate	13.93	163	233548	11.199	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN110918\
Data File : BN003476.D
Acq On : 09 Nov 2018 17:36
Operator : MJ/SJ
Sample : J5860-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EP2-SB-104(15-17)

Quant Time: Nov 12 03:46:43 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN102418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 07 08:25:32 2018
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.85 min Scan# 826
 Delta R.T. -0.01 min
 Lab File: BN003476.D
 Acq: 09 Nov 2018 17:36

Instrument :
 BNA_N
 ClientSampleId :
 EP2-SB-104(15-17)

Tot Ion	Ratio	Lower	Upper
152	100		
150	156.0	123.8	185.8
115	57.7	45.9	68.9




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
 Concen: 147.370 ng
 RT: 5.42 min Scan# 413
 Delta R.T. 0.01 min
 Lab File: BN003476.D
 Acq: 09 Nov 2018 17:36

Tgt Ion	Ratio	Lower	Upper
112	100		
64	52.0	46.6	69.8
63	25.7	23.8	35.8



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BNC
 Ion 63.00 (62.70 to 63.70): BNC





#7

Phenol-d6

Concen: 140.397 ng

RT: 7.01 min Scan# 684

Delta R.T. 0.00 min

Lab File: BN003476.D

Acq: 09 Nov 2018 17:36

Instrument :

BNA_N

ClientSampleId :

EP2-SB-104(15-17)


Tot Ion: 99 Resp: 1185449

Ion Ratio Lower Upper




99 100

42 14.5 13.0 19.6

71 29.7 24.5 36.7



Abundance Ion 99.00 (98.70 to 99.70): BN0
Ion 42.00 (41.70 to 42.70): BN0
Ion 71.00 (70.70 to 71.70): BN0



#10

Phenol

Concen: 9.765 ng

RT: 7.03 min Scan# 688

Delta R.T. 0.00 min

Lab File: BN003476.D

Acq: 09 Nov 2018 17:36


Tgt Ion: 94 Resp: 87008

Ion Ratio Lower Upper



94 100


65 26.4 7.7 47.7

66 34.5 15.8 55.8



Abundance Ion 94.00 (93.70 to 94.70): BN0
Ion 65.00 (64.70 to 65.70): BN0
Ion 66.00 (65.70 to 66.70): BN0






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.65 min Scan# 1302
Delta R.T. -0.01 min
Lab File: BN003476.D
Acq: 09 Nov 2018 17:36

Instrument :
BNA_N
ClientSampleId :
EP2-SB-104(15-17)

Tot Ion:	136	Resp:	438792
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	7.1	6.6	10.0
68	6.3	4.6	7.0






Abundance

Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E


Ion 54.00 (53.70 to 54.70): BNC

Ion 68.00 (67.70 to 68.70): BNC



#23
Nitrobenzene-d5
Concen: 89.121 ng
RT: 9.02 min Scan# 1025
Delta R.T. -0.01 min
Lab File: BN003476.D
Acq: 09 Nov 2018 17:36

Tgt Ion:	82	Resp:	681886
Ion	Ratio	Lower	Upper
82	100		
128	49.3	37.1	55.7
54	45.0	39.4	59.2






Abundance

Ion 82.00 (81.70 to 82.70): BNC

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BNC





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.49 min Scan# 1955

Delta R.T. 0.00 min

Lab File: BN003476.D

Acq: 09 Nov 2018 17:36

Instrument :

BNA_N

ClientSampleId :

EP2-SB-104(15-17)


Tot Ion:164 Resp: 257925

Ion Ratio Lower Upper

164 100

162 102.8 80.8 121.2




160 44.3 35.5 53.3



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#42

2,4,6-Tribromophenol

Concen: 168.039 ng

RT: 15.97 min Scan# 2208

Delta R.T. -0.01 min

Lab File: BN003476.D

Acq: 09 Nov 2018 17:36


Tgt Ion:330 Resp: 583606

Ion Ratio Lower Upper

330 100

332 96.0 76.6 114.8



141 27.0 24.5 36.7




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E






#45
 2-Fluorobiphenyl
 Concen: 95.276 ng
 RT: 13.10 min Scan# 1720
 Delta R.T. -0.01 min
 Lab File: BN003476.D
 Acq: 09 Nov 2018 17:36




Instrument :
 BNA_N
 ClientSampleId :
 EP2-SB-104(15-17)

Tot Ion:172 Resp: 1811674

Ion	Ratio	Lower	Upper
172	100		
171	35.7	28.2	42.4
170	23.4	18.4	27.6




Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#50
 Dimethylphthalate
 Concen: 11.199 ng
 RT: 13.93 min Scan# 1861
 Delta R.T. -0.02 min
 Lab File: BN003476.D
 Acq: 09 Nov 2018 17:36


Tgt Ion:163 Resp: 233548

Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.4	5.2
164	10.1	7.9	11.9



Abundance Ion 163.00 (162.70 to 163.70): E
 Ion 194.00 (193.70 to 194.70): E
 Ion 164.00 (163.70 to 164.70): E









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.23 min Scan# 2421
Delta R.T. -0.01 min
Lab File: BN003476.D
Acq: 09 Nov 2018 17:36

Instrument :
BNA_N
ClientSampleId :
EP2-SB-104(15-17)

Tot Ion:188 Resp: 557774
Ion Ratio Lower Upper
188 100
94 9.3 8.1 12.1
80 10.2 8.6 13.0




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BNC
Ion 80.00 (79.70 to 80.70): BNC






#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.40 min Scan# 3131
Delta R.T. -0.01 min
Lab File: BN003476.D
Acq: 09 Nov 2018 17:36

Tgt Ion:240 Resp: 450418
Ion Ratio Lower Upper
240 100
120 11.2 9.4 14.0
236 26.1 20.3 30.5



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#79

Terphenyl-d14

Concen: 111.310 ng

RT: 19.85 min Scan# 2866

Delta R.T. -0.01 min

Lab File: BN003476.D

Acq: 09 Nov 2018 17:36

Instrument :

BNA_N

ClientSampleId :

EP2-SB-104(15-17)


Tot Ion:244 Resp: 2344388

Ion Ratio Lower Upper

244 100

212 6.8 5.4 8.0




122 10.9 9.4 14.2



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#87

Perylene-d12

Concen: 20.000 ng

RT: 23.72 min Scan# 3524

Delta R.T. -0.01 min

Lab File: BN003476.D

Acq: 09 Nov 2018 17:36


Tgt Ion:264 Resp: 451946

Ion Ratio Lower Upper

264 100

260 23.4 18.5 27.7

265 21.7 16.6 25.0



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

