

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110920\
 Data File : BN012509.D
 Acq On : 09 Nov 2020 12:43
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 09 13:59:51 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 05:46:47 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.626	0.555	11.3	82	0.00
3	n-Nitrosodimethylamine	1.709	1.704	0.3	96	0.00
4 S	2-Fluorophenol	1.398	1.443	-3.2	96	0.00
5 S	Phenol-d6	1.636	1.829	-11.8	102	0.00
6	bis(2-Chloroethyl)ether	1.115	0.989	11.3	88	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.473	0.446	5.7	91	-0.01
9	Naphthalene	1.122	1.019	9.2	88	0.00
10	Hexachlorobutadiene	0.245	0.233	4.9	90	-0.01
11 SURR	2-Methylnaphthalene-d10	0.649	0.582	10.3	88	0.00
12	2-Methylnaphthalene	0.724	0.676	6.6	91	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	-0.01
14 S	2,4,6-Tribromophenol	0.290	0.306	-5.5	107	0.00
15 S	2-Fluorobiphenyl	1.560	1.416	9.2	88	-0.01
16	Acenaphthylene	2.005	1.783	11.1	90	0.00
17	Acenaphthene	1.276	1.155	9.5	91	0.00
18	Fluorene	1.627	1.481	9.0	91	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.097	0.098	-1.0	109	0.00
21	4-Bromophenyl-phenylether	0.231	0.208	10.0	92	0.00
22	Hexachlorobenzene	0.301	0.282	6.3	92	-0.01
23	Atrazine	0.240	0.222	7.5	94	0.00
24	Pentachlorophenol	0.141	0.144	-2.1	109	0.00
25	Phenanthrene	1.184	1.078	9.0	91	-0.01
26	Anthracene	1.096	1.004	8.4	94	0.00
27 SURR	Fluoranthene-d10	1.240	1.120	9.7	92	0.00
28	Fluoranthene	1.413	1.276	9.7	92	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	92	-0.01
30	Pyrene	1.344	1.234	8.2	94	0.00
31 S	Terphenyl-d14	0.954	0.878	8.0	93	0.00
32	Benzo(a)anthracene	1.273	1.192	6.4	95	0.00
33	Chrysene	1.333	1.201	9.9	92	0.00
34	Bis(2-ethylhexyl)phthalate	0.864	0.779	9.8	95	0.00
35	Indeno(1,2,3-cd)pyrene	2.006	2.172	-8.3	103	0.00
36 I	Perylene-d12	1.000	1.000	0.0	95	0.00
37	Benzo(b)fluoranthene	1.276	1.176	7.8	100	0.00
38	Benzo(k)fluoranthene	1.360	1.262	7.2	98	0.00
39 C	Benzo(a)pyrene	1.249	1.189	4.8	102	0.00
40	Dibenzo(a,h)anthracene	1.416	1.427	-0.8	102	0.00
41	Benzo(g,h,i)perylene	1.458	1.467	-0.6	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4