

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110920\
 Data File : BN012524.D
 Acq On : 09 Nov 2020 21:44
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 10 04:02:30 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 05 05:46:47 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	98	0.00
2	1,4-Dioxane	0.400	0.332	17.0	86	0.00
3	n-Nitrosodimethylamine	0.400	0.355	11.3	96	0.00
4 S	2-Fluorophenol	0.400	0.411	-2.7	107	0.00
5 S	Phenol-d6	0.400	0.392	2.0	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.376	6.0	105	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.381	4.8	102	-0.01
9	Naphthalene	0.400	0.370	7.5	99	0.00
10	Hexachlorobutadiene	0.400	0.375	6.3	99	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.376	6.0	103	0.00
12	2-Methylnaphthalene	0.400	0.379	5.3	103	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	92	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.280	30.0#	75	0.00
15 S	2-Fluorobiphenyl	0.400	0.396	1.0	101	-0.01
16	Acenaphthylene	0.400	0.373	6.8	99	0.00
17	Acenaphthene	0.400	0.388	3.0	103	0.00
18	Fluorene	0.400	0.377	5.8	100	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	97	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.049	87.8#	15	0.03
21	4-Bromophenyl-phenylether	0.400	0.350	12.5	97	0.00
22	Hexachlorobenzene	0.400	0.366	8.5	98	-0.01
23	Atrazine	0.400	0.364	9.0	101	0.00
24	Pentachlorophenol	0.400	0.090	77.5#	26	0.01
25	Phenanthrene	0.400	0.391	2.3	106	-0.01
26	Anthracene	0.400	0.370	7.5	103	0.00
27 SURR	Fluoranthene-d10	0.400	0.371	7.3	102	0.00
28	Fluoranthene	0.400	0.385	3.8	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	100	-0.01
30	Pyrene	0.400	0.384	4.0	107	0.00
31 S	Terphenyl-d14	0.400	0.381	4.8	105	0.00
32	Benzo(a)anthracene	0.400	0.357	10.8	99	0.00
33	Chrysene	0.400	0.387	3.3	108	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.415	-3.7	119	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.358	10.5	93	-0.01
36 I	Perylene-d12	0.400	0.400	0.0	96	0.00
37	Benzo(b)fluoranthene	0.400	0.360	10.0	99	0.00
38	Benzo(k)fluoranthene	0.400	0.372	7.0	100	0.00
39 C	Benzo(a)pyrene	0.400	0.360	10.0	98	0.00
40	Dibenzo(a,h)anthracene	0.400	0.355	11.3	91	0.00
41	Benzo(g,h,i)perylene	0.400	0.361	9.8	92	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4