

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110923\
 Data File : BN028533.D
 Acq On : 09 Nov 2023 23:57
 Operator : MA/JU
 Sample : 05082-13
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BH343

Quant Time: Nov 10 00:53:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 09 16:30:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.703	152	7478	0.400	ng/ul	0.00
4) Naphthalene-d8	10.496	136	25441	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.354	164	16501	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.103	188	37078	0.400	ng/ul	0.00
17) Chrysene-d12	21.312	240	28858	0.400	ng/ul	0.00
23) Perylene-d12	23.618	264	31008	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.151	96	37946	4.591	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.091	152	15013	0.392	ng/ul	0.00
18) Fluoranthene-d10	19.141	212	50296	0.542	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.189	88	1614	0.177	ng/ul#	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110923\
Data File : BN028533.D
Acq On : 09 Nov 2023 23:57
Operator : MA/JU
Sample : 05082-13
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BH343

Quant Time: Nov 10 00:53:49 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 09 16:30:32 2023
Response via : Initial Calibration

