

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110923\
 Data File : BN028536.D
 Acq On : 10 Nov 2023 01:45
 Operator : MA/JU
 Sample : 05082-16
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BH352

Quant Time: Nov 10 02:11:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 09 16:30:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.703	152	7662	0.400	ng/ul	0.00
4) Naphthalene-d8	10.495	136	26255	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.352	164	16355	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.102	188	36733	0.400	ng/ul	0.00
17) Chrysene-d12	21.312	240	30342	0.400	ng/ul	0.00
23) Perylene-d12	23.615	264	33372	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.151	96	35290	4.167	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.090	152	14892	0.377	ng/ul	0.00
18) Fluoranthene-d10	19.140	212	45645	0.467	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.184	88	12032	1.285	ng/ul#	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 10 02:11:44 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 09 16:30:32 2023
Response via : Initial Calibration

