

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110923\
 Data File : BN028548.D
 Acq On : 10 Nov 2023 09:33
 Operator : MA/JU
 Sample : 05044-04
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BHEF8

Quant Time: Nov 10 23:26:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 09 16:30:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.703	152	7165	0.400	ng/ul	0.00
4) Naphthalene-d8	10.491	136	23997	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.349	164	14969	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.099	188	33748	0.400	ng/ul	0.00
17) Chrysene-d12	21.314	240	27079	0.400	ng/ul	0.00
23) Perylene-d12	23.617	264	28827	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.151	96	32839	4.147	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.091	152	11602	0.321	ng/ul	0.00
18) Fluoranthene-d10	19.137	212	34243	0.393	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.184	88	2902	0.332	ng/ul#	85
15) Phenanthrene	17.141	178	3325	0.032	ng/ul#	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN110923\
Data File : BN028548.D
Acq On : 10 Nov 2023 09:33
Operator : MA/JU
Sample : 05044-04
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BHEF8

Quant Time: Nov 10 23:26:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 09 16:30:32 2023
Response via : Initial Calibration

